

^{129}Xe Ultrafast Z-spectroscopy enables micromolar detection of biosensors on a 1T benchtop spectrometer

Kévin Chighine¹, Estelle Léonce¹, Céline Boutin¹, Hervé Desvaux¹, and Patrick Berthault¹

¹Université Paris-Saclay, CNRS, CEA, Nanosciences et Innovation pour les Matériaux, la Biomédecine et l'Energie, 91191, Gif-sur-Yvette, France

Correspondence: Patrick Berthault (patrick.berthault@cea.fr)

Abstract. The availability of a benchtop NMR spectrometer, of low cost and easily transportable, can allow detection of low quantities of biosensors, provided that hyperpolarized species are used. Here we show that the micromolar threshold can easily be reached, by employing laser-polarized xenon and cage-molecules reversibly hosting it. Indirect detection of caged xenon is made *via* chemical exchange, using ultrafast Z-spectroscopy based on spatio-temporal encoding. On this non-dedicated
5 low-field spectrometer, several ideas are proposed to improve the signal.

1 Introduction

In this sad period overshadowed by pandemia, among the analytical methods aiming at imaging the lung-blood transfer, hyperpolarized xenon NMR/MRI increasingly interests the '*in vivo*' scientific community. While xenon nuclear polarization can easily be boosted *via* spin-exchange optical pumping (Walker and Happer, 1997, SEOP), the other interest of this noble gas
10 for NMR is that it exhibits a wide chemical shift range (more than 320 ppm for the monoatomic species) and is soluble in most biological fluids. Therefore, xenon is a powerful exogenous probe of the functioning of the air-blood barrier (Driehuys et al., 2006). Moreover, it is prone to open the way to molecular magnetic resonance imaging. In an approach pioneered by A. Pines and co-workers (Spence et al., 2001), xenon is reversibly encapsulated in molecular systems that are functionalized with biological ligands. This two-step procedure, where the bioprobe is first introduced and then hyperpolarized xenon is delivered,
15 benefits from the difference in resonance frequency between bound xenon and free xenon (in the gas phase or in the dissolved phase, cf. Berthault et al. (2009)).

In this method, most of the studies used cryptophane derivatives as xenon hosts, as despite a complex synthesis they are functionalizable by ligands (Brotin and Dutasta, 2009). For instance, to our knowledge, only one example of chemical functionalization of a xenon host other than cryptophane - a cucurbituril - has been reported in the literature (Truxal et al., 2019).
20 The approach of ^{129}Xe NMR-based biosensing using functionalized cryptophanes has yet been successfully applied *in vitro* for detection of small analytes (Tassali et al., 2014; Dubost et al., 2014; Jeong et al., 2015; Yang et al., 2016; Guo et al., 2016; Yang et al., 2017), of large biosystems (Wei et al., 2006; Chambers et al., 2009; Boutin et al., 2011; Rose et al., 2014; Taratula et al., 2015; Khan et al., 2015; Riggle et al., 2017; Milanole et al., 2017; Schnurr et al., 2020), or of change of physiological conditions: temperature (Schröder et al., 2008; Schilling et al., 2010) or pH (Léonce et al., 2018). To date, it has however never




Figure 1. Generic structure of the cryptophanes used in this study. For compound **1**: $m = 2$ and $n = 3$; for **2**: $m = 3$ and $n = 2$.

25 been used *in vivo*; only a proof of concept has been performed on rats using a non-functionalized cucurbituril (Hane et al., 2017). Several difficulties or obstacles have delayed *in vivo* applications, among which obviously the lack of sensitivity.

We made the remark that in a pre-clinical environment, it could be very useful to test the behaviour of such bioprobes in NMR, using a benchtop spectrometer, less cumbersome and less expensive than a high-field spectrometer. It could be placed very close to the optical pumping setup, working in flow or batch modes. The present work aims at assessing the feasibility of the detection of low concentrations of ¹²⁹Xe NMR-based biosensors using a non-dedicated benchtop spectrometer. After a brief description of our spin-exchange optical pumping setup working in the batch mode, direct and indirect detection techniques are studied both at low and high magnetic field. Theoretical considerations are given and practical ways of improvement are analyzed.

2 Results

35 2.1 Molecular systems

For such a study, we decided to use two water-soluble cryptophanes, synthesized by the group of T. Brotin at ENS Lyon, and previously characterized (Huber et al., 2006). Their generic structure is depicted in Fig. 1. In cryptophane **1** the two cyclotrivertilene bowls are connected by two O-(CH₂)₂-O linkers and a O-(CH₂)₃-O linker; in cryptophane **2** they are connected by two O-(CH₂)₃-O linkers and a O-(CH₂)₂-O linker. Xenon inside these cage-molecules resonates at 52 ppm (for Xe@**1**) and 42 ppm (for Xe@**2**) if one calibrates the signal of xenon free in water at 196 ppm. These two molecular systems could be used as probes for pH, as the chemical shift of caged xenon varies as a function of the concentration in H⁺ ions (Léonce et al., 2018).

2.2 Addition of hyperpolarized xenon into the samples




Figure 2. Principle of the experiments. Laser-polarized xenon is produced via Spin-Exchange Optical Pumping in the batch mode with the setup described in Chauvin et al. (2016). BE: Beam Expander; BS: polarization Beam Splitter cube; MI: mirror; $\lambda/4$: quarterwave plates; PW: Powermeter. The black vertical rectangle represent the coils, which deliver a magnetic field B colinear to the photon beam. A picture of the optical pumping cell is given in inset. After some minutes of optical pumping, frozen hyperpolarized xenon is transported to the NMR spectrometers. The disposition of the NMR tube inside the magnet, the static and radiofrequency fields as well as the axis of the gradient are indicated on the benchtop spectrometer.

The principle of the experiments is depicted in Fig. 2. Xenon enriched at 83 % in isotope 129 was polarized via spin-exchange optical pumping using a home-built setup already described in Chauvin et al. (2016). After about 10 minutes optical pumping, hyperpolarized xenon was collected frozen and transported immersed in liquid nitrogen inside a 0.3 T solenoid fed by a car battery. Then, in the fringe field of the unshielded 11.7 T magnet, xenon was heated and transferred to the capped NMR tube containing the sample, thanks to a vacuum line and a hollow spinner filled with liquid nitrogen. This procedure enabled

us to transfer all xenon above the solution without freezing it. Finally fast heating and vigorous shaking of the tube speeded up the dissolution of the noble gas into the sample of interest. This method, proved efficient for high magnetic field experiments, is not directly translatable to the NMR experiments in the benchtop spectrometer. All attempt to use the fringe field of the permanent magnet during the xenon transfer and shaking step led to disappointing results in terms of remaining polarization. This was expected as the magnet, of the Halbach type, delivers an horizontal static magnetic field with a steep slope in intensity and direction in the vertical dimension, and thus xenon may cross areas of null field during its introduction in the spectrometer. The least bad solution was thus to carry out the introduction of polarized xenon inside the NMR tube in the fringe field of the superconducting magnet, then to quickly shake the tube in this magnetic environment before transporting it to the benchtop spectrometer. The following shaking, carried out as close as possible to the permanent magnet, leads to a faster depolarization.

Without optimization other than the setting of the quarterwave plates, this robust setup gave us a useable polarization of *ca.* 0.15, as measured in the gas phase on the 11.7 T spectrometer by comparison with the thermal equilibrium NMR signal.

The pressure in the NMR tube on top of the solution was *ca.* 1 bar, as estimated post NMR experiment by weighing the tube before and after degassing.

2.3 Direct detection of ^{129}Xe NMR-based biosensors

Figure 3 displays a comparison of one-scan ^{129}Xe NMR spectra of the same sample of cryptophanes dissolved at a concentration of $77\ \mu\text{M}$ in D_2O , at 11.7 and 1 Tesla (same NMR tube). The two spectra were recorded with different xenon batches, but acquired in the same experimental conditions.

Not surprisingly, the signal-to-noise ratio is better at 11.7 Tesla than at 1 T. For information purposes only, they were measured at 2492 and 1142, respectively, taking into account the signal of xenon free in water. Obviously if the xenon nuclear polarization should be similar for the two experiments as it directly derives from the optical pumping step, the signal-to-noise ratios are expected to be very different due to the detection part. Indeed, in classical NMR, it is the magnetic induction, proportional to the temporal derivative of the component of the magnetization perpendicular to the static magnetic field that is detected, and thus the Larmor frequency on the one side, and all geometrical and electronic parameters defining the sensitivity (coil geometry, useful volume, quality factor, tuning frequency, filter and digitization...) on the other side would have to be considered for comparing the signal-to-noise ratio.

From Fig. 3, several remarks can be made. While at high field the ^{129}Xe NMR spectrum displays one signal at 196 ppm corresponding to xenon free in water and two distinct signals at high field corresponding to caged xenon (Xe@1 at 52 ppm and Xe@2 at 42 ppm according to Huber et al. (2006)), at low static magnetic field one of the latter signals disappears or is hardly observable. The full-width at half-maximum of the Xe@1 signals were roughly measured to 44 Hz and 20 Hz at 11.7 and 1 Tesla, respectively. The presence of a narrow peak at 11.7 T and a very broad line at 1T for Xe@2 reveals the weakness of the direct detection approach, even if one uses a fast repetition 'frequency-selective pulse - detection' (Berthault et al., 2008). When the lines are so broad and the signal-to-noise so low, each of these acquisitions brings a lot of noise and the resulting signal is difficult to distinguish from the baseline.




Figure 3. One-scan ^{129}Xe NMR spectra of the noble gas into a water solution of cryptophanes **1** and **2**, both at $77\ \mu\text{M}$ and 295 K, with the same pulse flip angle. Top spectrum performed at 11.7 Tesla ; bottom spectrum performed at 1 Tesla.

2.4 Indirect detection

It is well known that the detection of diluted species in exchange with a main spin reservoir can be facilitated by taking advantage of this exchange. This has given rise to the Chemical Exchange Saturation Transfer (CEST) experiments, which principle lies in the rf saturation at the frequency of these diluted species and the observation of the subsequent loss of magnetization of the main reservoir due to the exchange (see for instance Vinogradov et al. (2013)). Schröder and co-workers have extended this approach to hyperpolarized species, creating the Hyper-CEST sequence (Schroder et al., 2006).

But sequences of the CEST family are usually reserved to high magnetic fields. For instance, for detection of metabolites such as glutamate or carnosine, ^1H CEST is an efficient method (Cai et al., 2012; Bodet et al., 2015). But it would not come to mind to use a CEST sequence with a low magnetic field, as the frequency splitting between the two environments in exchange is only ca. 3.3 ppm (140 Hz at 1 Tesla). However, due to the very wide chemical shift range of xenon, this can be envisioned in the case of ^{129}Xe NMR-based biosensors.

Due to the huge advantage of a spectrum-per-spectrum averaging instead of a point-per-point averaging for hyperpolarized species, we decided to turn to the ultra-fast version of the CEST experiment, initially proposed by Jerschow and co-workers in ^1H (Xu et al., 2013). In this sequence, saturation is applied in the presence of a gradient, which amounts to saturate only a slice of the sample. After a read pulse, the receiver is open in the presence of a second gradient which decodes the profile of the sample along it. By subtracting two experiments recorded with and without saturation, one obtains the Z-spectrum. Note that this sequence can be combined to localized spectroscopy or imaging schemes, see for instance Döpfert et al. (2014b); Xu et al. (2016); Liu et al. (2016).

In the past, the Ultra-Fast Z-spectroscopy (UFZ) experiment was successfully applied in hyperpolarized ^{129}Xe NMR for the study of biosensors by us and others (Boutin et al., 2013; Döpfert et al., 2014a), and also for detection of low amounts of biological cells (Berthault and Boutin, 2015)).

Note that Gouilleux and co-workers were the first to use pulsed field gradients and to implement spatially-encoded pulse sequences on a low magnetic field spectrometer, which allowed access to experiments such as DOSY and Ultrafast COSY; see a review in Gouilleux et al. (2020) On another side, King et al. have demonstrated the performance of the Ultrafast multidimensional Laplace NMR - based also on spatial encoding - on a low field, single-sided magnet to derive T_1 - T_2 and D - T_2 correlation maps (King et al., 2016, 2018). This method reveals its full potential when used with DNP hyperpolarized species.

For this study, we used the pulse sequence depicted in Fig. 4a. It contains an rf offset switch: the first value, ω_B , is centered on the Xe@cryptophanes region (*i.e.* around 45 ppm), while the second value, ω_A , is applied on-resonance with the free xenon frequency. Such a sequence is not intended to provide the full Z-spectrum (saturation at the free xenon frequency would be disastrous for hyperpolarization), but is prone to reveal the presence of caged xenon (*cf.* Boutin et al. (2013) for details).

Figure 4b displays an ^{129}Xe UFZ-spectrum recorded on the cryptophane mixture at 11.7 Tesla and 295 K.

The apparent chemical shift splitting between the two dips can be measured at 25.4 ppm, which well corresponds to the real chemical shift splitting of 10 ppm when the ratio between the saturation and acquisition gradients is considered (see Theoretical part). The dip corresponding to Xe@2 is broader than that of Xe@1, which can be explained by a faster xenon in-out exchange.

Figure 5 displays an ^{129}Xe UFZ-spectrum recorded on the cryptophane mixture at 1 Tesla and 295 K. For the spatial encoding of the ^{129}Xe UFZ-spectroscopy experiment, on the Magritek SpinSolve C spectrometer we had the choice to use the shim system to create pulsed field gradients in the x , y or z directions, or to use the installed gradient system. The latter gradient is oriented horizontally while the main axis of the NMR tube is vertical (see Fig. 2). It could have been interesting to use the shim along the NMR tube axis, as the magnetization profile would have been along the largest dimension of the tube and would have given a flat profile. However, for stability reasons we decided to use the nominal gradient system. This explains the rounded shape of the magnetization profile envelope, corresponding to an axial projection of the tube. Note that quantitative interpretation of the UFZ-spectra, *e. g.* to extract exchange rates from a set of UFZ spectra, would need consideration of the excitation-detection profile of the rf coil (Ahola et al., 2015).

The gradient strength was firstly calibrated by acquiring a 1D axial profile of the tube in ^1H . Then we chose the value of the gradient simultaneous to ^{129}Xe saturation, G_{sat} , keeping in mind that it had to induce a spectral expansion of the signal




Figure 4. a) Pulse sequence used for the ^{129}Xe UFZ-spectroscopy. ω_B and ω_A denote the rf offset placed in the region of caged xenon and at the frequency of free xenon, respectively. b) ^{129}Xe UFZ-spectrum of the noble gas into the solution containing the cryptophane mixture at 11.7 T and 295 K. In black: the profile obtained after 2 s saturation at a field strength $B_1 = \omega_1/\gamma = 7\mu\text{T}$ (*On* experiment); in blue: same profile without saturation (*Off* experiment); in green: *Off* - *On*. Other important parameters: $G_{\text{sat}} = 35 \text{ mT}\cdot\text{m}^{-1}$; $G_{\text{acq}} = 90 \text{ mT}\cdot\text{m}^{-1}$.

saturation smaller than the difference between the free xenon frequency ω_A and the caged xenon frequency ω_B , thus filling the condition:

$$\gamma \cdot G_{\text{sat}} \cdot d < |\omega_A - \omega_B|$$

130 with d the inner diameter of the tube. In our case, $d = 0.43 \text{ cm}$, $\gamma = -11.79 \text{ MHz}\cdot\text{T}^{-1}$, $|\omega_A - \omega_B| \simeq 1800 \text{ Hz}$ induces that G_{sat} must be lower than $35.5 \text{ mT}\cdot\text{m}^{-1}$.

The profile of the UFZ-spectrum reveals two dips corresponding to xenon in the two cryptophanes, with an apparent frequency splitting of about 40 ppm, which is the expected value for Xe@1 and Xe@2 given the ratio $G_{\text{sat}}/G_{\text{acq}}$. The rf carrier having been placed on resonance with the Xe@1 frequency, we have assigned the dips as displayed on Fig. 5. The apparent
 135 reverse frequency axis is due to the relative sign of the two gradients. Obviously the separation between the dips is less net




Figure 5. ^{129}Xe UFZ-spectrum of the noble gas into the solution containing the cryptophane mixture at 1 T and 295 K, using the same pulse sequence as for Fig. 4 (sum of two experiments with saturation of 2 s and 4 s). The saturation offset was placed at the Xe@1 frequency. In black: the profile obtained after saturation at a field strength $B_1 = \omega_1/\gamma = 0.87 \mu\text{T}$ (*On experiment*); in blue: same profile without saturation (*Off experiment*); Other important parameters: $G_{\text{sat}} = 21 \text{ mT}\cdot\text{m}^{-1}$; $G_{\text{acq}} = 84 \text{ mT}\cdot\text{m}^{-1}$.

than at high field, and to confirm our assignment we have performed the same experiment with another xenon batch, simply inverting the sign of G_{sat} . The corresponding two-scan UFZ-spectrum is displayed in Fig. 6.

In order to model the behavior of the Z-spectra as a function of the experimental parameters, we have simulated the Hyper-CEST experiment (see Theory Section). As input of these simulations, R_{1A} can be measured in an experiment consisting in a series {small flip angle pulse - detection}, by taking into account the flip angle. R_{2A} can be roughly estimated from the linewidth of A, but with a large uncertainty. f can be directly determined from a simple ^{129}Xe spectrum as it is the ratio of the signal B to the signal A. But the parameter which remains difficult to predict or estimate is k_{out} .

In a first step, we used Equations 13 and 11 to simulate the trends on the ^{129}Xe UFZ spectra recorded at 11.7 and 1 T (Fig. 7). In Fig. 7a, in order to evidence the effect of the Larmor frequency, all other parameters have been kept identical (f is the fraction of caged xenon; k_{out} is the xenon out rate):




Figure 6. ^{129}Xe UFZ-spectrum of the noble gas into the solution containing the cryptophane mixture at 1 T and 295 K, using the same pulse sequence as for Fig. 5, except that the saturation gradient is opposite: $G_{\text{sat}} = -21 \text{ mT}\cdot\text{m}^{-1}$; $G_{\text{acq}} = 84 \text{ mT}\cdot\text{m}^{-1}$. Saturation delay: 3 s.

$f = 0.04$; $k_{\text{out}} = 50 \text{ s}^{-1}$; $\delta_A = 196 \text{ ppm}$; $\delta_B = 52 \text{ ppm}$; $R_{1A} = 0.01 \text{ s}^{-1}$; $R_{2A} = 50 \text{ s}^{-1}$; $\omega_1 = 521.4 \text{ rad}\cdot\text{s}^{-1}$ ($7 \mu\text{T}$) ; $t_{\text{sat}} = 1\text{s}$.

This simulation shows that the rf saturation strength must be reduced on the low field spectrometer in order to keep a flat baseline, due to the low difference between the free and bound xenon resonance frequencies.

150 Figure 7b displays the simulation for more realistic experimental conditions. The parameters were the following :

- Parameters common to both fields: $f = 0.04$; $k_{\text{out}} = 50 \text{ s}^{-1}$; $\delta_A = 196 \text{ ppm}$; $\delta_B = 52 \text{ ppm}$
- Spinsolve (Larmor frequency = 12.09 MHz): $R_{1A} = 0.01 \text{ s}^{-1}$; $R_{2A} = 20 \text{ s}^{-1}$; $\omega_1 = 64.7 \text{ rad}\cdot\text{s}^{-1}$ ($0.87 \mu\text{T}$) ; $t_{\text{sat}} = 3 \text{ s}$
- Avance (Larmor frequency = 138.36 MHz): $R_{1A} = 0.01 \text{ s}^{-1}$; $R_{2A} = 50 \text{ s}^{-1}$; $\omega_1 = 521.4 \text{ rad}\cdot\text{s}^{-1}$ ($7 \mu\text{T}$) ; $t_{\text{sat}} = 1 \text{ s}$

155 Comparison between the two ^{129}Xe Z-spectra simulated with realistic conditions evidences firstly the effect of the smaller frequency splitting on the Magritek. Despite lower relaxation rates and despite a weaker saturation strength at low field, the direct relaxation term $\lambda_{\text{direct}}(\omega_i, \omega_1)$ has a larger influence on the baseline in the Xe@cryptophane region (purple curve). This




Figure 7. Simulation of ^{129}Xe Z-spectra at 11.7 (green) and 1 (purple) Tesla. (a): Considering same relaxation times, same rf field strength and duration of the saturation ; (b): Considering a saturation of lower amplitude but larger duration for the low-field experiment. Details of the conditions for the simulation: see text.




Figure 8. Simulation of ^{129}Xe Z-spectra at 1 Tesla for a mixture of two cryptophanes according to the saturation field strength. Conditions for the simulation: see text.

term decreases the signal by about three percent. This reveals that the saturation field strength must stay small to ensure the success of the experiment. Secondly, such a small ω_1 requires a longer saturation time. With a saturation time t_{sat} three times longer than the one used at high field, the dip in the Xe@cryptophane region has almost the same depth.

160 The second simulation considered two cryptophanes with characteristics close to the ones used in the present study. The parameters of the ^{129}Xe Z-spectra were those of the real experiment on the benchtop spectrometer:

- Parameters common to both cryptophanes: Larmor frequency = 12.09 MHz ; $R_{1A} = 0.1 \text{ s}^{-1}$; $R_{2A} = 10 \text{ s}^{-1}$; $t_{\text{sat}} = 3 \text{ s}$
- Cryptophane 1: $f = 0.037$; $k_{\text{out}} = 50 \text{ s}^{-1}$; $\delta_B = 52 \text{ ppm}$
- Cryptophane 2: $f = 0.025$; $k_{\text{out}} = 100 \text{ s}^{-1}$; $\delta_B = 42 \text{ ppm}$




Figure 9. ^{129}Xe UFZ-spectrum of the noble gas into the solution containing cryptophane **2** at $19.2\ \mu\text{M}$. Magnetic field: 1 T ; temperature 295 K, same pulse sequence as for Fig. 4 (one scan). ω_B was placed at the Xe@2 frequency. Profile obtained after 3 s saturation at a field strength $B_1 = \omega_1/\gamma = 1.24\ \mu\text{T}$. $G_{\text{sat}} = 21\ \text{mT}\cdot\text{m}^{-1}$; $G_{\text{acq}} = 84\ \text{mT}\cdot\text{m}^{-1}$.

165 Figure 8 displays the effect of the saturation strength ω_1 on the aspect of the ^{129}Xe Z-spectrum of the cryptophane mixture (only the 0 - 100 ppm region is displayed). At low saturation strength, the signal of the cryptophane with the highest f/k_{out} is predominant, as expected (see theoretical part). Then, when ω_1 increases the second dip becomes more pronounced, larger than the first one due to the higher k_{out} value (Equation 9). But increasing further ω_1 leads to significant lowering of the baseline, thus of the contrast, and the signals are less separated. Thus a compromise has to be found to favor detection of the dips and
 170 maximize their separation. With the current conditions of the simulation, a value of ω_1 between 50 and 150 $\text{rad}\cdot\text{s}^{-1}$ seems the best.

These simulations have helped us to find the best experimental conditions.

In such an approach, the detection threshold can further be lowered, maybe at the price of a lack of discrimination between two minor sites in exchange with the main signal. As an example that does not seek to be a record, this experiment has been
 175 repeated with a solution containing cryptophane **2** at a concentration of $19.2\ \mu\text{M}$. Figure 9 displays the ^{129}Xe UFZ profile.

With respect to the previous experiments, only the saturation field strength has been slightly increased from 0.87 to 1.24 μT ($\omega_1 = 64.7$ to $92.4 \text{ rad}\cdot\text{s}^{-1}$). The dip corresponding to caged xenon appears clearly.

3 Materials and Methods

3.1 Preparation of the samples

180 A mother solution of cryptophane **1** was prepared by dissolution of 0.64 mg of the powder in 500 μL D_2O and 20 μL NaOD 0.1 M. The same preparation was done for cryptophane **2**. 50 μL of each of these solutions were mixed and then the solution was diluted by a factor 5 in D_2O , giving a resulting concentration of 77 μM for cryptophane **1** and 78 μM for cryptophane **2**. The solution was placed in a 5 mm-NMR tube equipped with a valve. The tube was degassed before each inroduction of hyperpolarized xenon by three cycles of helium bubbling then evacuation.

185 3.2 Observation of ^{129}Xe on the 11.7 T spectrometer

The experiments were performed at 295 K on a Bruker Avance II spectrometer equipped with a dual ^{129}Xe - ^1H probehead and a GREAT3/10 three-axis gradient amplifier.

3.3 Observation of ^{129}Xe on the benchtop spectrometer

The experiments were performed at 295 K on a Magritek SpinSolve Carbon spectrometer at a magnetic field of 1.02 T (^1H Larmor frequency 43.71 MHz), the temperature of the permanent magnet being 301 K. The X channel of the low-field spectrometer being initially optimized for ^{13}C (*i. e.* 10.9 MHz), it was necessary to lower the temperature of the permanent magnet in order to approach the ^{129}Xe Larmor frequency (12.1 MHz). Obviously, the situation is not optimal for ^{129}Xe , and the value of the 90° pulse was measured to 150 μs at full power. This was calibrated with a sample of thermal xenon inside dodecane doped with Gd^{3+} ions. Note that at such low field, the paramagnetic doping was not very efficient, and a T_1 of *ca.* 10 s was
195 found for ^{129}Xe in this solution. The same procedure was used to calibrate the B_1 strength value for saturation.

The SpinSolve software in expert mode was used to create the pulse programs that were not in the Magritek library. In particular the UltraFast Z-spectroscopy sequence was written and tested, first in ^1H NMR, then in ^{23}Na NMR, before the ^{129}Xe NMR experiments. The pulse program written for the Magritek Spinsolve spectrometer is provided in the Supplementary Information.

200 In order to apply exactly the same processing to the data acquired at 1 and 11.7 T and to compare them safely, a program was written in Python to convert the FID recorded under Spinsolve into JCAMP-DX data readable by the Bruker software, Topspin.

4 Theory of the Hyper-CEST experiment

Let us consider a cryptophane solubilized in water at a sub-millimolar concentration. A small part of hyperpolarized xenon, when introduced in solution, will be reversibly caged in cryptophane. The ^{129}Xe spectrum will thus exhibit two signals: free dissolved xenon (pool A of magnetization M_A^0 , giving rise to the main signal at ω_A) and xenon inside the cryptophane (pool B of magnetization M_B^0 , very minor signal at ω_B). Two types of xenon in-out exchange coexist: Simple dissociation:



(X for xenon, C for cryptophane), and degenerate (or kick-out) exchange, dependent on the ratio xenon concentration to cryptophane concentration (Korchak et al., 2016) :



where the asterisk denotes the hyperpolarized state. As the two xenon pools have different resonance frequencies, these processes can be seen as a unique exchange between free and encapsulated xenon, characterized by the rates k_{in} and k_{out} at steady-state:

$$k_{in} = k_{out} \frac{M_B^0}{M_A^0} = k_{out} \times f \quad (3)$$

recalling that f is the fraction of caged xenon.

The Hyper-CEST experiment consists in saturating with a (CW) rf irradiation of strength ω_1 in the Xe@cryptophane region (pool B), and detecting the influence on the main Xe signal (pool A). This has two consequences:

◇ A direct effect linked to the rf saturation at an offset $\Delta_i = \omega_i - \omega_A$ from the main resonance, which tilts the magnetization by an angle $\theta_i = \tan^{-1} \frac{\omega_1}{|\Delta_i|}$ with θ_i in $[-\pi/2 : \pi/2]$ (Desvaux and Berthault, 1999). The magnetization transverse to this effective field is averaged out by rf field inhomogeneity and transverse relaxation. The magnetization aligned with the effective field is going to relax. This effect, present even in the absence of exchange, is characterized by the depolarization rate:

$$\lambda_{\text{direct}}(\omega_i, \omega_1) = R_{1A} \cos^2 \theta_i + R_{2A} \sin^2 \theta_i \quad (4)$$

where R_{1A} and R_{2A} are the longitudinal and transverse relaxation rates of free xenon, respectively. In the range of chemical exchange observed with cryptophanes $k = k_{in} + k_{out} \gg R_{1A}, R_{2A}$, given the low proportion of xenon caged in the cryptophane, one has:

$$R_1 = (1 - f)R_{1A} + fR_{1B} \simeq R_{1A} \quad (5)$$

$$R_2 = (1 - f)R_{2A} + fR_{2B} \simeq R_{2A} \quad (6)$$

230 ◇ An indirect depolarization due to the saturation transfer from the pool B to the pool A (CEST effect). If the saturation is applied exactly on-resonance with the caged xenon frequency ($\omega_i = \omega_B$), with the assumption $k_{\text{out}} \gg R_2^B$ and $k_{\text{out}} \gg k_{\text{in}}$, the depolarization rate linked to saturation at the frequency of pool B is given by (Zaiss et al., 2012; Kunth et al., 2014):

$$\lambda_{\text{on}}(\omega_1) \simeq f k_{\text{out}} \frac{\omega_1^2}{\omega_1^2 + k_{\text{out}}^2} \quad (7)$$

This has enabled Kunth and co-workers to distinguish three cases (Kunth et al., 2015):

- 235 • $\omega_1 \gg f k_{\text{out}}$: $\lambda_{\text{on}} \simeq f k_{\text{out}}$. *Maximum depolarization rate.*
 • $\omega_1 = f k_{\text{out}}$: $\lambda_{\text{on}} \simeq f k_{\text{out}}/2$. *Fifty percent of the maximum possible depolarization rate.*
 • $\omega_1 \ll f k_{\text{out}}$: $\lambda_{\text{on}} \simeq (f/k_{\text{out}})\omega_1^2$. *Parabolic behavior with the saturation strength.*

In order to understand the UFZ-spectroscopy experiment, one needs to consider the difference frequency between the frequency of the rf irradiation, ω_i , and the frequency of pool B , ω_B , which has a Lorentzian shape as a function of ω_i (Zaiss et al., 2012).

240

$$\lambda_{\text{CEST}}(\omega_i, \omega_1) = \frac{\lambda_{\text{on}}(\omega_1) \frac{\Gamma(\omega_1)^2}{4}}{\frac{\Gamma(\omega_1)^2}{4} + (\omega_i - \omega_B)^2} \quad (8)$$

with the full-width at half-maximum value of the depolarization rate given by:

$$\Gamma(\omega_1) \simeq 2\sqrt{\omega_1^2 + k_{\text{out}}^2} \quad (9)$$

Thus, the total depolarization rate is:

$$245 \quad \lambda(\omega_i, \omega_1) = \lambda_{\text{direct}}(\omega_i, \omega_1) + \lambda_{\text{CEST}}(\omega_i, \omega_1) = R_{1A} \cos^2 \theta_i + R_{2A} \sin^2 \theta_i + \frac{\lambda_{\text{on}}(\omega_1) \frac{\Gamma(\omega_1)^2}{4}}{\frac{\Gamma(\omega_1)^2}{4} + (\omega_i - \omega_B)^2} \quad (10)$$

which can be re-written:

$$\lambda(\omega_i, \omega_1) = R_{1A} \frac{1}{1 + \left(\frac{\omega_1}{\omega_i - \omega_A}\right)^2} + R_{2A} \frac{\left(\frac{\omega_1}{\omega_i - \omega_A}\right)^2}{1 + \left(\frac{\omega_1}{\omega_i - \omega_A}\right)^2} + \frac{f k_{\text{out}} \cdot \omega_1^2}{\omega_1^2 + k_{\text{out}}^2 + (\omega_i - \omega_B)^2} \quad (11)$$

If one considers the experiment at 11.7 Tesla with the usual modest saturation strength, $|\omega_i - \omega_A| \gg \omega_1$, Equation 11 can be simplified to:

$$250 \quad \lambda(\omega_i, \omega_1) = R_{1A} + \frac{f k_{\text{out}} \cdot \omega_1^2}{\omega_1^2 + k_{\text{out}}^2 + (\omega_i - \omega_B)^2} \quad (12)$$

The z-magnetization after a saturation of duration t_{sat} is :

$$Z(\omega_i, \omega_1) = M_A^0 \cdot e^{-\lambda(\omega_i, \omega_1) t_{\text{sat}}} \quad (13)$$

Specificity of the Ultra-Fast Z-spectroscopy

The Ultra-Fast Z-spectroscopy (UFZ) consists in applying the rf saturation in the presence of a magnetic field gradient G_{sat} and after a read pulse detecting the signal in the presence of another gradient G_{acq} . The obtained profile reflects the magnetization all along the sample in the G_{acq} gradient direction with dips at positions where saturation is effective according to the rf offset and the gradient G_{sat} . Thus after subtraction from the profile acquired through the same sequence without saturation and application of a scaling factor of the intensity for taking into account relaxation at the transient high polarization, it corresponds to the Z-spectrum. After the read pulse, the acquisition in the presence of a gradient G_{acq} makes that all spectral information is spread on a frequency range equal to $\gamma \cdot G_{\text{acq}} \cdot r$, where r is the dimension of the sample along the gradient axis. In summary, two signals separated on a normal NMR spectrum by a value $\Delta\omega_{B_{12}}$ are now separated in the UFZ-spectrum by:

$$\Delta\omega_{B_{12}}^{\text{UFZ}} = \Delta\omega_{B_{12}} \frac{G_{\text{acq}}}{G_{\text{sat}}} \quad (14)$$

This shows that for a fixed acquisition gradient the biggest apparent separation between the dips is obtained by minimizing the saturation gradient. The price to pay is that the dips or one of the dips can exit from the magnetization envelope if the condition $2|\omega_i - \omega_{B_{1,2}}|/G_{\text{sat}} > \gamma \cdot r$ is encountered.

5 Areas for improvement

The use of Hyper-CEST experiments at low magnetic field (1 T) for detection of low quantities of ^{129}Xe NMR-based biosensors is possible, even with a spectrometer not dedicated to ^{129}Xe observation. Such a benchtop spectrometer, placed close to the optical pumping setup and near the imager could represent a helpful tool to i) know the xenon hyperpolarization level and ii) to know the thermodynamics and kinetics of the complex between the noble gas and the biosensor.

However, some issues have been encountered with these experiments. The most important is the difficulty to keep the hyperpolarization between two scans. Indeed shaking the NMR tube in the absence of a magnetic field high enough and sufficiently homogenous leads to fast depolarization. For each introduction of laser-polarized xenon into the NMR tube, we were limited to two to three experiments. This fast relaxation being linked to the diffusion of the noble gas into field gradients (in amplitude and in direction, see Cates et al. (1988)), a solution to counter this would be to increase the pressure inside the NMR tube, by increasing the amount of hyperpolarized xenon or by adding another gas such as nitrogen. Another solution under study would be to build an homogenous field tunnel between the production site of the hyperpolarized species and the spectrometer, such as in dissolution Dynamic Nuclear Polarization (Milani et al., 2015). In any case, the ultrafast version of this experiment, the UFZ-spectroscopy, exhibits several advantages with respect to the classical version of the Hyper-CEST.

Working with a spectrometer not tuned to the frequency of interest was only possible as i) it is a low field, ii) the lowering of the magnet temperature could improve the situation. However some concern may still appear when using a long rf saturation in the ^{129}Xe UFZ experiments, and every solution to bring the radio frequency of the X channel of the spectrometer closer to the resonance frequency of the nuclei of interest is appealing. An easy and low-cost solution was found through placing a solenoid around the NMR tube, or even better by wrapping two copper foils around the NMR tube on either side of the region to be

285 detected, as displayed in Fig. 10a-b, a setup inspired by Wheeler and Conradi (2012). The resulting inductive tuning decreases the inductance, and thereby induces a raise on the resonance frequency of the circuit.

Figure 10c-d shows the wobble curves obtained with the naked NMR tube (c) and with copper foils manually adjusted on the NMR tube axis (d). Remarkably, the quality factor of the probe is not degraded by the presence of the copper foils, and even a better matching is obtained. At the ^{129}Xe frequency, the radiofrequency reflected amplitude at the probe which is ca. 35
290 μV without the foils (Fig. 10c) falls to 3 μV with them (Fig. 10d).

As can be seen on Fig. 10b, given the gap between the two copper foils, which should be smaller than the original detection zone, a partial rf shielding of the sample must occur. This decreases the detection volume accordingly.

Note also that such a simple setup gives access to all nuclei in the region 10.9 - 12.1 MHz, *i.e.* ^{13}C , ^{27}Al , ^{23}Na and ^{129}Xe . It is *a priori* useable on every benchtop spectrometer, as they are commonly based on a Halbach magnet for the static magnetic
295 field and a solenoid for the rf coil, which principal axis is colinear to the NMR tube and the magnet aperture. Creating a inductive coupling with this arrangement is therefore an easy task.

Author contributions. PB conceived and designed the experiments. KC had the idea and developed the coils to shift the working frequency of the benchtop spectrometer. EL, PB and CB performed the NMR experiments at high and low magnetic field. HD and PB wrote the simulation programs. PB and EL processed the data. PB wrote the manuscript, which was reviewed by all the authors.

300 *Competing interests.* The authors declare no competing interests.

Acknowledgements. The authors warmly thank Thierry Brotin (ENS Lyon, France) for providing samples of cryptophanes, Jean-Claude Berthet (CEA Saclay, France) for the preparation of the SEOP cells, and Craig Eccles (Magritek, Germany) for his help on the SpinSolve Expert software. Support from French National Research Agency (ANR) is greatly acknowledged (Project-ANR-17-LCV2-0002 DESIR, and Project-ANR-19-CE19-0024 PHOENIX).




Figure 10. Example of effect observed on the wobble curve of the X channel on the Magritek Spinsolve Carbon 43 when simple copper foils are wrapped around the NMR tube. a): Scheme of the setup, with the copper foils wrapped around the NMR tube ; b) Picture of one of these tubes. As we were limited to an outer diameter of 5 mm, the lower part of the NMR tube receiving the foils was only 4 mm ; c) Wobble curve of the X channel without the copper foils ; d) Wobble curve with the copper coils.

305 References

- Ahola, S., Zhivonitko, V. V., Mankinen, O., Zhang, G., Kantola, A. M., Chen, H.-Y., Hilty, C., Koptug, I. V., and Telkki, V.-V.: Ultrafast multidimensional Laplace NMR for a rapid and sensitive chemical analysis, *Nat Commun*, 6, 8363, <https://doi.org/10.1038/ncomms9363>, 2015.
- Berthault, P. and Boutin, C.: Chapter 14. Biosensing and Study of Biological Cells using Hyperpolarized ^{129}Xe , in: *New Developments in NMR*, edited by Meersmann, T. and Brunner, E., pp. 261–271, Royal Society of Chemistry, Cambridge, <https://doi.org/10.1039/9781782628378-00261>, 2015.
- Berthault, P., Bogaert-Buchmann, A., Desvaux, H., Huber, G., and Boulard, Y.: Sensitivity and Multiplexing Capabilities of MRI Based on Polarized ^{129}Xe Biosensors, *J. Am. Chem. Soc.*, 130, 16 456–16 457, <https://doi.org/10.1021/ja805274u>, 2008.
- Berthault, P., Huber, G., and Desvaux, H.: Biosensing using laser-polarized xenon NMR/MRI, *Progress in Nuclear Magnetic Resonance Spectroscopy*, 55, 35–60, <https://doi.org/10.1016/j.pnmrs.2008.11.003>, 2009.
- Bodet, O., Goerke, S., Behl, N. G. R., Roeloffs, V., Zaiss, M., and Bachert, P.: Amide proton transfer of carnosine in aqueous solution studied *in vitro* by WEX and CEST experiments: Study of Amide Proton Transfer in Carnosine-Water System, *NMR Biomed.*, 28, 1097–1103, <https://doi.org/10.1002/nbm.3343>, 2015.
- Boutin, C., Stopin, A., Lenda, F., Brotin, T., Dutasta, J.-P., Jamin, N., Sanson, A., Boulard, Y., Leteurtre, F., Huber, G., Bogaert-Buchmann, A., Tassali, N., Desvaux, H., Carrière, M., and Berthault, P.: Cell uptake of a biosensor detected by hyperpolarized ^{129}Xe NMR: The transferrin case, *Bioorganic & Medicinal Chemistry*, 19, 4135–4143, <https://doi.org/10.1016/j.bmc.2011.05.002>, 2011.
- Boutin, C., Léonce, E., Brotin, T., Jerschow, A., and Berthault, P.: Ultrafast Z-Spectroscopy for ^{129}Xe NMR-Based Sensors, *J. Phys. Chem. Lett.*, 4, 4172–4176, <https://doi.org/10.1021/jz402261h>, 2013.
- Brotin, T. and Dutasta, J.-P.: Cryptophanes and Their Complexes—Present and Future, *Chem. Rev.*, 109, 88–130, <https://doi.org/10.1021/cr0680437>, 2009.
- Cai, K., Haris, M., Singh, A., Kogan, F., Greenberg, J. H., Hariharan, H., Detre, J. A., and Reddy, R.: Magnetic resonance imaging of glutamate, *Nat Med*, 18, 302–306, <https://doi.org/10.1038/nm.2615>, 2012.
- Cates, G. D., Schaefer, S. R., and Happer, W.: Relaxation of spins due to field inhomogeneities in gaseous samples at low magnetic fields and low pressures, *Phys. Rev. A*, 37, 2877–2885, <https://doi.org/10.1103/PhysRevA.37.2877>, 1988.
- Chambers, J. M., Hill, P. A., Aaron, J. A., Han, Z., Christianson, D. W., Kuzma, N. N., and Dmochowski, I. J.: Cryptophane Xenon-129 Nuclear Magnetic Resonance Biosensors Targeting Human Carbonic Anhydrase, *J. Am. Chem. Soc.*, 131, 563–569, <https://doi.org/10.1021/ja806092w>, 2009.
- Chauvin, C., Liagre, L., Boutin, C., Mari, E., Léonce, E., Carret, G., Coltrinari, B., and Berthault, P.: Note: Spin-exchange optical pumping in a van, *Review of Scientific Instruments*, 87, 016 105, <https://doi.org/10.1063/1.4940928>, 2016.
- Desvaux, H. and Berthault, P.: Study of dynamic processes in liquids using off-resonance rf irradiation, *Progress in Nuclear Magnetic Resonance Spectroscopy*, 35, 295–340, [https://doi.org/10.1016/S0079-6565\(99\)00011-4](https://doi.org/10.1016/S0079-6565(99)00011-4), 1999.
- Döpfert, J., Witte, C., and Schröder, L.: Fast Gradient-Encoded CEST Spectroscopy of Hyperpolarized Xenon, *ChemPhysChem*, 15, 261–264, <https://doi.org/10.1002/cphc.201300888>, 2014a.
- Döpfert, J., Zaiss, M., Witte, C., and Schröder, L.: Ultrafast CEST imaging, *Journal of Magnetic Resonance*, 243, 47–53, <https://doi.org/10.1016/j.jmr.2014.03.008>, 2014b.

- Driehuys, B., Cofer, G. P., Pollaro, J., Mackel, J. B., Hedlund, L. W., and Johnson, G. A.: Imaging alveolar-capillary gas transfer using hyperpolarized ^{129}Xe MRI, *Proceedings of the National Academy of Sciences*, 103, 18 278–18 283, <https://doi.org/10.1073/pnas.0608458103>, 2006.
- Dubost, E., Dognon, J.-P., Rousseau, B., Milanole, G., Dugave, C., Boulard, Y., Léonce, E., Boutin, C., and Berthault, P.: Understanding a Host-Guest Model System through ^{129}Xe NMR Spectroscopic Experiments and Theoretical Studies, *Angew. Chem. Int. Ed.*, 53, 9837–9840, <https://doi.org/10.1002/anie.201405349>, 2014.
- Gouilleux, B., Farjon, J., and Giraudeau, P.: Gradient-based pulse sequences for benchtop NMR spectroscopy, *Journal of Magnetic Resonance*, 319, 106 810, <https://doi.org/10.1016/j.jmr.2020.106810>, 2020.
- Guo, Q., Zeng, Q., Jiang, W., Zhang, X., Luo, Q., Zhang, X., Bouchard, L.-S., Liu, M., and Zhou, X.: A Molecular Imaging Approach to Mercury Sensing Based on Hyperpolarized ^{129}Xe Molecular Clamp Probe, *Chem. Eur. J.*, 22, 3967–3970, <https://doi.org/10.1002/chem.201600193>, 2016.
- Hane, F. T., Li, T., Smylie, P., Pellizzari, R. M., Plata, J. A., DeBoef, B., and Albert, M. S.: In vivo detection of cucurbit[6]uril, a hyperpolarized xenon contrast agent for a xenon magnetic resonance imaging biosensor, *Sci Rep*, 7, 41 027, <https://doi.org/10.1038/srep41027>, 2017.
- Huber, G., Brotin, T., Dubois, L., Desvaux, H., Dutasta, J.-P., and Berthault, P.: Water Soluble Cryptophanes Showing Unprecedented Affinity for Xenon: Candidates as NMR-Based Biosensors, *J. Am. Chem. Soc.*, 128, 6239–6246, <https://doi.org/10.1021/ja060266r>, 2006.
- Jeong, K., Slack, C. C., Vassiliou, C. C., Dao, P., Gomes, M. D., Kennedy, D. J., Truxal, A. E., Sperling, L. J., Francis, M. B., Wemmer, D. E., and Pines, A.: Investigation of DOTA-Metal Chelation Effects on the Chemical Shift of ^{129}Xe , *ChemPhysChem*, 16, 3573–3577, <https://doi.org/10.1002/cphc.201500806>, 2015.
- Khan, N. S., Riggle, B. A., Seward, G. K., Bai, Y., and Dmochowski, I. J.: Cryptophane-Folate Biosensor for ^{129}Xe NMR, *Bioconjugate Chem.*, 26, 101–109, <https://doi.org/10.1021/bc5005526>, 2015.
- King, J. N., Lee, V. J., Ahola, S., Telkki, V.-V., and Meldrum, T.: Ultrafast Multidimensional Laplace NMR Using a Single-Sided Magnet, *Angew. Chem. Int. Ed.*, 55, 5040–5043, <https://doi.org/10.1002/anie.201511859>, 2016.
- King, J. N., Fallorina, A., Yu, J., Zhang, G., Telkki, V.-V., Hilty, C., and Meldrum, T.: Probing molecular dynamics with hyperpolarized ultrafast Laplace NMR using a low-field, single-sided magnet, *Chem. Sci.*, 9, 6143–6149, <https://doi.org/10.1039/C8SC01329B>, 2018.
- Korchak, S., Kilian, W., Schröder, L., and Mitschang, L.: Design and comparison of exchange spectroscopy approaches to cryptophane–xenon host–guest kinetics, *Journal of Magnetic Resonance*, 265, 139–145, <https://doi.org/10.1016/j.jmr.2016.02.005>, 2016.
- Kunth, M., Witte, C., and Schröder, L.: Quantitative chemical exchange saturation transfer with hyperpolarized nuclei (qHyper-CEST): Sensing xenon-host exchange dynamics and binding affinities by NMR, *The Journal of Chemical Physics*, 141, 194 202, <https://doi.org/10.1063/1.4901429>, 2014.
- Kunth, M., Witte, C., and Schröder, L.: Continuous-wave saturation considerations for efficient xenon depolarization: Saturation Pulse Considerations for Efficient Xenon Depolarization, *NMR in Biomedicine*, 28, 601–606, <https://doi.org/10.1002/nbm.3307>, 2015.
- Léonce, E., Dognon, J.-P., Pitrat, D., Mulatier, J.-C., Brotin, T., and Berthault, P.: Accurate pH Sensing using Hyperpolarized ^{129}Xe NMR Spectroscopy, *Chem. Eur. J.*, 24, 6534–6537, <https://doi.org/10.1002/chem.201800900>, 2018.
- Liu, Z., Dimitrov, I. E., Lenkinski, R. E., Hajibeigi, A., and Vinogradov, E.: UCEPR: Ultrafast localized CEST-spectroscopy with PRESS in phantoms and in vivo: Ultrafast Localized CEST-Spectroscopy with PRESS, *Magn. Reson. Med.*, 75, 1875–1885, <https://doi.org/10.1002/mrm.25780>, 2016.

- Milani, J., Vuichoud, B., Bornet, A., Miéville, P., Mottier, R., Jannin, S., and Bodenhausen, G.: A magnetic tunnel to shelter hyperpolarized fluids, *Review of Scientific Instruments*, 86, 024 101, <https://doi.org/10.1063/1.4908196>, 2015.
- 380 Milanole, G., Gao, B., Paoletti, A., Pieters, G., Dugave, C., Deutsch, E., Rivera, S., Law, F., Perfettini, J.-L., Mari, E., Léonce, E., Boutin, C., Berthault, P., Volland, H., Fenaille, F., Brotin, T., and Rousseau, B.: Bimodal fluorescence/ ^{129}Xe NMR probe for molecular imaging and biological inhibition of EGFR in Non-Small Cell Lung Cancer, *Bioorganic & Medicinal Chemistry*, 25, 6653–6660, <https://doi.org/10.1016/j.bmc.2017.11.002>, 2017.
- 385 Riggle, B. A., Greenberg, M. L., Wang, Y., Wissner, R. F., Zemerov, S. D., Petersson, E. J., and Dmochowski, I. J.: A cryptophane-based “turn-on” ^{129}Xe NMR biosensor for monitoring calmodulin, *Org. Biomol. Chem.*, 15, 8883–8887, <https://doi.org/10.1039/C7OB02391J>, 2017.
- Rose, H. M., Witte, C., Rossella, F., Klippel, S., Freund, C., and Schroder, L.: Development of an antibody-based, modular biosensor for ^{129}Xe NMR molecular imaging of cells at nanomolar concentrations, *Proceedings of the National Academy of Sciences*, 111, 11 697–11 702, <https://doi.org/10.1073/pnas.1406797111>, 2014.
- 390 Schilling, F., Schröder, L., Palaniappan, K. K., Zapf, S., Wemmer, D. E., and Pines, A.: MRI Thermometry Based on Encapsulated Hyperpolarized Xenon, *Chem. Eur. J. of Chem. Phys.*, 11, 3529–3533, <https://doi.org/10.1002/cphc.201000507>, 2010.
- Schnurr, M., Volk, I., Nikolenko, H., Winkler, L., Dathe, M., and Schröder, L.: Functionalized Lipopeptide Micelles as Highly Efficient NMR Depolarization Seed Points for Targeted Cell Labelling in Xenon MRI, *Adv. Biosys.*, 4, 1900 251, <https://doi.org/10.1002/adbi.201900251>, 2020.
- 395 Schroder, L., Lowery, T. J., Hilty, C., Wemmer, D. E., and Pines, A.: Molecular Imaging Using a Targeted Magnetic Resonance Hyperpolarized Biosensor, *Science*, 314, 446–449, <https://doi.org/10.1126/science.1131847>, 2006.
- Schröder, L., Meldrum, T., Smith, M., Lowery, T. J., Wemmer, D. E., and Pines, A.: Temperature Response of ^{129}Xe Depolarization Transfer and Its Application for Ultrasensitive NMR Detection, *Phys. Rev. Lett.*, 100, 257 603, <https://doi.org/10.1103/PhysRevLett.100.257603>, 2008.
- 400 Spence, M. M., Rubin, S. M., Dimitrov, I. E., Ruiz, E. J., Wemmer, D. E., Pines, A., Yao, S. Q., Tian, F., and Schultz, P. G.: Functionalized xenon as a biosensor, *Proceedings of the National Academy of Sciences*, 98, 10 654–10 657, <https://doi.org/10.1073/pnas.191368398>, 2001.
- Taratula, O., Bai, Y., D’Antonio, E. L., and Dmochowski, I. J.: Enantiopure cryptophane- ^{129}Xe nuclear magnetic resonance biosensors targeting carbonic anhydrase, *Supramolecular Chemistry*, 27, 65–71, <https://doi.org/10.1080/10610278.2014.906601>, 2015.
- 405 Tassali, N., Kotera, N., Boutin, C., Léonce, E., Boulard, Y., Rousseau, B., Dubost, E., Taran, F., Brotin, T., Dutasta, J.-P., and Berthault, P.: Smart Detection of Toxic Metal Ions, Pb^{2+} and Cd^{2+} , Using a ^{129}Xe NMR-Based Sensor, *Anal. Chem.*, 86, 1783–1788, <https://doi.org/10.1021/ac403669p>, 2014.
- Truxal, A. E., Cao, L., Isaacs, L., Wemmer, D. E., and Pines, A.: Directly Functionalized Cucurbit[7]uril as a Biosensor for the Selective Detection of Protein Interactions by ^{129}Xe hyperCEST NMR, *Chem. Eur. J.*, 25, 6108–6112, <https://doi.org/10.1002/chem.201900610>, 410 2019.
- Vinogradov, E., Sherry, A. D., and Lenkinski, R. E.: CEST: From basic principles to applications, challenges and opportunities, *Journal of Magnetic Resonance*, 229, 155–172, <https://doi.org/10.1016/j.jmr.2012.11.024>, 2013.
- Walker, T. G. and Happer, W.: Spin-exchange optical pumping of noble-gas nuclei, *Rev. Mod. Phys.*, 69, 629–642, <https://doi.org/10.1103/RevModPhys.69.629>, 1997.

- 415 Wei, Q., Seward, G. K., Hill, P. A., Patton, B., Dimitrov, I. E., Kuzma, N. N., and Dmochowski, I. J.: Designing ^{129}Xe NMR Biosensors for Matrix Metalloproteinase Detection, *J. Am. Chem. Soc.*, 128, 13 274–13 283, <https://doi.org/10.1021/ja0640501>, 2006.
- Wheeler, D. D. and Conradi, M. S.: Practical exercises for learning to construct NMR/MRI probe circuits, *Concepts Magn. Reson.*, 40A, 1–13, <https://doi.org/10.1002/cmr.a.21221>, 2012.
- Xu, X., Lee, J.-S., and Jerschow, A.: Ultrafast Scanning of Exchangeable Sites by NMR Spectroscopy, *Angew. Chem. Int. Ed.*, 52, 8281–420 8284, <https://doi.org/10.1002/anie.201303255>, 2013.
- Xu, X., Yadav, N. N., Song, X., McMahon, M. T., Jerschow, A., van Zijl, P. C., and Xu, J.: Screening CEST contrast agents using ultrafast CEST imaging, *Journal of Magnetic Resonance*, 265, 224–229, <https://doi.org/10.1016/j.jmr.2016.02.015>, 2016.
- Yang, S., Jiang, W., Ren, L., Yuan, Y., Zhang, B., Luo, Q., Guo, Q., Bouchard, L.-S., Liu, M., and Zhou, X.: Biothiol Xenon MRI Sensor Based on Thiol-Addition Reaction, *Anal. Chem.*, 88, 5835–5840, <https://doi.org/10.1021/acs.analchem.6b00403>, 2016.
- 425 Yang, S., Yuan, Y., Jiang, W., Ren, L., Deng, H., Bouchard, L. S., Zhou, X., and Liu, M.: Hyperpolarized ^{129}Xe Magnetic Resonance Imaging Sensor for H_2S , *Chem. Eur. J.*, 23, 7648–7652, <https://doi.org/10.1002/chem.201605768>, 2017.
- Zaiss, M., Schnurr, M., and Bachert, P.: Analytical solution for the depolarization of hyperpolarized nuclei by chemical exchange saturation transfer between free and encapsulated xenon (HyperCEST), *The Journal of Chemical Physics*, 136, 144 106, <https://doi.org/10.1063/1.3701178>, 2012.