


Correction of field instabilities in biomolecular solid-state NMR by simultaneous acquisition of a frequency reference

Václav Římal¹, Morgane Callon¹, Alexander A. Malär¹, Riccardo Cadalbert¹, Anahit Torosyan¹, Thomas Wiegand¹, Matthias Ernst¹, Anja Böckmann², Beat H. Meier¹

5 ¹Physical Chemistry, ETH Zurich, Zurich, CH-8093, Switzerland

²Molecular Microbiology and Structural Biochemistry, UMR 5086 CNRS/Université de Lyon, 69367 Lyon, France

Correspondence: Beat H. Meier (beme@ethz.ch)



Abstract. With the advent of faster magic-angle spinning (MAS) and higher magnetic fields, the resolution of biomolecular solid-state nuclear magnetic resonance (NMR) spectra has been continuously increasing. As a direct consequence, the always narrower spectral lines, especially in proton-detected spectroscopy, are also becoming more sensitive to temporal instabilities of the magnetic field in the sample volume. Field drifts in the order of tenths of ppm occur after probe insertion or temperature change, during cryogen refill, or are intrinsic to the superconducting high-field magnets, particularly in the months after charging.

As an alternative to a field–frequency lock based on deuterium solvent resonance rarely available for solid-state NMR, we present a strategy to compensate non-linear field drifts using simultaneous acquisition of a frequency reference (SAFR). It is based on the acquisition of an auxiliary 1D spectrum in each scan of the experiment. Typically, a small-flip-angle pulse is added at the beginning of the pulse sequence. Based on the frequency of the maximum of the solvent signal, the field evolution in time is reconstructed and used to correct the raw data after acquisition, thereby acting in its principle as a digital lock system. The general applicability of our approach is demonstrated on 2D and 3D protein spectra during various situations with a non-linear field drift. SAFR with small-flip-angle pulses causes no significant loss in sensitivity or increase in experimental time in protein spectroscopy. The correction leads to the possibility of recording high-quality spectra in a typical biomolecular experiment even during non-linear field changes in the order of 0.1 ppm h⁻¹ without the need for hardware solutions, such as stabilizing the temperature of the magnet bore. The improvement of linewidths and peak shapes turns out to be especially important for ¹H-detected spectra under fast MAS, but the method is suitable for the detection of carbon or other nuclei as well.

1 Introduction

Solid-state nuclear magnetic resonance (NMR) witnesses an ongoing increase in spectral resolution, in particular in biomolecular applications, and proton linewidths in the order of 10 Hz (or 0.01 ppm at 1200 MHz) are possible in ¹H-detected spectra under fast magic-angle spinning (MAS) for perdeuterated and fully back-exchanged proteins (Penzel et al., 2019). To fully exploit this high resolution, crucial to extract as much information as possible, the static magnetic field B_0 must be stable within 1 ppb (part per billion) during the duration of the entire experiment (minutes to days). Otherwise, adequate corrections must be applied. The stability is particularly critical at a high magnetic field (Callon et al., 2021; Nimerovsky et al., 2021). Alternatively to correcting the field, the spectrometer frequency could be adjusted in real time. While this is conceivable with modern hardware, we are not aware of such a solution in biomolecular NMR and significant protocol changes would be necessary.

In solution-state NMR, the B_0 instabilities are routinely compensated for by a deuterium field–frequency lock in real time. In solid-state NMR spectroscopy, the implementation of a lock proved difficult in practice and commercial probeheads are rarely

equipped with a lock channel. Modern actively shielded superconducting magnets are reasonably stable in a magnetically tranquil environment and at a constant laboratory temperature with the highest change often given by a linear drift of the field as a function of time. Spectrometers typically provide a hardware linear drift correction that compensates for this field drift, which can be adjusted every few weeks or months by a calibration experiment. More importantly, insertion and removal of the NMR probe from the magnet bore and any changes in the temperature of the sample or the shim cylinder create a relatively strong non-linear drift that lasts, for high-field magnets, for hours (Malär et al., 2021), delaying the start of an experiment. Alternatives to the field–frequency lock approaches, that are applicable to solid-state NMR, too, have been used for reducing the field drift. Malär et al. (2021) describe a spectrometer equipped with a magnet-bore heater system that prevents major long-term field drifts due to temperature changes inside the magnet bore, which greatly increases stability and shortens time constants to reach a stable field after a disturbance. External locks can monitor the ^2D or ^7Li resonance frequency of an auxiliary sample located in the proximity of the main sample, which is then utilized for the lock (Paulson and Zilm, 2005; Takahashi et al., 2012). On benchtop low-field NMR spectrometers, a “SoftLock” system monitoring the non-deuterated solvent signal is available (Minkler et al., 2020). In magnetic resonance imaging, the field–frequency lock (Henry et al., 1999) has advanced to feedback control based on field mapping (Vionnet et al., 2021).

Another solution that eliminates the effect of the field drift in the spectra is data correction after the acquisition. Referencing to an internal ^{13}C resonance was documented (Kupče and Freeman, 2010). Interleaved navigator scans (Thiel et al., 2002) or spectral registration (Near et al., 2015) can be used in medical magnetic resonance spectroscopy in vivo. More recently, a posteriori linear drift correction was implemented for solid-state NMR spectroscopy (Najbauer and Andreas, 2019). The latter method profits from a fast and easy application without any requirement of extra hardware, because one only needs to measure a reference spectrum before and after a multidimensional experiment. However, the application is restricted to linear drifts during the entire experiment.

Our work presented here extends the linear drift compensation (Najbauer and Andreas, 2019) for a general non-linear case with no strict assumptions on the expected progress of the field over time including, e.g., the changes appearing during a helium fill. This is necessary especially for experiments started several hours after sample insertion or after or during refill of liquid helium or nitrogen to the magnet, when long-lasting field perturbations emerge. We propose to monitor the magnetic field for each trace of a multidimensional experiment. On current spectrometers, the field monitoring can proceed as simultaneous acquisition of a frequency reference (SAFR): a reference spectrum with a small-flip-angle pulse is recorded before each scan of the experiment. For high-field magnets, the flip angle can be chosen short enough (we used 1° or less) not to cause any significant impact on the rest of the experiment. Simultaneous acquisition of several free-induction decays (FIDs) in one scan is becoming increasingly popular for efficient combination of several multidimensional spectra into one pulse program (Gallo et al., 2019; Gopinath and Veglia, 2020; Stanek et al., 2020; Kupče and Claridge, 2017; Sharma et al., 2016), while in SAFR, the additional FID serves for the frequency calibration of the main experiment. We present various examples of ^1H - and ^{13}C -

detected 2D and 3D MAS spectra that benefit from SAFR and the subsequent correction of the multidimensional FID. The pulse sequences and the correction program are published along with this article (Římal, 2021).

75 2 Theory

2.1 Phase differences during unstable magnetic field

All our treatment deals with phase-sensitive data, both in direct and indirect dimensions of a multidimensional experiment. In a magnetic field with a time-independent magnitude B_0 , the NMR signal in the time domain (defined by the variable t_{acq}) is expressed as

$$80 \quad S(t_{\text{acq}}) = S_x(t_{\text{acq}}) + iS_y(t_{\text{acq}}), \quad (1)$$

where we identify x and y with the real and imaginary axes, respectively. In a multidimensional experiment, Eq. (1) applies both to direct and indirect acquisition with a general time coordinate t_{acq} . For the direct acquisition of an FID, we will further use the specific symbol t_{dir} for the time t_{acq} (in a 2D typically t_2). Any indirect evolution block will be emphasised by referring to t_{acq} by t_{indir} (in a 2D typically t_1).

85 If the magnetic field changes in time, the detected coherence acquires an additional phase $\Delta\varphi$ relative to the evolution in the constant field B_0 and the detected values are then $S_x^*(t_{\text{acq}})$ and $S_y^*(t_{\text{acq}})$. In the following, we describe how the ideal complex $S(t_{\text{acq}})$ from Eq. (1) can be reconstructed from $S_x^*(t_{\text{acq}})$ and $S_y^*(t_{\text{acq}})$ given that the time dependence of the magnetic field is known – either measured by SAFR as proposed in this work or obtained from another source.

2.2 Correction of the direct dimension

90 We assume that the field change is slow compared to the direct acquisition period, hence the k -th FID of a multidimensional experiment is recorded under a constant field with the magnetic induction $B_0 + \Delta B_k$. The phase difference of the time-domain signal of an isotope with a gyromagnetic ratio γ depends linearly on the time variable t_{dir} :

$$\Delta\varphi_k(t_{\text{dir}}) = \gamma\Delta B_k t_{\text{dir}}. \quad (2)$$

Therefore, the complex experimental data can be corrected using rotations by $-\Delta\varphi_k(t_{\text{dir}})$ for each time point in t_{dir} :

$$95 \quad S(t_{\text{dir}}) = S^*(t_{\text{dir}})e^{-i\Delta\varphi_k(t_{\text{dir}})} = S^*(t_{\text{dir}})e^{-i\gamma\Delta B_k t_{\text{dir}}}, \quad (3)$$

where $S_x^*(t_{\text{dir}})$ and $S_y^*(t_{\text{dir}})$ have been combined into a complex FID:

$$S^*(t_{\text{dir}}) = S_x^*(t_{\text{dir}}) + iS_y^*(t_{\text{dir}}). \quad (4)$$

2.3 Correction of the indirect dimensions

Even after the corrections are done in the direct dimension, one still needs to compensate for the effect of the field change ΔB_k on the spin evolution during all the indirect time periods present in the pulse sequence. For one indirect dimension, there is a particular value of the evolution delay $t_{\text{indir}} = q\Delta t_{\text{indir}}$, which is incremented stepwise within the experiment, connected with the FID k describing the increment q in the indirect dimension. In general, every FID k has a different value of ΔB_k . Because of that, the phase difference of every acquired data point must be treated with its own ΔB_k , such that

$$\Delta\varphi_k' = \gamma_{\text{indir}}\Delta B_k t_{\text{indir}} = \gamma_{\text{indir}}\Delta B_k q\Delta t_{\text{indir}}, \quad (5)$$

where γ_{indir} is the gyromagnetic ratio of the isotope evolving during the indirect evolution period.

In order to be able to rotate the acquired data back to the corrected points $S(t_{\text{indir}})$ accurately, it is important to realize that different fields will be present during the acquisition of the real and imaginary parts of the data that form a hypercomplex point. This can be particularly important for adjacent planes in 3D experiments and yet more for hyperplanes in higher dimensions, where significant wall-clock time has elapsed between the respective acquisitions. According to Eq. (5), a pair of the detected values $S_x^*(t_{\text{indir}})$ and $S_y^*(t_{\text{indir}})$ thus carries unequal phase differences $\Delta\varphi_k'$ and $\Delta\varphi_l'$, respectively, because they are recorded during FIDs k and l . Therefore, $S_x^*(t_{\text{indir}})$ and $S_y^*(t_{\text{indir}})$ no longer belong to a single complex number as for the direct dimension in Eq. (4), but are x and y coordinates of two points S_k^* and S_l^* , respectively, which are images of S rotated by angles $\Delta\varphi_k'$ and $\Delta\varphi_l'$ (Fig. 1). For States and States-TPPI (time-proportional phase incrementation) modes (States et al., 1982; Bodenhausen et al., 1984) with t_{indir} (and q as well) being the same for k and l , however, their amplitude is maintained (i.e., $|S_k^*| = |S_l^*|$). S_x^* and S_y^* can be viewed as the coordinates of the point S in a frame with x -axis rotated by $-\Delta\varphi_k'$ and y -axis by $-\Delta\varphi_l'$, which is obtained after a transformation by a generally non-orthogonal matrix

$$\mathbf{R} = \begin{pmatrix} \cos\Delta\varphi_k' & -\sin\Delta\varphi_k' \\ \sin\Delta\varphi_l' & \cos\Delta\varphi_l' \end{pmatrix}. \quad (6)$$

The desired coordinates of the point S in the original frame, i. e., the values corrected for the field drift, are then obtained by performing a back-transformation:


$$\begin{pmatrix} S_x \\ S_y \end{pmatrix} = \mathbf{R}^{-1} \begin{pmatrix} S_x^* \\ S_y^* \end{pmatrix}, \quad (7)$$

where

$$\mathbf{R}^{-1} = \frac{1}{\cos\Delta\varphi_k'\cos\Delta\varphi_l' + \sin\Delta\varphi_k'\sin\Delta\varphi_l'} \begin{pmatrix} \cos\Delta\varphi_l' & \sin\Delta\varphi_k' \\ -\sin\Delta\varphi_l' & \cos\Delta\varphi_k' \end{pmatrix}. \quad (8)$$

In the special case of $\Delta\varphi_k' = \Delta\varphi_l'$, Eq. (8) reduces to a single orthogonal rotation by $-\Delta\varphi_k'$, consistently with Eq. (3) valid for the direct dimension.

125 For TPPI, S_k^* and S_l^* are acquired with different values of t_{indir} , which renders the above approach invalid. A workaround for TPPI is to convert the data to States–TPPI first, which is possible by forward and backward Fourier transforms with appropriate modes. Without this, mirrored signals appear in the corrected spectra.



130 **Figure 1:** Without the magnetic-field drift, the values S_x and S_y , which form the Cartesian coordinates of the point S , would be measured as the phase-sensitive signal in the indirect dimension under States or States–TPPI mode (red). During an experiment with a time-dependent magnetic field, the x and y coordinates of the points S_k^* and S_l^* are acquired, respectively (blue). S_k^* and S_l^* are images of the point S after rotations by $\Delta\phi_k'$ and $\Delta\phi_l'$, respectively.

2.4 Application of SAFR

135 The field changes ΔB_k for every FID k can be measured by SAFR accompanying a multidimensional experiment. We prepared an AU program “safrcorr”, which can be run directly in Bruker Topspin, that applies the theoretical considerations described above to the experimental data. By this program, the raw data (before Fourier transform) are read. The direct dimension is corrected for every FID according to Eq. (3). For each indirect dimension, the phase differences $\Delta\phi_k'$ are calculated by Eq. (5) and used to correct the data points using Eq. (7). In this way, all the points in the time domain are compensated for the field
 140 drift along all dimensions. A new dataset is created, which can be further processed by standard means.

3 Materials and methods

3.1 Samples

Uniformly ^{13}C – ^{15}N labelled PYRIN-domain filaments of mouse apoptosis-associated speck-like protein containing a caspase-recruitment domain (ASC) were expressed in *E. coli* as described previously (Ravotti et al., 2016). A deuterated and 100 %
 145 back-exchanged ASC sample (dASC) was prepared in the same way except that deuterated water, glucose, and ammonium chloride were used. Exchangeable sites were then re-protonated by solvent exchange with H_2O . Published protocols for preparation of uniformly ^{13}C – ^{15}N labelled samples were followed for Rpo4/7*, a complex of two subunits of RNA polymerase II

Rpo4C36S/Rpo7K123C (Torosyan et al., 2019; Werner and Grohmann, 2011), and for HET-s(218–289) amyloid fibrils from the filamentous fungus *Podospora anserina* (Van Melckebeke et al., 2010).

150

3.2 Solid-state NMR spectroscopy

The NMR experiments were performed using Bruker triple-resonance MAS probeheads on two wide-bore 850 MHz Bruker Avance III HD (running Topspin 3 software) and a standard-bore 1200 MHz Bruker Avance NEO (Topspin 4) spectrometers. ¹H-detected spectra were acquired using 0.7 mm rotors at a 100 kHz MAS rate. ¹³C-detected spectra were acquired using 3.2 mm rotors in E-free probes (Gor'kov et al., 2007). The MAS rates for HET-s(218–289) were 17 kHz and 20 kHz for 850 MHz and 1200 MHz, respectively; the MAS rate for experiments on ¹³C-adamantane was 11 kHz. All spectra were processed in Topspin 4 and presented using CcpNmr Analysis 2 (Stevens et al., 2011, Vranken et al., 2005). Basic acquisition and processing parameters are shown in Tables S2 and S3 in the Supplement.

155

3.3 Pulse programs with SAFR

The SAFR block usually consists of a small-flip-angle pulse on protons and the acquisition of an FID preceding the main experiment (Fig. 2). It is repeated every scan and summed up and stored to disk at the same time as the data of the main experiment, i. e., before the acquisition of the next row of a multidimensional experiment is started. No differences in total experimental time and longitudinal relaxation to equilibrium are introduced since the entire SAFR block is placed during the relaxation delay and the small-flip-angle pulse does not disturb the equilibrium polarization noticeably. There are options in the pulse programs, which switch the SAFR block off (keeping the overall timing unchanged), include decoupling on an additional channel, or acquire a one-dimensional spectrum by the same pulse program. We distinguish homonuclear and heteronuclear SAFR with respect to the nucleus detected in the main experiment. Homonuclear variants work with one receiver and use separated memory buffers (tested on Bruker Avance III HD and NEO, but should be possible on most currently used spectrometers). In contrast, heteronuclear cases need hardware that allows acquisition on multiple receivers, which is possible with modern commercial spectrometers (default on Bruker Avance NEO).

160

165

Representative examples of homonuclear, ¹H-detected pulse sequences are 2D SAFR–hNH and SAFR–hCH correlations with cross-polarization (CP) transfers and MISSISSIPPI water suppression (Zhou and Rienstra, 2008) shown in Fig. 2 (a). A small-flip-angle pulse (we used 0.5° or 1.0°) in the SAFR block, which minimizes signal losses in the main part (Fig. S1 in the Supplement), is sufficient for detecting the water signal in a protein sample. The equivalence of 2D experiments with and without SAFR has been tested (Fig. S2 in the Supplement). Schemes of 2D SAFR–hNH and SAFR–hCH with INEPT transfers and 3D SAFR–hCNH pulse programs can be found in the Supplement (Fig. S3 and Fig. S4, respectively). Even a combination of SAFR with multiple acquisition periods in the main experiment, based on dual-acquisition MAS (DUMAS) hCH–hNH after excitation by simultaneous CP (Gopinath and Veglia, 2020), has been prepared and tested (Fig. S5 in the Supplement). Despite

170

its limited applicability, we also implemented an X-detected SAFR for X-detected experiments, which can be run when multiple receivers are not available (Fig. S6 in the Supplement).

180




Figure 2: Pulse programs for ^1H -detected SAFR (red, flip angle ϑ) before a 2D correlation experiment. (a) CP-based ^1H -detected 2D hCH or hNH. (b) 2D ^{13}C - ^{13}C DARR. Filled black rectangles represent 90-degree pulses, empty rectangles and curved shapes are CP transfers, and grey blocks indicate composite-pulse decoupling (CPD), MISSISSIPPI water suppression (MISS), and the DARR. Optionally, decoupling of the third nucleus can be turned on during t_1 and t_2 in (a).

185

When multiple receivers allow detection on two channels in one pulse sequence, X-detected experiments can be accompanied with SAFR on ^1H . Such a heteronuclear case is demonstrated on a ^{13}C -detected 2D SAFR–DARR, dipolar-assisted rotational resonance (Takegoshi et al., 2001; Takegoshi et al., 2003), in Fig. 2 (b). The same principles as for homonuclear SAFR described above apply. A simple ^{13}C -detected 2D ^1H – ^{13}C correlation and a ^{13}C -detected 3D hNCC involving a ^{13}C DREAM, dipolar recoupling enhanced by amplitude modulation (Verel et al., 2001), are shown in Figures S7 and S8 in the Supplement, respectively, both with ^1H SAFR. ^1H -detected 3D SAFR–hCAco[C, NH] experiment with ^1H SAFR including an additional ^{13}C acquisition yielding a 2D CACO correlation (Gallo et al., 2019) was also tested (Fig. S9 in the Supplement).

190

3.4 Data correction

195

For the calculations described in Sect. 2, we have written an AU program “safrcorr”. The code builds upon the grounds laid for the linear drift compensation by Najbauer and Andreas (2019) with permission from the authors. It can be run directly in Topspin (Bruker), but its core is in C++, allowing modifications to other data formats. The safrcorr program reads the Fourier-transformed reference spectra and finds the global maximum of each. The peak position is refined by a parabolic interpolation through three adjacent data points (Press et al., 2016). In this way, the information about the field drift ΔB_k is obtained for every FID of the main experiment. The time-domain data of the main experiment are then corrected in all dimensions according to Eqs. (3) and (7) and saved to disk under a new experiment number. The chemical shifts of the spectral maxima and their

200

frequency differences relative to the first FID are also displayed in a window and stored in a plain-text file. The time needed for the correction is only a few seconds even for spectra with more than 10000 FIDs. The typical disk space occupied by all data after the correction reaches five times the size of the same experiment without SAFR (the main experiment and the SAFR data acquired, their copies after they are split into two individual datasets, and the corrected data), but most of the processed spectra and also the raw data after the split can be safely deleted to free some storage space. Further description of the program and its use in connection with the SAFR pulse programs can be found in the Supplement.


4 Results

We present here a collection of ^1H - and ^{13}C -detected protein 2D and 3D spectra with SAFR and show the effect of the correction of the field drift. We focus on the non-linear time dependence of the magnetic field. Although strong non-linearities were encountered in the natural drift of the 1.2 GHz magnet several months after its installation (Fig. S10 in the Supplement), the drift at later stages when SAFR was developed showed mostly linear trends during reasonably long experiments. Nevertheless, SAFR is valuable in connection with additional perturbations usually arising, e. g. sample change in probeheads that need to be removed, probehead change, sample-temperature changes, the refill of cryogenic liquids for the magnet, or environmental magnetic field changes (possibly coming from other devices in the proximity of the magnet). Even though parts of these field changes can be reduced by other means, such as a bore-temperature control system, it would require dedicated hardware that can be expensive or unavailable and imperfect; SAFR offers a software solution that is independent of the source and time-course of the perturbations because it uses the information obtained directly from the sample space. In addition, we demonstrate the performance of SAFR during intentionally introduced strong field changes by manipulations with the Z^0 shim current, which serve as a proof of concept of the new method and its general applicability.


4.1 Compensation of thermal effects

Sample cooling is needed to compensate for the temperature rises by friction during fast-MAS experiments and a temperature of the input gas as low as 240 K is needed. Even after reaching an equilibrium within the sample, which can be monitored by the ^1H chemical shifts of H_2O in biological samples (Böckmann et al., 2009; Gottlieb et al., 1997), it can take hours until the body of the probehead and the shim cylinder, as well as the magnet bore, reach their thermal equilibrium temperatures without a system that maintains the temperature of these parts (Malär et al., 2021). A similar effect occurs when a change in the sample temperature is required. Due to the temperature dependence of the magnetic susceptibility and the thermal expansion of the materials used, these cooling or heating processes are inevitably connected with a drift of the magnetic field. Again, the equilibration can take hours without a magnet-bore heater system (Malär et al., 2021).

230 As an example, we measured a 2D SAFR–hNH spectrum of the per-deuterated PYRIN-domain of apoptosis-associated speck-
like protein (Sborgi et al., 2015; Ravotti et al., 2016), dASC, 91 residues, shortly after cooling of the incoming nitrogen gas
from 280 K to 240 K had started. The bore-temperature control system, present and functional on the setup used, was disabled
during this procedure. SAFR with a 0.5° flip angle was used to determine the resonance frequency of water in the sample. The
signal of SAFR is strong enough after a 0.5° pulse and shows that only a negligible part of the protein amide magnetization is
235 excited (see Fig. 3 for a comparison of SAFR and spin-echo 1D ^1H spectra). The correction of the field drift of more than 0.1
ppm clearly improves spectral resolution in both dimensions (Fig. 4). Triangular line shapes that are present in the uncorrected
spectrum acquired during the unstable conditions are eliminated and the spectral resolution is enhanced. After the correction,
the differences to control spectra recorded under constant field (7 and 20 hours later, when the hardware components have
reached their thermal equilibrium) are insignificant and in the same range as between each of the two control experiments
240 performed (Fig. S11 in the Supplement). In this way, employing SAFR to compensate the temperature instabilities allows
shortening the delay needed, after setup of the spectrometer or a sample change, before a high-quality spectrum can be ac-
quired, without the requirement of additional hardware equipment such as a magnet-bore temperature control system. Com-
pared to the linear drift correction, no assumptions on the linearity of the time-dependence of the magnetic field are necessary.




245 **Figure 3:** 1D ^1H spectra of dASC acquired by SAFR after a 0.5° flip angle during hNH (green) and by a spin-echo pulse sequence (black). 256 scans were recorded in both cases (850 MHz). (a) Full spectra scaled to match their maximal intensities. (b) The spectra in (a) multiplied 50-fold in intensities, showing negligible excitation of the amide region in the 0.5° SAFR spectrum.



250 **Figure 4: 2D SAFR-hNH of dASC 36 min after the start of cooling down (850 MHz). (a) The spectrum without (blue) and with (red) the field-drift correction. 1D traces along the horizontal dashed lines are shown at the top. (b), (c) Two selected details of the spectral regions indicated by the dashed rectangles in (a). (d) The evolution of the proton frequency correction.**

4.2 2D acquisition during helium fill at 850 MHz


2D hNH CP experiments were acquired on dASC during the helium filling on an 850 MHz magnet. SAFR using the water resonance of the sample was used to monitor the field drift. The 2D experiment was recorded in four identical blocks of eight
255 scans and the 2D FIDs were summed up before Fourier transform. We demonstrate that even when the individual experimental times were set relatively short (1 h 23 min), non-linearities in the field evolution occur (Fig. 5). The drift mostly affects the positions of the resonances, but the line shapes are distorted as well, visible mainly in experiments 2 and 3 in Fig. 5 (b). Both the positions and the shapes of the peaks are restored after the correction, such that possible differences to a control experiment remain negligible (Fig. S12 in the Supplement). Moreover, the summed spectra in Fig. 5 (a) and (b) emphasize the peak
260 broadening that would be caused by the spectral summation without the drift correction and, at the same time, that SAFR doesn't require chemical-shift calibration of individual spectra when the whole progress of the field drift as in Fig. 5 (c) is known.



265 **Figure 5: 2D SAFR-hNH of dASC during helium fill (850 MHz).** (a) The spectrum obtained from summing raw data (before Fourier
transform) of four experiments together before (blue) and after drift correction (red). 1D traces along the horizontal dashed lines
are shown at the top. (b) A selected peak, indicated by the dashed lines in (a), shown separately for the 2D spectra numbered 1–4
and their sum (Σ) before (blue) and after the drift correction (red). Contours in all the panels are plotted at the same intensities per
scan. (c) The frequency evolution in terms of H_2O chemical-shift difference measured by SAFR. Different colours correspond to the
270 separate 2D experiments 1–4 (each taking 1 h 23 min with 512 rows, 8 scans per row). The magnet was depressurized around row
50, filling started after row 220, and ended before row 600.

4.3 3D acquisition during helium fill on a 1200 MHz magnet

The performance of SAFR was further exploited during a 3D experiment. Figure 6 shows a CP-based 3D SAFR–hCANH spectrum of dASC (pulse program in Fig. S4 in the Supplement) in a 1200 MHz magnet during helium fill. The improvement after the drift correction is obvious. The distortions in the spectrum before the correction would make it useless which is explained by the strong field drift observed in Fig. 6 (b). Note that the field drift stayed significant for two days; without SAFR, this instrument time couldn't have been used for a reliable 3D. The periodic field oscillations in the inset of Fig. 6 (b) are not caused by the actual experiment, but were inherent to the magnet control system at that time and were later improved by the manufacturer. The strongest influence of the drift is seen along the ^{13}C axis, which corresponds to the evolution time that is incremented stepwise after a particular NH plane (122 rows) is acquired, i.e., ^{13}C belongs to the outermost loop in the pulse program. Without SAFR, one could consider that, in ppm units, the resonances are narrower along the ^{13}C dimension than ^{15}N (1.1 ppm and 1.0 ppm line widths in ^{13}C compared to 1.4 ppm and 2.0 ppm in ^{15}N for the two peaks shown in Fig. 6, respectively), which makes ^{13}C more sensitive to the effects of the field drift. Therefore, the uncorrected spectrum would in principle slightly profit from exchanging the order of the loops. Whereas this and other alternative approaches, such as splitting the experiment into several blocks with smaller number of transients and varying the dimension order, would bring only minor improvements to the uncorrected spectrum but would still lead to line broadening after averaging similarly as in Fig. 5 discussed above, SAFR and the subsequent spectral correction are independent on these technical adjustments of the pulse program and yield the same final results.



290

Figure 6: 3D SAFR–hCANH of dASC during and after helium fill (1200 MHz). (a) Two selected peaks and their surroundings viewed in three possible plane orientations before (blue) and after the drift correction (red). The perpendicular cross-sections are taken at the positions of the peaks after the correction, indicated by the dashed red lines. Differently to this, the ^{15}N – ^{13}C plane of the uncorrected spectrum is taken along the dashed blue lines. (b) The frequency evolution in terms of H_2O chemical shift difference measured by SAFR. Insets show details of the regions marked by rectangles (different aspect ratios). The helium filling started before the acquisition began.

295

4.4 2D acquisition during nitrogen fill at 1200 MHz

The 2D SAFR–hCH of a complex of two subunits of archaeal RNA polymerase II (Torosyan et al., 2019), Rpo4/7*, where only Rpo7 (187 residues) was ^{13}C and ^{15}N labelled, in Fig. 7 shows an application where no strong field drift was anticipated but turned out to be present during the experiment. Generally, it is considered safe to measure solid-state spectra during the refill of liquid nitrogen, but a field change of around 0.1 ppm was observed. Because it happened mostly in the second half of the acquisition time, its overall influence on the spectrum is not that strong. Nevertheless, the narrowest peaks are affected, as amplified by the strongly resolution-enhanced processing in Fig. 7: errors in resonance positions in multiples of 0.01 ppm and even false peak doubling (panel b of Fig. 7) can arise when the drift is not compensated for.

305




Figure 7: 2D SAFR–hCH of Rpo4/7* during nitrogen fill (1200 MHz) with processing-enhanced resolution (shifted squared sine bell with parameter SSB = 5 in both dimensions). (a) C_{α} – H_{α} region of 2D hCH spectrum before (blue) and after the drift correction (red). (b) A detail of the 2D hCH spectrum outside of the region shown in (a). Peak maxima of the corrected spectrum are marked by red crosses. (c) The frequency evolution in terms of H_2O chemical-shift difference measured by SAFR. The nitrogen filling started around row 400.

4.5 Limitations

The correction scheme assumes that the solvent (water) chemical shift is constant. It is however known that it is temperature-dependent and changes by 0.0111 ppm/°C (Gottlieb et al., 1997), or 13.3 Hz/°C on a 1200 MHz system. The temperature of the sample must therefore be kept constant within a fraction of a degree (or several degrees, depending on the spectral line-widths) to avoid broadening effects or t_1 noise. Of course, similar requirements apply to the conventional field–frequency locks. We have checked that the application of 1H -SAFR and subsequent drift correction under ordinary conditions do not introduce extra broadening or noise (Fig. S13 in the Supplement) and that a purely linear field drift is compensated for without any differences to the linear drift correction previously published by Najbauer and Andreas (2019), as shown in the Supplement (Fig. S14).

Although we expect that the drift correction is mostly needed for 1H -detected spectroscopy, there might be relatively rare cases of its usage accompanying X-nucleus detection. 1H -SAFR can also be used, but only on spectrometers capable of detection on

multiple receivers during one experiment. For single-receiver equipment, ^{13}C -SAFR is in principle possible for strong and non-linear field changes but can lead to t_1 noise after a correction relying on a broad reference resonance (Figures S15, S16, and S17 in the Supplement).

Furthermore, we assume that the perturbations to be corrected are slow and no changes occur within the duration of a single FID. In addition, we corrected each FID of an experiment (summed up after several scans needed for phase cycling and improvement of the signal-to-noise ratio) as a whole entity. At the cost of higher storage demand and additional pre-processing steps, each scan can be corrected individually in principle, but this did not turn out to be necessary (Figures S18 and S19 in the Supplement show an extreme case with an artificially created jump in the field value).

5 Discussion and conclusions

We have presented a new strategy to eliminate line-broadening and spectral artefacts arising from the instabilities of the magnetic field in high-resolution solid-state biomolecular spectroscopy. By adding a small-flip-angle experiment at the beginning of the pulse sequence, we obtain a frequency reference (SAFR) which can be used to correct digitally the FIDs. An automated procedure facilitating these steps as well as some example pulse programs commonly employed in biological solid-state NMR are developed and made available. In its principle and outcome, SAFR is equivalent to a lock system: it detects the position of the solvent (or other intense and preferably narrow) peak and corrects the acquired data for the corresponding frequency shift. The main difference is that the correction using SAFR is performed after, not during the acquisition. Our work has demonstrated the importance of a lock in various 2D and 3D correlation experiments under different conditions: an often encountered challenge is the insertion of a sample to the probehead and its temperature change, which can delay the start of a conventional acquisition by several hours. SAFR overcomes this problem without the need for a hardware solution, such as a bore-temperature control system. Next, the filling of the cryogenic liquids into the magnet induces non-linear transient field disturbances, which last for hours or even days. Again, their effects on the NMR spectra can be corrected by SAFR. Finally, the field strength is affected by external fields or temperature changes in the room that are unpredictable, but SAFR will be helpful in these cases as well.

SAFR requires a well-resolved and intense resonance line in the spectrum. The ^1H resonance of solvent water present in hydrated microcrystalline or sedimented protein samples typically serves this purpose (Böckmann et al., 2009; Lacabanne et al., 2019). Alternatively, any strong and sufficiently narrow peak can be used as the reference. As the water resonance is temperature dependent, the sample temperature needs to be stable during the experiments within about 1 °C. Normally, one FID (one row of a multidimensional experiment with its number of scans) is corrected with a single frequency value. If a significant field fluctuation appears within this time, one could store the SAFR spectra for every scan (or a chosen number of scans) and apply SAFR to this entity. We have not implemented this as it seems to be of marginal interest, realizing that most

of such circumstances could be treated by separating the experiment in shorter identical blocks with lower numbers of scans. Besides this, we note that peaks that are aliased or folded along an indirect dimension cannot be accurately corrected by SAFR; a proper designation and a separate treatment of such resonances would be necessary.

In our experience, highly resolved ¹H-detected 2D and 3D spectra benefit from the use of the drift correction the most. Indeed, SAFR significantly increases the available amount of the precious spectrometer time. Multidimensional NMR experiments can be safely recorded even during the helium or nitrogen refill. Although designed specifically for cases known to cause field changes, SAFR can be safely and without additional assumptions used even when no drifts are anticipated to ensure unbiased results of the experiment.

6 Code availability

All the pulse programs listed in Table S1 in the Supplement and the AU program safrcorr are freely available at <https://doi.org/10.5905/ethz-1007-453> (Římal, 2021) under the Apache License, Version 2.0. Details about the composition of the programs and a typical experimental workflow are described in the Supplement.

7 Data availability

The experimental data are available at <https://doi.org/10.3929/ethz-b-000522147> (Římal and Meier, 2021).

8 Supplement link

the link to the supplement will be included by Copernicus

9 Author contribution

VŘ, TW, ME, AB, and BHM contributed to the concept of the method. VŘ performed, analysed, and visualized the experiments and wrote the pulse and AU programs. MC and AAM carried out the experiment on Rpo4/7*. VŘ, MC, AAM, and TW tested the software. RC expressed the ASC and HET-s(218–289) protein fibrils and AT the Rpo4/7* protein complex. TW analysed the field drift of the 1.2 GHz magnet. BHM conceived and supervised the project. VŘ and BHM prepared the original draft. All authors reviewed and edited the article.

375 10 Competing interests

ME is an executive editor and AB and BHM are members of the editorial board of Magnetic Resonance. The peer-review process was guided by an independent editor, and the authors have also no other competing interests to declare. The other authors declare that they have no conflict of interest.

11 Financial support

380 The research has been supported by an ERC Advanced Grant (BHM, grant number 741863, Faster), by the Swiss National Science Foundation (BHM, grant numbers 200020_159707 and 200020_188711), an ETH Research Grant (TW, grant number ETH-43 17-2), the French Agence Nationale de Recherches sur le Sida et les hépatites virales (ANRS, ECTZ71388 & ECTZ100488), and the CNRS (CNRS-Momentum 2018), the LABEX ECOFECT (ANR-11-LABX-0048) within the Université de Lyon program Investissements d’Avenir (ANR-11-IDEX-0007).

385 12 References

- Böckmann, A., Gardiennet, C., Verel, R., Hunkeler, A., Loquet, A., Pintacuda, G., Emsley, L., Meier, B. H. and Lesage, A.: Characterization of different water pools in solid-state NMR protein samples, *J. Biomol. NMR*, 45(3), 319–327, <https://doi.org/10.1007/s10858-009-9374-3>, 2009.
- Bodenhausen, G., Kogler, H. and Ernst, R. R.: Selection of coherence-transfer pathways in NMR pulse experiments, *J. Magn. Reson.*, 58(3), 370–388, [https://doi.org/10.1016/0022-2364\(84\)90142-2](https://doi.org/10.1016/0022-2364(84)90142-2), 1984.
- 390 Callon, M., Malär, A. A., Pfister, S., Římal, V., Weber, M. E., Wiegand, T., Zehnder, J., Chávez, M., Cadalbert, R., Deb, R., Däpp, A., Fogeron, M.-L., Hunkeler, A., Lecoq, L., Torosyan, A., Zyla, D., Glockshuber, R., Jonas, S., Nassal, M., Ernst, M., Böckmann, A. and Meier, B. H.: Biomolecular solid-state NMR spectroscopy at 1200 MHz: the gain in resolution, *J. Biomol. NMR*, 75(6–7), 255–272, <https://doi.org/10.1007/s10858-021-00373-x>, 2021.
- 395 Gallo, A., Franks, W. T. and Lewandowski, J. R.: A suite of solid-state NMR experiments to utilize orphaned magnetization for assignment of proteins using parallel high and low gamma detection, *J. Magn. Reson.*, 305, 219–231, <https://doi.org/10.1016/j.jmr.2019.07.006>, 2019.
- Gopinath, T. and Veglia, G.: Proton-detected polarization optimized experiments (POE) using ultrafast magic angle spinning solid-state NMR: Multi-acquisition of membrane protein spectra, *J. Magn. Reson.*, 310, 106664, <https://doi.org/10.1016/j.jmr.2019.106664>, 2020.
- 400 Gor’kov, P. L., Witter, R., Chekmenev, E. Y., Nozairov, F., Fu, R. and Brey, W. W.: Low-E probe for ^{19}F - ^1H NMR of dilute biological solids, *J. Magn. Reson.*, 189(2), 182–189, <https://doi.org/10.1016/j.jmr.2007.09.008>, 2007.

- Gottlieb, H. E., Kotlyar, V. and Nudelman, A.: NMR Chemical Shifts of Common Laboratory Solvents as Trace Impurities, *J. Org. Chem.*, 62(21), 7512–7515, <https://doi.org/10.1021/jo971176v>, 1997.
- 405 Henry, P.-G., van de Moortele, P.-F., Giacomini, E., Nauwerth, A. and Bloch, G.: Field-frequency locked in vivo proton MRS on a whole-body spectrometer, *Magn. Reson. Med.*, 42(4), 636–642, [https://doi.org/10.1002/\(SICI\)1522-2594\(199910\)42:4<636::AID-MRM4>3.0.CO;2-I](https://doi.org/10.1002/(SICI)1522-2594(199910)42:4<636::AID-MRM4>3.0.CO;2-I), 1999.
- Kupče, Ě. and Claridge, T. D. W.: NOAH: NMR Supersequences for Small Molecule Analysis and Structure Elucidation, *Angew. Chemie Int. Ed.*, 56(39), 11779–11783, <https://doi.org/10.1002/anie.201705506>, 2017.
- 410 Kupče, Ě. and Freeman, R.: Molecular structure from a single NMR sequence (fast-PANACEA), *J. Magn. Reson.*, 206(1), 147–153, <https://doi.org/10.1016/j.jmr.2010.06.018>, 2010.
- Lacabanne, D., Fogeron, M.-L., Wiegand, T., Cadalbert, R., Meier, B. H. and Böckmann, A.: Protein sample preparation for solid-state NMR investigations, *Prog. Nucl. Magn. Reson. Spectrosc.*, 110, 20–33, <https://doi.org/10.1016/j.pnmrs.2019.01.001>, 2019.
- 415 Malär, A. A., Völker, L. A., Cadalbert, R., Lecoq, L., Ernst, M., Böckmann, A., Meier, B. H. and Wiegand, T.: Temperature-Dependent Solid-State NMR Proton Chemical-Shift Values and Hydrogen Bonding, *J. Phys. Chem. B*, 125(23), 6222–6230, <https://doi.org/10.1021/acs.jpcc.1c04061>, 2021.
- Minkler, M. J., Kim, J. M., Shinde, V. V. and Beckingham, B. S.: Low-field ¹H NMR spectroscopy: Factors impacting signal-to-noise ratio and experimental time in the context of mixed microstructure polyisoprenes, *Magn. Reson. Chem.*, 58(12), 1168–
- 420 1176, <https://doi.org/10.1002/mrc.5022>, 2020.
- Najbauer, E. E. and Andreas, L. B.: Correcting for magnetic field drift in magic-angle spinning NMR datasets, *J. Magn. Reson.*, 305, 1–4, <https://doi.org/10.1016/j.jmr.2019.05.005>, 2019.
- Near, J., Edden, R., Evans, C. J., Paquin, R., Harris, A. and Jezzard, P.: Frequency and phase drift correction of magnetic resonance spectroscopy data by spectral registration in the time domain, *Magn. Reson. Med.*, 73(1), 44–50,
- 425 <https://doi.org/10.1002/mrm.25094>, 2015.
- Nimerovsky, E., Movellan, K. T., Zhang, X. C., Forster, M. C., Najbauer, E., Xue, K., Dervişođlu, R., Giller, K., Griesinger, C., Becker, S. and Andreas, L. B.: Proton Detected Solid-State NMR of Membrane Proteins at 28 Tesla (1.2 GHz) and 100 kHz Magic-Angle Spinning, *Biomolecules*, 11(5), 752, <https://doi.org/10.3390/biom11050752>, 2021.
- Paulson, E. K. and Zilm, K. W.: External field-frequency lock probe for high resolution solid state NMR, *Rev. Sci. Instrum.*,
- 430 76(2), 026104, <https://doi.org/10.1063/1.1841972>, 2005.
- Penzel, S., Oss, A., Org, M. L., Samoson, A., Böckmann, A., Ernst, M. and Meier, B. H.: Spinning faster: protein NMR at MAS frequencies up to 126 kHz, *J. Biomol. NMR*, 73(1–2), 19–29, <https://doi.org/10.1007/s10858-018-0219-9>, 2019.
- Press, W. H., Teukolsky, S. A., Vetterling, W. T. and Flannery, B. P.: Minimization or Maximization of Functions, in *Numerical Recipes in C: The Art of Scientific Computing*, pp. 394–455, Cambridge University Press, Cambridge, 1992.

- 435 Ravotti, F., Sborgi, L., Cadalbert, R., Huber, M., Mazur, A., Broz, P., Hiller, S., Meier, B. H. and Böckmann, A.: Sequence-specific solid-state NMR assignments of the mouse ASC PYRIN domain in its filament form, *Biomol. NMR Assign.*, 10(1), 107–115, <https://doi.org/10.1007/s12104-015-9647-6>, 2016.
- Římal, V.: Pulse and AU programs for correcting field instabilities in NMR by simultaneous acquisition of a frequency reference (SAFR), ETH Zurich [code], <https://doi.org/10.5905/ethz-1007-453>, 2021.
- 440 Římal, V. and Meier, B. H.: Experimental Data for Correction of Field Instabilities in Biomolecular Solid-State NMR by Simultaneous Acquisition of a Frequency Reference, ETH Zurich [data set], <https://doi.org/10.3929/ethz-b-000522147>, 2021.
- Sborgi, L., Ravotti, F., Dandey, V. P., Dick, M. S., Mazur, A., Reckel, S., Chami, M., Scherer, S., Huber, M., Böckmann, A., Egelman, E. H., Stahlberg, H., Broz, P., Meier, B. H. and Hiller, S.: Structure and assembly of the mouse ASC inflammasome by combined NMR spectroscopy and cryo-electron microscopy, *Proc. Natl. Acad. Sci.*, 112(43), 13237–13242, <https://doi.org/10.1073/pnas.1507579112>, 2015.
- 445 Sharma, K., Madhu, P. K. and Mote, K. R.: A suite of pulse sequences based on multiple sequential acquisitions at one and two radiofrequency channels for solid-state magic-angle spinning NMR studies of proteins, *J. Biomol. NMR*, 65(3–4), 127–141, <https://doi.org/10.1007/s10858-016-0043-z>, 2016.
- Stanek, J., Schubeis, T., Paluch, P., Güntert, P., Andreas, L. B. and Pintacuda, G.: Automated Backbone NMR Resonance Assignment of Large Proteins Using Redundant Linking from a Single Simultaneous Acquisition, *J. Am. Chem. Soc.*, 142(12), 5793–5799, <https://doi.org/10.1021/jacs.0c00251>, 2020.
- 450 States, D. ., Haberkorn, R. . and Ruben, D. .: A two-dimensional nuclear Overhauser experiment with pure absorption phase in four quadrants, *J. Magn. Reson.*, 48(2), 286–292, [https://doi.org/10.1016/0022-2364\(82\)90279-7](https://doi.org/10.1016/0022-2364(82)90279-7), 1982.
- Stevens, T. J., Fogh, R. H., Boucher, W., Higman, V. A., Eisenmenger, F., Bardiaux, B., Van Rossum, B. J., Oschkinat, H. and Laue, E. D.: A software framework for analysing solid-state MAS NMR data, *J. Biomol. NMR*, 51(4), 437–447, <https://doi.org/10.1007/s10858-011-9569-2>, 2011.
- 455 Takahashi, M., Ebisawa, Y., Tennmei, K., Yanagisawa, Y., Hosono, M., Takasugi, K., Hase, T., Miyazaki, T., Fujito, T., Nakagome, H., Kiyoshi, T., Yamazaki, T. and Maeda, H.: Towards a beyond 1 GHz solid-state nuclear magnetic resonance: External lock operation in an external current mode for a 500 MHz nuclear magnetic resonance, *Rev. Sci. Instrum.*, 83(10), 105110, <https://doi.org/10.1063/1.4757576>, 2012.
- 460 Takegoshi, K., Nakamura, S. and Terao, T.: ^{13}C – ^1H dipolar-assisted rotational resonance in magic-angle spinning NMR, *Chem. Phys. Lett.*, 344(5–6), 631–637, [https://doi.org/10.1016/S0009-2614\(01\)00791-6](https://doi.org/10.1016/S0009-2614(01)00791-6), 2001.
- Takegoshi, K., Nakamura, S. and Terao, T.: ^{13}C – ^1H dipolar-driven ^{13}C – ^{13}C recoupling without ^{13}C rf irradiation in nuclear magnetic resonance of rotating solids, *J. Chem. Phys.*, 118(5), 2325–2341, <https://doi.org/10.1063/1.1534105>, 2003.

- 465 Thiel, T., Czisch, M., Elbel, G. K. and Hennig, J.: Phase coherent averaging in magnetic resonance spectroscopy using interleaved navigator scans: Compensation of motion artifacts and magnetic field instabilities, *Magn. Reson. Med.*, 47(6), 1077–1082, <https://doi.org/10.1002/mrm.10174>, 2002.
- Torosyan, A., Wiegand, T., Schledorn, M., Klose, D., Güntert, P., Böckmann, A. and Meier, B. H.: Including Protons in Solid-State NMR Resonance Assignment and Secondary Structure Analysis: The Example of RNA Polymerase II Subunits Rpo4/7, 470 *Front. Mol. Biosci.*, 6(October), 1–8, <https://doi.org/10.3389/fmolb.2019.00100>, 2019.
- Van Melckebeke, H., Wasmer, C., Lange, A., AB, E., Loquet, A., Böckmann, A. and Meier, B. H.: Atomic-Resolution Three-Dimensional Structure of HET-s(218–289) Amyloid Fibrils by Solid-State NMR Spectroscopy, *J. Am. Chem. Soc.*, 132(39), 13765–13775, <https://doi.org/10.1021/ja104213j>, 2010.
- Verel, R., Ernst, M. and Meier, B. H.: Adiabatic dipolar recoupling in solid-state NMR: The DREAM scheme, *J. Magn. Reson.*, 475 150(1), 81–99, <https://doi.org/10.1006/jmre.2001.2310>, 2001.
- Vionnet, L., Aranovitch, A., Duerst, Y., Haeberlin, M., Dietrich, B. E., Gross, S. and Pruessmann, K. P.: Simultaneous feedback control for joint field and motion correction in brain MRI, *Neuroimage*, 226, 117286, <https://doi.org/10.1016/j.neuroimage.2020.117286>, 2021.
- Vranken, W. F., Boucher, W., Stevens, T. J., Fogh, R. H., Pajon, A., Llinas, M., Ulrich, E. L., Markley, J. L., Ionides, J. and 480 Laue, E. D.: The CCPN data model for NMR spectroscopy: Development of a software pipeline, *Proteins Struct. Funct. Bioinforma.*, 59(4), 687–696, <https://doi.org/10.1002/prot.20449>, 2005.
- Werner, F. and Grohmann, D.: Evolution of multisubunit RNA polymerases in the three domains of life, *Nat. Rev. Microbiol.*, 9(2), 85–98, <https://doi.org/10.1038/nrmicro2507>, 2011.
- Zhou, D. H. and Rienstra, C. M.: High-performance solvent suppression for proton detected solid-state NMR, *J. Magn. Reson.*, 485 192(1), 167–172, <https://doi.org/10.1016/j.jmr.2008.01.012>, 2008.