# Time-domain R-PDLF NMR for molecular structure determination in complex lipid membranes

Anika Wurl<sup>1</sup>, Kay Saalwächter<sup>1</sup>, and Tiago Mendes Ferreira<sup>1</sup>

<sup>1</sup>NMR group, Institute for Physics, Martin Luther University Halle-Wittenberg

**Correspondence:** Tiago Mendes Ferreira (tiago.ferreira@physik.uni-halle.de)

**Abstract.** Proton-detected local field (PDLF) NMR spectroscopy, using magic-angle spinning and dipolar recoupling, is presently the most powerful experimental technique for obtaining atomistic structural information from small molecules undergoing anisotropic motion. Common examples include peptides, drugs, or lipids in model membranes and molecules that form liquid crystals. The measurements on complex systems are however compromised by the larger number of transients required. Retaining sufficient spectral quality in the direct dimension requires that the indirect time-domain modulation becomes too short for yielding dipolar splittings in the frequency domain. In such cases the dipolar couplings can be obtained by fitting the experimental data, however ideal models often fail to fit the data properly. Here, we present a new methodology for analysing R-symmetry based PDLF NMR experiments that significantly improves fitting accuracy and therefore enables to address more complex systems. The new methodology consists of fitting the time-domain data with NMR simulations accounting for RF spatial inhomogeneity. This makes possible the following three important improvements: the use of shorter experiments that enable to measure samples with a higher level of complexity; the measurement of C-H bond order parameters with smaller magnitudes  $|S_{CH}|$  and of smaller variations of  $|S_{CH}|$  caused by perturbations of the system; and the determination of  $|S_{CH}|$ values with small differences from distinct sites having the same chemical shift. The increase in accuracy is demonstrated by comparison with <sup>2</sup>H NMR quadrupolar echo experiments on mixtures of deuterated and non-deuterated dimyristoylphosphatidylcholine (DMPC). The new method presented enables an unprecedented level of structural detail and will be highly useful for investigating complex membrane systems as exemplified here with membranes composed of a brain lipid extract with many distinct lipid types.

## 1 Introduction

The methodology for characterizing the molecular structure in biological systems has been advancing rapidly over recent years (Cheng, 2018; Gauto et al., 2019; Wu and Lander, 2020). Among the various experimental techniques available, solid-state NMR spectroscopy provides the most powerful methods for investigating the molecular structure of smaller molecules (<30 kDa) undergoing anisotropic motion in membranes such as lipids, peptides and membrane proteins (Andersson et al., 2017; Löser et al., 2018; Bacle et al., 2021; Zerweck et al., 2017; Aisenbrey et al., 2019; Cady et al., 2007; Park et al., 2012; Mandala et al., 2020; Umegawa et al., 2018; Mandala et al., 2018).

Of these techniques, researchers have consistently used <sup>2</sup>H NMR spectroscopy since the 1970s for the molecular structural characterization of both lipids and peptides (Seelig, 1977; Davis, 1983; Strandberg and Ulrich, 2004; Leftin and Brown, 2011). The popularity of this method is due to both its simplicity and the high accuracy of the experiments. However, it requires specific <sup>2</sup>H isotopic labelling for site resolution. This requirement severely limits its application to the investigation of biological extracts or complex model membranes that mimic biological systems. Alternatives to <sup>2</sup>H NMR are separated local field (SLF) 2D NMR experiments that make use of <sup>1</sup>H–<sup>13</sup>C heteronuclear dipolar recoupling during the indirect time (Hester et al., 1976). This technique is very advantageous compared to <sup>2</sup>H NMR because SLF delivers the same type of information with <sup>13</sup>C chemical shift selectivity and does not require isotopic labelling. These SLF <sup>1</sup>H–<sup>13</sup>C heteronuclear dipolar recoupling NMR techniques fall into two groups, carbon-detected local field (CDLF) and proton-detected local field (PDLF) experiments with the latter having enhanced resolution (Nakai and Terao, 1992; Schmidt-Rohr et al., 1994; Bärenwald et al., 2016). Here, we discuss only PDLF techniques which provide C–H bond site resolution and are therefore directly comparable to <sup>2</sup>H NMR.

25

A number of PDLF pulse sequences have been implemented to investigate lipid membranes as well as other systems (Griffin, 1998; De Paëpe, 2012; Molugu et al., 2017). The main difference between the distinct PDLF sequences reported are the heteronuclear dipolar recoupling blocks used to counteract the effect of magic angle spinning (MAS) on the anisotropic Hamiltonian terms during the indirect dimension. Several recoupling pulse sequences have been designed for this purpose such as in REDOR (Gullion), DROSS (Gross et al., 1997; Leftin et al., 2014), R-symmetry based pulse sequences (Levitt, 2007; Dvinskikh et al., 2004; Hou et al., 2011) and others (De Paëpe, 2012). In recent years, a R-symmetry based PDLF sequence dubbed R-PDLF designed by Dvinskikh and coworkers (Dvinskikh et al., 2004), with R18<sup>7</sup> recoupling blocks, has been frequently used to successfully investigate lipid membranes (Dvinskikh et al., 2005; Ferreira et al., 2013; Löser et al., 2018; Bacle et al., 2021; Fridolf et al., 2022). The R-symmetry sequences are particularly advantageous since they enable to simultaneously recouple the heteronuclear and decouple the homonuclear dipolar interactions. Moreover, R-symmetry recoupling can be applied directly to an IS spin ensemble in thermal equilibrium with no need of generating transverse magnetization before the recoupling blocks and therefore avoiding potential  $T_2$  losses during the recoupling period, i.e. increasing sensitivity. It is well known that, as with other dipolar recoupling pulse sequences, R-symmetry recoupling is rather sensitive to RF imperfections (Nishimura et al., 2001; Schanda et al., 2011) and therefore to the RF spatial inhomogeneity across the sample investigated – one unavoidable and unique feature of the MAS probe used to perform experiments (Tošner et al., 2017). To minimize this problem, windowed R-symmetry sequences have been proposed that are not as sensitive to RF inhomogeneity as the original windowless R-symmetry sequences (Gansmüller et al., 2013; Lu et al., 2016). These windowed versions, however, require using higher RF radiofrequency fields which is a practical bottleneck due to hardware limitations.

In general, the strategy for developing better dipolar recoupling methods has been to minimize the sensitivity to RF spatial inhomogeneity, by using/designing dipolar recoupling sequences that are less sensitive to RF imperfections and reducing the sample studied to a narrow volume, as well as using recoupling pulse sequences with a high scaling factor that enable to measure smaller dipolar couplings (Dvinskikh et al., 2005; Chevelkov et al., 2009; Schanda et al., 2011; Gansmüller et al., 2013; Lu et al., 2016; Asami and Reif, 2017; Jain et al., 2019; Nimerovsky and Soutar, 2020). This last point is of significant importance, since higher scaling factors enable to obtain dipolar splittings in the indirect frequency domain with a reduced

maximum time in the indirect dimension. PDLF experiments have been mostly analysed by "reading off" the dipolar splittings obtained in the indirect frequency domain of the 2D spectra. In contrast to the standard analysis of REDOR or DIPSHIFT dipolar evolutions (Gullion; deAzevedo et al., 2008), PDLF measurements have been rarely analysed by employing fitting procedures. The few TD analysis attempts reported assumed samples under a homogeneous RF field (Gross et al., 1997; Gansmüller et al., 2013; Fridolf et al., 2022), however, it is well known that dipolar recoupling sequences are highly sensitive to RF inhomogeneity (Schanda et al., 2011).

Here, we present a new approach that enables to increase the accuracy and applicability of R-symmetry based PDLF NMR considerably through time-domain analysis (rather than frequency domain) of the dipolar modulation by explicitly taking into account the RF spatial inhomogeneity of the used probe explicitly, i.e., rather than trying to minimize the effect of RF spatial inhomogeneity we include it in the NMR simulations used to fit the experimental measurements. The use of RF inhomogeneity for data analysis has been done recently for investigating protein molecular structure (Xue et al., 2022). Here we apply this strategy to analyse R-PDLF experiments that were proven suitable to investigate molecular structure in lipid membranes. The novelty of our new methodology presented is solely in the new experimental analysis procedure, no special hardware or different setup of the experiment is needed. We demonstrate the method on a DMPC/DMPCd sample to allow comparison of the <sup>1</sup>H-<sup>13</sup>C dipolar couplings determined with a R18<sup>7</sup><sub>1</sub>-type PDLF sequence (R-PDLF (Dvinskikh et al., 2004, 2005)) with the C-D bond order parameters determined from <sup>2</sup>H quadrupolar couplings. We then implemented our new method to investigate a system of complex lipid membranes composed of a brain lipid extract. The new method enabled us to measure several dipolar couplings from the membrane hydrophilic layer composed of a set of many types of distinct phospholipid headgroups. We believe that this new strategy is a highly useful tool in structural biology for investigating both model systems and complex membrane systems with a low signal to noise ratio such as biological lipid extracts. In addition it has the potential for investigating in-cell molecular structures.

#### 2 Methodological Framework

The typical approach for determining C-H bond order parameter magnitudes with PDLF dipolar recoupling NMR is by performing a Fourier transform,

$$I(\omega_1) = \int_0^{+\infty} s(t_1) \exp(-i\omega_1 t_1) dt_1 \tag{1}$$

of the time-domain recoupled dipolar modulation over the indirect dimension,  $s(t_1)$ , and measuring the splitting(s) of the resulting spectral dipolar line shape (Dvinskikh et al., 2004; Gross et al., 1997). If the scaling factor,  $\kappa$ , for the particular pulse sequence under use is known, determining the magnitude of a C-H bond order parameter from the splitting(s) observed is straightforward by using

$$|S_{\rm CH}| = \kappa^{-1} \frac{\Delta \nu}{d_{\rm CH}} = \left| \frac{1}{2} \left\langle 3\cos^2 \theta - 1 \right\rangle \right| \tag{2}$$




Figure 1. The effect of using a limited number of time-domain points to describe a superposition of two Pake patterns with distinct splittings.

where  $d_{\rm CH}$  is the magnitude of the rigid dipolar coupling for a static C-H bond equal to  $\approx$ 22 kHz, the angle  $\theta$  is between the C-H bond and the symmetry axis of its uniaxial motion (in case of lipid membranes this is the bilayer normal), and the angular brackets denote a time average over a time interval up to approximately the inverse of the splitting measured.

95

100

For a dipolar modulation corresponding to a single C-H bond order parameter and acquired under optimal experimental conditions, the use of equation 2 allows accurate values to be obtained provided that the digitised signal,  $s(t_1)$ , extends to a time longer than the reciprocal of the scaled dipolar coupling frequency (e.g. 4 times the reciprocal gives approximately 10% error). If  $s(t_1)$  is instead a superposition of two distinct contributions with different  $S_{CH}$ , for resolving the two components in the spectrum, the signal acquisition must be, at least, longer than half of a beat period defined by the frequencies of the crystallite orientations responsible for the two splittings (Lindon and Ferrige, 1980). This is illustrated in Figure 1 for a superposition of two Pake patterns. Therefore, both for single-component dipolar modulations with low order parameters and for two-component dipolar modulations with a small frequency difference between the two components, a large number of indirect dimension points is needed in 2D dipolar recoupling experiments which severely increases the experimental time, especially for samples requiring the acquisition of a large number of transients in the direct dimension. Moreover, there is a maximum limit to the number of indirect dimension points acquired both due to RF heating of samples and to the need of maintaining the integrity of the NMR probe.




Figure 2. Numerical simulations of the R-PDLF pulse sequence for a fixed dipolar coupling displaying the effects of RF nutation frequency deviation from ideal settings (left) and  $^{1}$ H chemical shift offset (right) given in ppm units. The solid red and faint red curves correspond to the dipolar modulation with ideal settings which yield a dipolar splitting equal to  $0.315S_{\rm CH}d_{\rm CH}$ . Negative and positive chemical offset shifts with the same magnitude have exactly the same effect. All simulations in this work were performed for a  $^{1}$ H Larmor frequency of 400 MHz.

105

115

Time-domain analysis can be used to circumvent the limitations outlined above entirely. The practical bottleneck for time-domain analysis is the RF inhomogeneity across the sample, intrinsic to the majority of experimental setups (Tošner et al., 2017), combined with the sensitivity of recoupling pulse sequences to RF strength accuracy and the effect of chemical shift offsets (Schanda et al., 2011; Lu et al., 2016). Figure 2 illustrates such dependencies for the R-PDLF pulse sequence (Dvinskikh et al., 2004), showing a non-linear dependence of the dipolar modulation on the RF pulse power level deviation from the ideal value and on the <sup>1</sup>H chemical shift offset. The effect of RF miscalibration becomes rather pronounced at RF frequencies higher or lower than 6% of the ideal frequency. This means that in typical experiments, for which the RF spatial inhomogeneity profile may reach 80% of the maximum value in the outer parts of the sample (Tošner et al., 2017), one should expect deviations from the ideal dipolar lineshape. The effect of the <sup>1</sup>H chemical shift offset is almost negligible up to 4 ppm above which large deviations from the ideal behaviour occur. For achieving optimal resolution in an R-PDLF experiment the chemical shift offset should therefore be no more than 4 ppm at a magnetic field inducing a Larmor frequency for <sup>1</sup>H equal to 400 MHz.

The outcome of an experiment is always the sum over all the detectable spatially distributed sample volumes. Therefore the experimental data measured with an R-PDLF experiment is a sum over dipolar modulations, each modulation with a characteristic lineshape that depends on the local RF field. To simulate realistic R-PDLF experiments accounting for RF inhomogeneity, we first measured the RF inhomogeneity in the probe used in this work by the method suggested by Odedra and Wimperis (Odedra and Wimperis, 2013) (details given as SI). The simulation data was then generated by integrating over R-PDLF NMR simulations that covered the measured RF spatial inhomogeneity profile. This was done by composing a MATLAB script with SIMPSON call-outs enabling to build an R-PDLF simulation database with realistic modulations for a range of  $S_{\rm CH}$  values. The database generated included the effect of chemical shift offsets and RF miscalibration with or without a built in RF spatial inhomogeneity profile. The MATLAB/SIMPSON files and the NMR database generated are available as open data (https://github.com/tfmFerreira/inhomogeneous-rf-nmr.git).

The result of accounting for RF inhomogeneity in the R-PDLF simulations is shown in Figure 3 for a two-component dipolar modulation, consisting of two distinct C–H bond order parameters of 0.03 and 0.04. The main effect of including RF spatial inhomogeneity in the simulation is the damping of the signal and the non-zero long time average of the modulation which gives rise to a middle peak in the dipolar spectrum as also reported previously by Polenova and coworkers (Lu et al., 2016). Due to the limited number of points used, the distinct dipolar couplings are not visible in either dipolar spectrum of the two simulations. Figure 3D shows the result of adding random noise to the simulated data in Figure 3C and fitting the data either with ideal settings or by taking the RF spatial inhomogeneity profile into account. The fit performed with NMR simulations having ideal settings gave values of 0.034 and 0.094, thus precluding the use of such fitting procedure for accurate analysis under such conditions. In contrast, the fit performed based on NMR simulations accounting for RF spatial inhomogeneity matches the data almost perfectly - as expected since these simulations were used to generate the original data - giving order parameters values of 0.028 and 0.040, very close to the original values 0.03 and 0.04 used. This suggests that, if the (coil dependent) RF spatial inhomogeneity across the sample is measured with reasonable accuracy, it would be possible to simulate a set of realistic data for the complete range of dipolar couplings (0-22 kHz for <sup>1</sup>H-<sup>13</sup>C). This would better describe the experimental measurements and therefore enable a more accurate time-domain analysis.

### 140 2.1 Results and Discussion

120


130

135


145

#### 2.2 Time-domain fitting enables higher accuracy and shorter experiments

From the description shown in the preceding section it is clear that RF spatial inhomogeneity affects the dipolar modulation in R-PDLF experiments. Therefore, for accurate fits of experimental data, this effect needs to be taken into account. In this and the following section we demonstrate that accounting for the effect of RF inhomogeneity in time-domain fits of experimental data indeed leads to highly accurate fits of the experimental data and consequently to a considerable improvement of the accuracy of the C–H bond order parameters determined. Instead of in the time domain, the fitting could also be performed in the frequency domain, i.e. by performing the Fourier transform of the experimental data in the indirect dimension and by then fitting this data with the Fourier transform of the numerical simulations. Here, we simply use time domain fits and avoid performing




**Figure 3.** Simulation of the effect of RF spatial inhomogeneity on R-PDLF dipolar modulation and illustration of the advantage of time-domain over frequency-domain analysis of the data. (A) R-PDLF pulse sequence used in this work (Dvinskikh et al., 2004). (B) Ideal R-PDLF dipolar modulation with two components having C—H bond order parameters equal to 0.03 and 0.04 (grey lines display the individual components used). In the frequency domain, the two splittings can not be distinguished, due to the limited number of points used in the time domain. (C) The same as in (B) but including the effect of RF spatial inhomogeneity taken as a Gaussian distribution that reflects the RF inhomogeneity of the coil used in this work (details in the SI). (D) Simulation of an experimental result with low signal-to-noise ratio by addition of random noise to the simulated curve in (C).On the bottom plot, the curves show the result of fitting the data with two components using numerical simulations performed with ideal settings (red) and by accounting for the RF spatial inhomogeneity profile (blue).




**Figure 4.** The proposed method applied to a sample of POPE MLVs. (A) <sup>13</sup>C rINEPT spectrum of the POPE MLVs. (B) Time-domain (TD) fit accounting for RF inhomogeneity of the dipolar dimension of a R-PDLF experiment with a total of 8 points (thick solid line). The experimental dipolar spectrum is shown on the right together with the fit prediction of the experimental dipolar spectrum (using a total of 32 time-domain points). The thin lines show the time-domain curves that correspond to the error limits as defined in the SI. (C) The same as in (B) but fitting a total of 32 experimental points in the indirect dimension. A single-exponential decay of 14 Hz was multiplied to the simulated data. (D) Order parameter estimated by the time-domain fit (blue) and calculated from the dipolar splitting in the Fourier transform of the experimental data (gray) as function of the total number of points in the indirect dimension used.

the additional step of the Fourier transform which is unnecessary and has an extra computational time cost. Figure 4 shows the application of this time-domain analysis to a sample of 1-palmitoyl-2-oleoyl-sn-glycero-3-phosphoethanolamine (POPE) MLVs, illustrating the analysis method on the  $\beta$  carbon of the POPE headgroup. POPE is one of the most abundant lipids



**Figure 5.** The proposed methodology applied to a sample of a DMPC/DMPCd54 liquid crystalline system. (A) <sup>13</sup>C rINEPT spectrum and labels used to identify different carbons. (B) Time-domain fits neglecting the RF inhomogeneity across the sample. (C) Time-domain fits accounting for the RF inhomogeneity across the sample using one single dipolar coupling as fit parameter. (D) The same as in (C) but using two dipolar couplings as fit parameters assuming a two-component dipolar modulation.

in cellular membranes. It has recently been investigated with R-PDLF NMR in the context of the NMRlipids project. Such project aims to resolve the atomistic molecular structure of the most abundant lipids in nature. To this end, the project uses a combination of solid-state NMR experiments and all-atom molecular dynamics simulations (Bacle et al., 2021). The time-domain curve fitted to the experimental dipolar modulation measured for the  $\beta$  carbons gives an estimation of  $S_{\rm CH}$  =0.032  $\pm 0.006$  for this C-H bond order parameter magnitude using only the first 8 points along the indirect dimension. Such fit enables to predict the experimental dipolar spectrum that one would obtain using a higher number of indirect dimension points



**Figure 6.** <sup>2</sup>H NMR spectrum of the DMPC/DMPCd54 liquid crystalline system and symbols showing the positions used to determine order parameters.

as shown in Figures 4B and 4C. Figure 4D shows how the time-domain fit procedure enables to estimate an accurate order parameter using much shorter experiments in comparison to the use of the typical Fourier transform methodology. As it will be shown in a later section, this is particularly useful for complex systems for which only a limited number of indirect dimension points can be measured. Moreover, the fit procedure enables the estimation of the uncertainty of the value measured (the detailed description for defining the error bar is given as SI) which is more difficult to define with the Fourier transform method. Based on previous  $^2$ H NMR measurements, the  $\beta$  methylene carbon is expected to have the equivalent order parameters for the two C–H bonds (Gally et al., 1975). In the next section we demonstrate the usefulness of R-PDLF time-domain fits to analyse cases of two-component dipolar modulations and estimate the accuracy of the order parameters determined with the proposed method by comparison with  $^2$ H NMR quadrupolar splittings.

# 2.3 Comparison with <sup>2</sup>H NMR quadrupolar splittings

160

165

To further test the accuracy of our method, we used a water/DMPC/DMPCd54 liquid crystalline system ( $L_{\alpha}$  phase) such that it was possible to measure the  $^2$ H NMR spectrum of the perdeuterated acyl chains of the DMPCd54 molecules and compare it with the  $^1$ H- $^1$ 3C dipolar couplings determined from the acyl chains of the molecules with natural abundance of isotopes. Figure 5 shows the R-PDLF dipolar modulation for a number of resolved carbons from the liquid crystalline system and their corresponding fits with and without accounting for RF inhomogeneity. The improvement of the fits by accounting for RF inhomogeneity is evident, especially for the smaller couplings. We stress here that the RF inhomogeneity profile used

Table 1. Comparison of the order parameter magnitudes,  $|S_{\rm CH}|$ , determined from the DMPC/DMPCd54 liquid crystalline sample for the  $\omega$ ,  $\omega-1$  and  $\omega-2$  carbons from using the observed splitting in the frequency domain (FT) of the R-PDLF experiment, from time-domain (TD) fits with numerical simulations accounting for the effect of RF inhomogeneity on R-PDLF experiments using one and two components, and from quadrupolar splittings observed with a  $^2$ H quadrupolar echo.

Carbon	FT	TD fit with RF	<sup>2</sup> H NMR	
label		1 comp.	2 comp.	
$\omega$	0.0023	$0.0023 \pm 0.006$	$0.0023 \pm 0.004$	0.0023
	-	-	$0.0023 \pm 0.004$	-
$\omega$ -1	0.083	$0.090 \pm 0.009$	$0.096 \pm 0.009$	0.098
	-	-	$0.085 \pm 0.009$	0.083
ω-2	0.114	$0.118 \pm 0.01$	$0.122 \pm 0.011$	0.127
	-	-	$0.115 \pm 0.011$	0.115

in the simulations was experimentally measured and modelled by a Gaussian profile. The RF inhomogeneity is therefore an objectively quantified a priori constraint that we have obtained from an independent experiment. Moreover, since each of the three carbon peaks analysed is a sum of components from the two individual acyl chains, sn-1 and sn-2, we also performed the time-domain fits of the dipolar modulations using two components. In the case of the acyl chain methyl groups  $\omega$ , a single dipolar coupling minimizes the root mean square deviation (RMSD). On the other hand, for the  $\omega-1$  and  $\omega-2$  methylenes, two dipolar couplings enable a better fit than a single value which suggests a distinct motional geometry in the sn-1 and sn-2 acyl chains for these methylenes (a detailed analysis is given in the SI). The nonequivalence of the order parameters from the distinct chains is in agreement with previous investigations of specifically deuterated DPPC molecules (Seelig and Niederberger, 1974).

175

185

The  $^2$ H NMR spectrum measured from DMPCd54 molecules is shown in Figure 6. The five lowest quadrupolar splittings were used to calculate C–D bond order parameters. The results are included in Table 1 together with the results from the time-domain fits of the dipolar modulations from the molecules with natural abundance of isotopes. The agreement between the values obtained by using the different techniques is striking. The maximum difference between the order parameter magnitudes determined with the methodology proposed and the values determined from the  $^2$ H NMR spectrum is below  $\pm 0.005$ , much below the errors reported previously (up to  $\pm 0.02$ ) when doing similar comparisons between dipolar recoupling experiments with  $^2$ H NMR spectroscopy (Gross et al., 1997; Dvinskikh et al., 2005; Ferreira et al., 2013). Here, we used values taken from previous studies for the rigid  $^1$ H- $^{13}$ C dipolar and  $^2$ H quadrupolar couplings, namely 22 kHz and 126 kHz, respectively (Dvinskikh et al., 2004; Davis et al., 2009). We may therefore exclude a significant isotope effect on the molecular dynamics of the acyl chains in the water/DMPC/DMPCd54 liquid crystalline system. Moreover, we can unambiguously assign the C–D order parameters measured with  $^2$ H NMR based on the two-component fits of the R-PDLF data. The measurement of

the  $^2$ H NMR spectrum alone (from perdeuterated acyl chains) does not provide information concerning the origin of each of the distinct quadrupolar splittings observed. However, R-PDLF enables chemical shift selectivity in the direct dimension. By fitting the indirect dimension time-domain of the R-PDLF experiment for the  $\omega$ ,  $\omega$ -1 and  $\omega$ -2 carbons with two components (one per acyl chain), one readily enables the assignment of the five lowest quadrupole couplings in the  $^2$ H NMR spectrum.

The agreement between the  $^2$ H NMR and R-PDLF values obtained highlights the usefulness of the methodology described here over the conventional frequency-domain analysis: Such distinct dipolar couplings cannot be observed in the dipolar spectrum due to the limited number of points measured as described previously. Note that the only fitting parameters in the plots shown in Figure 5 are the dipolar couplings (i.e. order parameters), namely just one fitting parameter in case of a one component fit and two fitting parameters in case of a two-component fit. In the case of the two-component fits we have assumed that the intensity for each component was equal. This is a reasonable assumption in this case, because the expected differences between the order parameters and the dynamics of the two acyl chains are very small. Therefore, the rINEPT intensity of each component should be very similar according to the theoretical model for the dependence of  $^{13}$ C rINEPT intensity on CH order parameter and correlation times described by Topgaard and coworkers Nowacka et al. (2010, 2013). In general though, two-component fits may require fitting also the relative magnitude of each component for obtaining a good fit. Here we opted for a minimal amount of fitting parameters, i.e. the simplest model enabling to match the experimental data within experimental uncertainty. For determining the numbers shown in Table 1, we have used an iterative fit procedure as explained in the supplementary information were we also include a phenomenological single exponential relaxation. This did not affect the order parameters of the two-component fits determined for the carbons  $\omega$ -1 and  $\omega$ -2. For  $\omega$ , without inclusion of the single exponential relaxation, the order parameters calculated were 0.026 and 0.020.

In contrast to the POPE case described in the preceding section, for the DMPC/DMPCd54 system, including the exponential decay to account for relaxation effects affects the RMSD much less significantly (detailed comparison in the SI). We believe that this relates to the different membrane curvatures of the samples. While the POPE MLVs where prepared at excess hydration, the DMPC/DMPCd54 MLVs were prepared by placing the dry lipid film into a desiccator for about one day. The reduced hydration of the DMPC/DMPCd54 MLVs is expected to yield long range flat membranes eliminating relaxation effects from the reorientation motion due to the lateral diffusion of the lipid molecules. The POPE MLVs at excess hydration may have a higher curvature with lateral diffusion contributions to the spin relaxation during the dipolar recoupling dimension. However, for some other lipid systems, such as POPC and POPC/cholesterol that were prepared at excess hydration, a slower relaxation during the indirect dimension of the R-PDLF experiment was observed as well (results not shown). We refrained from investigating such relaxation effects in this work.

# 2.4 Application to a brain lipid extract

200

205


210

215

220

225

To showcase the applicability of the presented methodology we investigated the molecular structure of a complex membrane system composed of a brain lipid extract with several lipid types present. The complex lipid mixture consists of the chloroform:methanol extract of porcine brain tissue as purchased from AVANTI LIPIDS.



**Figure 7.** Phospholipid headgroup C–H bond order parameter magnitudes determined from a brain lipid extract using time-domain fits accounting for  $B_1$  inhomogeneity. The polarization transfer method used was rINEPT, recycle delay was 5 s, MAS rate was 5 kHz and a total of 4096 scans were acquired for each point in the indirect dimension which amounts roughly to 2 days of experimental time. (A) Chemical structures and carbon labels of the most abundant phospholipids in the brain lipid extract: phosphatidylethalonamine (PE), phosphatidylcholine (PC), and phosphatidylserine (PS).(B)  $^1$ H- $^{13}$ C rINEPT spectrum of the brain lipid extract membranes together with reference spectra for the most abundant lipids. The additional labels c and s were used to identify cholesterol and Galactose peaks, respectively. (C) Application of the proposed methodology for fitting the dipolar modulations and determining the corresponding order parameter magnitudes for selected carbons.

Figure 7 shows the <sup>13</sup>C spectrum of the brain lipid extract at excess hydration together with spectra acquired from single phospholipid model systems. Due to the much higher complexity of the brain lipid extract sample in comparison to model membrane systems (usually composed of 1-3 distinct components), the number of transients required to achieve a reasonable signal to noise ratio is considerably higher. Here, 4096 transients were acquired for each point in the indirect dimension with a total experimental time cost of approximately 48 hours (in comparison to the 256 transients used for the simple model sys-

tems). The comparison of the  $^{13}$ C spectra measured from the brain extract with the spectra from the simple models enables to identify a number of peaks from the different components in the extract, namely the  $\alpha$ ,  $\beta$  and  $\gamma$  segments of the headgroup of phosphatidylcholine (PC), the  $\alpha$  and  $\beta$  segments of phosphatidylethanolamine (PE), the  $\alpha$  and  $\beta$  segments of phosphatidylserine (PS), several cholesterol peaks and some galactose peaks most likely due to the presence of galactose cerebrosides in the mixture. Although only approximately 40% of the molar composition was known prior to the solid-state NMR experiments, from inspection of the  $^{13}$ C rINEPT spectrum, and knowing the lipid composition of myelin that makes up most of the dry weight of brain tissue (Norton and Autilio, 1966), it is clear that most of the remaining unknown molar content is cholesterol plus a considerable fraction of galactose cerebrosides (GalCer).

Figure 7C shows the indirect time-domain modulation for a number of identified peaks in the spectra measured by R-PDLF spectroscopy. In this case, due to the large number of transients required to achieve a a reasonable signal to noise ratio (4096 scans), only very short indirect dimension modulations can be measured. Namely, only 8 points in the indirect dimension were acquired for the brain lipid extract sample. Such a reduced total indirect dimension time is not sufficient to obtain dipolar splittings after Fourier transformation. Within such an experimental setup the Fourier transform of these time-domain modulations does not enable resolution for determining dipolar splittings. However, application of time-domain fitting including RF inhomogeneity enables to determine the dipolar couplings and therefore estimate C–H bond order parameters. The estimated order parameters are presented in Table 2.

The headgroup order parameters determined from the set of different phospholipid types in the brain lipid extract are very much in line with what is observed in corresponding lipid model membranes. The only exception is the PS  $\beta$  order parameter with a rather small order parameter equal to 0.04. This suggests that the brain lipid extract may contain calcium ions that are known to significantly affect the PS headgroup orientation. A thorough analysis of the order parameters in terms of the molecular structure in the complex brain lipid extract system will not be given here. Such analysis should be done in combination with experiments and MD simulations on PC/PE/PS/cholesterol/GalCer model membranes which resemble the lipid extract composition, to investigate potential phase separation behavior, effects of distinct ions present in the system and pH, or ultimately the effect of macromolecules such as basic myelin protein, the most abundant protein in myelin. We believe that such a combined approach will lead to a more accurate description of the lipid/protein interactions in myelin, which may be important in the context of a number of diseases related to demyelination. The methodology proposed here will enable to perform such investigations with higher accuracy using shorter experiments and without the requirement of isotopic labelling. Moreover, we envision applications to many other complex membranes such as bacterial membrane models, eukaryotic membrane models with cell type or organelle specific compositions, technological systems such as lipid-nanoparticles, as well as to address the molecular structure/orientation of drugs, peptides or other molecules with anisotropic motion in lipid membranes.

# 3 Conclusions

235

240

245

250

255

260

In summary, we have described a PDLF NMR methodology that consists of performing time-domain fits of the dipolar modulation with numerical simulations that account for the RF spatial inhomogeneity of the probe used. The RF inhomogeneity

**Table 2.** Order parameter magnitudes determined from a brain lipid extract using the methodology proposed in this work compared to  ${}^{2}H$  NMR values measured from phospholipid model systems reported previously.  ${}^{a)}$   $T=40^{\circ}C$ ,  ${}^{b)}$   $T=30^{\circ}C$  and  ${}^{c)}$   $T=23^{\circ}C$ 

Carbon	Brain extract	Model system	
segment	$ S_{ m CH} $	$ S_{\mathrm{CD}} $	model
$\alpha$ (PC)	$0.044\pm0.015$	$0.048^{a}$	DPPC/chol (Brown and Seelig, 1978)
		$0.049^{b}$	DOPC (Ulrich and Watts, 1994)
		$0.048^{c}$	POPC (Bechinger and Seelig, 1991)
0 ( <b>P.C</b> )	0.01   0.015	0.0249	DDDG/ 1 1 /D 19 19 19 19 19 19 19 19 19 19 19 19 19
$\beta$ (PC)	$0.01 \pm 0.015$	$0.024^{a}$	DPPC/chol (Brown and Seelig, 1978)
		$0.029^{b}$	DOPC (Ulrich and Watts, 1994)
		$0.044^{c}$	POPC (Bechinger and Seelig, 1991)
$\gamma$ (PC)	$0.001 \pm 0.016$	$0.008^{a}$	DPPC/chol (Brown and Seelig, 1978)
α (PE)	$0.066 \pm 0.024$	0.063 <sup>a</sup>	DPPE/chol (Brown and Seelig, 1978)
β (PE)	$0.019 \pm 0.012$	$0.020^{a}$	DPPE/chol (Brown and Seelig, 1978)
α (PS)	$0.026 \pm 0.018$	$0.016^{a}$	POPC/POPS (Roux and Bloom, 1990)
β (PS)	$0.040 \pm 0.039$	$0.091^{a}$	POPC/POPS (Roux and Bloom, 1990)
		$0.062^{a}$	+ 1 M CaCl <sub>2</sub> (Roux and Bloom, 1990)

is experimentally measured and is an a priori constraint. A dipolar modulation of a single component is therefore fitted with only two parameters, the dipolar coupling and a phenomenological relaxation time to account for exponential relaxation. The proposed methodology enables one to determine C–H bond order parameters from simple model systems with a much higher accuracy than previously and enables to investigate complex lipid membrane systems that were so far inaccessible using the conventional PDLF NMR methodology. We believe that the method presented will be extremely useful in the future concerning molecular structural investigations of complex systems such as multi-component models, lipid extracts and lipid membranes

with drugs, peptides or other molecules incorporated. Moreover, the higher accuracy and farther reach of the method will also be fundamental to validate molecular dynamics simulations for which structural experimental data was not accessible so far.

Code and data availability. https://github.com/tfmFerreira/inhomogeneous-rf-nmr.git

Author contributions. T.M.F. idealized the project. T.M.F. and A.W. developed the code. T.M.F and A.W. performed experiments and processed the experimental data. T.M.F., A.W. and K.S. analysed and discussed the results and manuscript. T.M.F. wrote the manuscript.

Competing interests. no competing interests

Acknowledgements. T.M.F. greatly acknowledges Alexey Krushelnitsky for invaluable support and discussions. This research study was funded by the German Research Foundation (Deutsche Forschungsgemeinschaft,DFG) [project number 189853844, TRR 102 (T.M.F. and A.W.)].

295

- Aisenbrey, C., Salnikov, E., and Bechinger, B.: Solid-State NMR Investigations of the MHC II Transmembrane Domains: Topological Equilibria and Lipid Interactions, J. Mem. Biol., 252, 371–384, https://doi.org/10.1007/s00232-019-00071-8, 2019.
- Andersson, J. M., Grey, C., Larsson, M., Ferreira, T. M., and Sparr, E.: Effect of cholesterol on the molecular structure and transitions in a clinical-grade lung surfactant extract, P. Natl. Acad. Sci. USA, 114, E3592–E3601, https://doi.org/10.1073/pnas.1701239114, 2017.
- Asami, S. and Reif, B.: Comparative Study of REDOR and CPPI Derived Order Parameters by 1H-Detected MAS NMR and MD Simulations, J. Phys. Chem. B, 121, 8719–8730, https://doi.org/10.1021/acs.jpcb.7b06812, 2017.
  - Bacle, A., Buslaev, P., Garcia-Fandino, R., Favela-Rosales, F., Ferreira, T., Fuchs, P., Gushchin, I., Javanainen, M., Kiirikki, A., Madsen, J., Melcr, J., Milán Rodríguez, P., Miettinen, M., Ollila, O., Papadopoulos, C., Peón, A., Piggot, T., Piñeiro, A., and Virtanen, S.: Inverse Conformational Selection in Lipid–Protein Binding, J. Am. Chem. Soc., 143, 13701–13709, https://doi.org/10.1021/jacs.1c05549, 2021.
- 290 Bechinger, B. and Seelig, J.: Conformational changes of the phosphatidylcholine headgroup due to membrane dehydration. A 2H-NMR study, Chemistry and Physics of Lipids, 58, 1–5, https://doi.org/https://doi.org/10.1016/0009-3084(91)90105-K, 1991.
  - Brown, M. and Seelig, J.: Influence of cholesterol on the polar region of phosphatidylcholine and phosphatidylethanolamine bilayers, Biochemistry, 17, 381–384, https://doi.org/10.1021/bi00595a029, 1978.
  - Bärenwald, R., Achilles, A., Lange, F., Ferreira, T. M., and Saalwächter, K.: Applications of Solid-State NMR Spectroscopy for the Study of Lipid Membranes with Polyphilic Guest (Macro)Molecules, Polymers, 8, https://doi.org/10.3390/polym8120439, 2016.
  - Cady, S., Goodman, C., Tatko, C., DeGrado, W., and Hong, M.: Determining the Orientation of Uniaxially Rotating Membrane Proteins Using Unoriented Samples: A 2H, 13C, and 15N Solid-State NMR Investigation of the Dynamics and Orientation of a Transmembrane Helical Bundle, J. Am. Chem. Soc., 129, 5719–5729, https://doi.org/10.1021/ja070305e, 2007.
  - Cheng, Y.: Membrane protein structural biology in the era of single particle cryo-EM, Curr. Opin. Struct. Biol., 52, 58–63, https://doi.org/https://doi.org/10.1016/j.sbi.2018.08.008, 2018.
  - Chevelkov, V., Fink, U., and Reif, B.: Accurate Determination of Order Parameters from 1H,15N Dipolar Couplings in MAS Solid-State NMR Experiments, J. Am. Chem. Soc., 131, 14018–14022, https://doi.org/10.1021/ja902649u, 2009.
  - Davis, J.: The description of membrane lipid conformation, order and dynamics by 2H-NMR, Biochimica et Biophysica Acta (BBA) Reviews on Biomembranes, 737, 117–171, https://doi.org/https://doi.org/10.1016/0304-4157(83)90015-1, 1983.
- Davis, J. H., Clair, J. J., and Juhasz, J.: Phase equilibria in DOPC/DPPC-d62/cholesterol mixtures, Biophys. J., 96, 521–539, https://doi.org/10.1016/j.bpj.2008.09.042, 2009.
  - De Paëpe, G.: Dipolar Recoupling in Magic Angle Spinning Solid-State Nuclear Magnetic Resonance, Annu. Rev. Phys. Chem., 63, 661–684, https://doi.org/10.1146/annurev-physchem-032511-143726, 2012.
- deAzevedo, E., Saalwachter, K., Pascui, O., de Souza, A., Bonagamba, T., and Reichert, D.: Intermediate motions as studied by solid-state separated local field NMR experiments, J. Chem. Phys., 128, 104 505, https://doi.org/10.1063/1.2831798, 2008.
  - Dvinskikh, S., Castro, V., and Sandström, D.: Efficient solid-state NMR methods for measuring heteronuclear dipolar couplings in unoriented lipid membrane systems, Phys. Chem. Phys., 7, 607—613, https://doi.org/10.1039/b418131j, 2005.
  - Dvinskikh, S. V., Zimmermann, H., Maliniak, A., and Sandstrom, D.: Measurements of motionally averaged heteronuclear dipolar couplings in MAS NMR using R-type recoupling, J. Magn. Reson., 168, 194–201, 2004.

Ferreira, T. M., Coreta-Gomes, F., Ollila, O. H. S., Moreno, M. J., Vaz, W. L. C., and Topgaard, D.: Cholesterol and POPC segmental order parameters in lipid membranes: solid state 1HĐ13C NMR and MD simulation studies, Phys. Chem. Chem. Phys., 15, 1976–1989, https://doi.org/10.1039/C2CP42738A, 2013.

320

- Fridolf, S., Hamid, M., Svenningsson, L., Skepö, M., Sparr, E., and Topgaard, D.: Molecular dynamics simulations and solid-state nuclear magnetic resonance spectroscopy measurements of C–H bond order parameters and effective correlation times in a POPC-GM3 bilayer, Phys. Chem. Chem. Phys., 24, 25588–25601, https://doi.org/10.1039/D2CP02860C, 2022.
- Gally, H. U., Niederberger, W., and Seelig, J.: Conformation and motion of the choline head group in bilayers of dipalmitoyl-3-sn-phosphatidylcholine, Biochemistry, 14, 3647–3652, https://doi.org/10.1021/bi00687a021, 1975.
- Gansmüller, A., Simorre, J.-P., and Hediger, S.: Windowed R-PDLF recoupling: A flexible and reliable tool to characterize molecular dynamics, J. Mag. Reson., 234, 154–164, https://doi.org/l0.1016/j.jmr.2013.06.017, 2013.
- Gauto, D., Estrozi, L., Schwieters, C., Effantin, G., Macek, P., Sounier, R., Sivertsen, A., Schmidt, E., Kerfah, R., Mas, G., Colletier, J., Güntert, P., Favier, A., Schoehn, G., Schanda, P., and Boisbouvier, J.: Integrated NMR and cryo-EM atomic-resolution structure determination of a half-megadalton enzyme complex, Nat. Commun., 10, 2697, https://doi.org/10.1038/s41467-019-10490-9, 2019.
  - Griffin, R.: Dipolar recoupling in MAS spectra of biological solids, Nat. Struct. Biol., 5, 508-512, https://doi.org/10.1038/749, 1998.
  - Gross, J. D., Warschawski, D. E., and Griffin, R. G.: Dipolar Recoupling in MAS NMR: A Probe for Segmental Order in Lipid Bilayers, J. Am. Chem. Soc., 119, 796–802, 1997.
  - Gullion, T.: Introduction to rotational-echo, double-resonance NMR, Concepts in Magnetic Resonance, 10, 277–289, https://doi.org/https://doi.org/10.1002/(SICI)1099-0534(1998)10:5<277::AID-CMR1>3.0.CO;2-U.
  - Hester, R., Ackerman, J., Neff, B., and Waugh, J.: Separated Local Field Spectra in NMR: Determination of Structure of Solids, Phys. Rev. Lett., 36, 1081–1083, https://doi.org/10.1103/PhysRevLett.36.1081, 1976.
- Hou, G., Byeon, I., Ahn, J., Gronenborn, A., and Polenova, T.: 1H–13C/1H–15N Heteronuclear Dipolar Recoupling by R-Symmetry Sequences Under Fast Magic Angle Spinning for Dynamics Analysis of Biological and Organic Solids, J. Am. Chem Soc., 133, 18646–18655, https://doi.org/10.1021/ja203771a, 2011.
  - Jain, M., Mote, K., Hellwagner, J., Rajalakshmi, G., Ernst, M., Madhu, P., and Agarwal, V.: Measuring strong one-bond dipolar couplings using REDOR in magic-angle spinning solid-state NMR, J. Chem. Phys., 150, 134 201, https://doi.org/10.1063/1.5088100, 2019.
- Leftin, A. and Brown, M. F.: An NMR database for simulations of membrane dynamics, Biochim. Biophys. Acta-Biomembr., 1808, 818 839, https://doi.org/10.1016/j.bbamem.2010.11.027, 2011.
  - Leftin, A., Molugu, T., Job, C., Beyer, K., and Brown, M.: Area per Lipid and Cholesterol Interactions in Membranes from Separated Local-Field 13C NMR Spectroscopy, Biophys. J., 107, 2274–2286, https://doi.org/https://doi.org/10.1016/j.bpj.2014.07.044, 2014.
- Levitt, M.: Symmetry-Based Pulse Sequences in Magic-Angle Spinning Solid-State NMR, https://doi.org/https://doi.org/10.1002/9780470034590.emrstmt 2007.
  - Lindon, J. and Ferrige, A.: Digitisation and data processing in Fourier transform NMR, Prog. Nucl. Magn. Reson. Spectrosc., 14, 27–66, https://doi.org/https://doi.org/10.1016/0079-6565(80)80002-1, 1980.
  - Löser, L., Saalwächter, K., and Ferreira, T.: Liquid-liquid phase coexistence in lipid membranes observed by natural abundance 1H-13C solid-state NMR, Phys. Chem. Chem. Phys., 20, 9751–9754, https://doi.org/10.1039/C8CP01012A, 2018.
- 350 Lu, X., Zhang, H., Lu, M., Vega, A. J., Hou, G., and Polenova, T.: Improving dipolar recoupling for site-specific structural and dynamics studies in biosolids NMR: windowed RN-symmetry sequences, Phys. Chem. Chem. Phys., 18, 4035–4044, https://doi.org/10.1039/C5CP07818K, 2016.

- Mandala, V., Williams, J., and Hong, M.: Structure and Dynamics of Membrane Proteins from Solid-State NMR, Annu. Rev. Biophys., 47, 201–222, https://doi.org/10.1146/annurev-biophys-070816-033712, 2018.
- Mandala, V., Loftis, A., Shcherbakov, A., Pentelute, B., and Hong, M.: Atomic structures of closed and open influenza B M2 proton channel reveal the conduction mechanism, Nat. Struct. Mol. Biol., 27, 160–167, https://doi.org/10.1038/s41594-019-0371-2, 2020.
  - Molugu, T., Lee, S., and Brown, M.: Concepts and Methods of Solid-State NMR Spectroscopy Applied to Biomembranes, Chem. Rev., 117, 12 087–12 132, https://doi.org/10.1021/acs.chemrev.6b00619, 2017.
  - Nakai, T. and Terao, T.: Measurements of heteronuclear dipolar powder patterns due only to directly bonded couplings, Magn. Reson. Chem., 30, 42–44, https://doi.org/https://doi.org/10.1002/mrc.1260300109, 1992.

360

375

- Nimerovsky, E. and Soutar, C.: A modification of γ-encoded RN symmetry pulses for increasing the scaling factor and more accurate measurements of the strong heteronuclear dipolar couplings, J. Magn. Reson., 319, 106827, https://doi.org/https://doi.org/10.1016/j.jmr.2020.106827, 2020.
- Nishimura, K., Fu, R., and Cross, T. A.: The Effect of RF Inhomogeneity on Heteronuclear Dipolar Recoupling in Solid State NMR: Practical Performance of SFAM and REDOR, J. Mag. Reson., 152, 227 233, https://doi.org/https://doi.org/10.1006/jmre.2001.2410, 2001.
  - Norton, W. and Autilio, L.: The lipid composition of purified bovine brain myelin, J. Neurochem., 13, 213–222, https://doi.org/https://doi.org/10.1111/j.1471-4159.1966.tb06794.x, 1966.
  - Nowacka, A., Mohr, P., Norrman, J., Martin, R., and Topgaard, D.: Polarization Transfer Solid-State NMR for Studying Surfactant Phase Behavior, Langmuir, 26, 16848–16856, https://doi.org/10.1021/la102935t, 2010.
- Nowacka, A., Bongartz, N., Ollila, O., Nylander, T., and Topgaard, D.: Signal intensities in 1H–13C CP and INEPT MAS NMR of liquid crystals, J. Magn. Reson., 230, 165–175, https://doi.org/https://doi.org/10.1016/j.jmr.2013.02.016, 2013.
  - Odedra, S. and Wimperis, S.: Imaging of the B1 distribution and background signal in a MAS NMR probehead using inhomogeneous B0 and B1 fields, J. Mag. Reson., 231, 95 99, https://doi.org/https://doi.org/10.1016/j.jmr.2013.04.002, 2013.
  - Park, S., Das, B., Casagrande, F., Tian, Y., Nothnagel, H., Chu, M., Kiefer, H., Maier, K., De Angelis, A., Marassi, F., and Opella, S.: Structure of the chemokine receptor CXCR1 in phospholipid bilayers, Nature, 491, 779–784, https://doi.org/10.1038/nature11580, 2012.
  - Roux, M. and Bloom, M.: Calcium, magnesium, lithium, sodium, and potassium distributions in the headgroup region of binary membranes of phosphatidylcholine and phosphatidylserine as seen by deuterium NMR, Biochemistry, 29, 7077–7089, https://doi.org/10.1021/bi00482a019, 1990.
  - Schanda, P., Meier, B., and Ernst, M.: Accurate measurement of one-bond H–X heteronuclear dipolar couplings in MAS solid-state NMR, J. Magn. Reson., 210, 246–259, https://doi.org/https://doi.org/10.1016/j.jmr.2011.03.015, 2011.
  - Schmidt-Rohr, K., Nanz, D., Emsley, L., and Pines, A.: NMR Measurement of Resolved Heteronuclear Dipole Couplings in Liquid Crystals and Lipids, J. Phys. Chem., 98, 6668–6670, https://doi.org/10.1021/j100078a002, 1994.
  - Seelig, J.: Deuterium magnetic resonance: theory and application to lipid membranes, Quarterly Reviews of Biophysics, 10, 353–418, https://doi.org/10.1017/S0033583500002948, 1977.
- Seelig, J. and Niederberger, W.: Deuterium-labeled lipids as structural probes in liquid crystalline bilayers. Deuterium magnetic resonance study, J. Am. Chem. Soc., 96, 2069–2072, https://doi.org/10.1021/ja00814a014, 1974.
  - Strandberg, E. and Ulrich, A.: NMR methods for studying membrane-active antimicrobial peptides, Concepts in Magnetic Resonance Part A, 23A, 89–120, https://doi.org/https://doi.org/10.1002/cmr.a.20024, 2004.
- Tošner, Z., Purea, A., Struppe, J. O., Wegner, S., Engelke, F., Glaser, S. J., and Reif, B.: Radiofrequency fields in MAS solid state NMR probes, Journal of Magnetic Resonance, 284, 20–32, https://doi.org/https://doi.org/10.1016/j.jmr.2017.09.002, 2017.

- Ulrich, A. S. and Watts, A.: Molecular response of the lipid headgroup to bilayer hydration monitored by <sup>2</sup>H NMR, Biophys. J., 66, 1441–1449, https://doi.org/10.1016/S0006-3495(94)80934-8, 1994.
- Umegawa, Y., Matsumori, N., and Murata, M.: Chapter Two Recent Solid-State NMR Studies of Hydrated Lipid Membranes, vol. 94 of *Annual Reports on NMR Spectroscopy*, pp. 41–72, Academic Press, https://doi.org/https://doi.org/10.1016/bs.arnmr.2017.12.003, 2018.
- Wu, M. and Lander, G.: How low can we go? Structure determination of small biological complexes using single-particle cryo-EM, Curr. Opin. Struct. Biol., 64, 9–16, https://doi.org/10.1016/j.sbi.2020.05.007, 2020.
  - Xue, K., Nimerovsky, E., Tekwani Movellan, K., Becker, S., and Andreas, L.: Backbone Torsion Angle Determination Using Proton Detected Magic-Angle Spinning Nuclear Magnetic Resonance, J. Phys. Chem. Lett., 13, 18–24, https://doi.org/10.1021/acs.jpclett.1c03267, 2022.
- Zerweck, J., Strandberg, E., Kukharenko, O., Reichert, J., Buerck, J., Wadhwani, P., and Ulrich, A.: Molecular mechanism of synergy between the antimicrobial peptides PGLa and magainin 2, Sci. Rep., 7, https://doi.org/10.1038/s41598-017-12599-7, 2017.