

Light-coupled cryogenic probes to detect low micromolar samples and allow automated NMR platform

Wolf Wüster¹, Pit Gebbers¹, Alois Renn², Matthias Bütkofer², Sophie Rüdiger,² Roland P. Riek², and Felix Torres^{2,3}

5 ¹Zurich University of Applied Sciences, School of Engineering, Forschungsschwerpunkt Angewandte Optik, Technikumstrasse 9, 8400 Winterthur, Switzerland

²Federal Institute of Technology Zurich (ETHZ), Institute of Molecular Physical Science, Vladimir-Prelog-Weg 2, 8093, Zürich, Switzerland

³NexMR AG, wiesenstrasse 10A, 8952, Schlieren, Switzerland

10 *Correspondence to:* Roland Riek (roland.riek@phys.chem.ethz.ch) and Felix Torres (ftorres@nexmr.com)

Abstract. Recent advances in NMR fragment screening use sample illumination to boost NMR sensitivity, reduce measurement time to a few seconds, and reduce sample concentration to a few micromolars. Nevertheless, the absence of a fully automated solution to measure several hundreds of samples with photoinduced hyperpolarization limits the large-scale applicability of the method. We present a setup to couple an optical fiber with a cryogenic probe using the flow-cell accessory port. This setup is compatible with commercially available autosamplers, enabling the fully automated measurement of several hundreds of samples per day.

1 Introduction

Photochemically induced dynamic nuclear polarization (photo-CIDNP) enhances NMR sensitivity thanks to the radical pair recombination after nuclear spin-dependent singlet-triplet mixing in a magnetic field.(Ward and Lawler, 1967; Bargon and Fischer, 1967; Kaptein and Oosterhoff, 1969; Closs, 1969) Typically, a photosensitizer, e.g., fluorescein or Atto Thio 12, is excited with light to generate the radical pair with a small molecule of interest, e.g., tryptophan tyrosine, histidine, or diverse heteroaromatic scaffolds.(Okuno and Cavagnero, 2016; Morozova and Ivanov, 2019; Hore and Broadhurst, 1993; Torres, 2021; Torres et al., 2023; Torres et al., 2021) After recombination, the nuclear spin population of the small molecules is out of Boltzmann equilibrium, yielding signal-to-noise enhancements (SNE) in the range of 20 to 100-fold, depending on the photosensitizer-molecule pair, its magnetic parameters, g-factors, the hyperfine couplings, and the magnetic field.(Torres, 2021; Torres et al., 2021; Sobol et al., 2019) Photo-CIDNP is typically performed in an aqueous buffer at room temperature and requires a few seconds of light irradiation. Therefore, it is an ideal hyperpolarization method for biological application, including fragment screening, where thousands of fragment molecules are screened against a biological target to identify an interaction. Photo-CIDNP screening of fragment libraries was reported with an experimental time of a few seconds and low micromolar concentration at high and low magnetic fields.(Torres et al., 2023; Stadler et al., 2023) Other applications potentially requiring high throughput were described recently, such as metabolomic.(Kuhn et al., 2024) However, the sample

illumination requires the insertion of an optical fiber into the NMR tube, impeding the use of automatic sampling machines, which would break the fiber. This drawback is major as the sample needed to be inserted by hand, reducing the throughput and monopolizing the staff's time over a repetitive task. While a flowthrough NMR where an HPLC is in fluidic connection
35 with a flowcell integrating an optical fiber was designed to solve this issue,(Torres et al., 2023) this solution presents the disadvantage of regular clogging due to the accumulation of small molecules or proteins in micrometer-sized tubing. NMRtorch is an alternative option to illuminate the sample with NMR tubes without an optic fiber, particularly in the case of high optical density due to the dye present in diverse light-coupled NMR applications comprising photo-CIDNP.(Bramham,
2022) Nevertheless, integrating NMRtorch to commercially available autosamplers is not straightforward and would require
40 hardware adaptation.

Furthermore, the samples measured for fragment screening contain a relatively low concentration of photosensitizer (2-10 μM),(Torres et al., 2023; Stadler et al., 2023) and the optical density is not critical to obtain sufficiently high SNE. Previous designs from others inspired this work to irradiate the NMR tubes from the bottom. These previous works integrate light with quartz rode or optic mirrors into room temperature probes, and they use flat-bottom NMR tubes to improve the coupling of the
45 light into the sample.(Kuhn, 2013; Kuprov and Hore, 2004) More recent work uses optic fibers inserted into drilled room temperature probes and modified NMR tubes.(Tolstoy et al., 2009; Koeppe et al., 2011) The objective of the present work is to demonstrate that it is possible to integrate an optical fiber without any hardware modification, which is more modular and can be inserted into a cryogenic probe using the flow-cell accessory port. This presents the advantage of benefit from the state-of-the-art performances of cryogenic probes, and it is easily installed using off-the-shelf and affordable laser components.
50 Finally, we show that it is possible to take advantage of the spherical shape of the bottoms of standard NMR tubes and the resulting lensing effect to achieve sufficient light-sample coupling. The platform presented in this manuscript is compatible with fully automated solutions such as commercially available autosamplers. It uses standard NMR consumables, guaranteeing a straightforward implementation for the photo-CIDNP NMR field.

2 Results and discussion

55 We use the flow-cell accessory port of a cryogenic probe to insert an optical multi-mode fiber in a non-invasive way (Figure S1). The full functionality of the cryogenic probe is always granted while the fiber is inserted. We take advantage of the fact that cryogenic probes are designed to be operated with flow cell tubes inserted at the bottom of the cryogenic probe. We find that an optical multimode fiber with a maximum outer diameter of roughly 1 mm can be used instead of a flow tube. The fiber can be inserted into the cryogenic probe and brought into the NMR tube's proximity without constriction. The illumination
60 scheme is depicted in Figure 1A. The multimode fiber is inserted until it reaches the lower end of the NMR tube. The light emerges from the optical fiber as a light cone whose divergence angle depends on the fiber's numerical aperture (NA). The light cone illuminates the bottom of the NMR sample and is refracted into the sample due to the refractive index difference of

air ($n=1$) and the borosilicate glass of the NMR sample tube ($n=1.47$). In the following, we describe the results for 3 mm tubes as they are

65 Due to the spherical shape of the bottom of the NMR tube, the light is transmitted into the sample liquid (Figure 1B). The shape of the tube bottom can be described as a spherical convex lens with two different curvatures. For a standard 3 mm NMR tube, we find an outer curvature of 1.3 mm, an inner curvature of 1.6 mm, and a tube wall thickness of 0.35 mm (Figure S2). We simulated the light path via optical ray tracing (Figure 1A and 1B). For this, the software package Ansys Zemax OpticStudio was used in non-sequential mode. The light is confined to the liquid sample and the tube walls via total internal
70 reflection resulting from the difference in the refractive index of air ($n=1.00$) and the sample, which is essentially water ($n=1.33$). Light rays that travel in the liquid at a maximum angle of 41.3° with respect to the sample wall will be confined within the sample and the tube walls. The confinement of the light within the sample was verified experimentally by bringing a multimode fiber close to an NMR sample tube filled with fluorescein (Figure 1C).

75 As the light emerges as a cone from the optical fiber output, the effect of the distance between the optic fiber output and the bottom of the tube was evaluated. As a result of the lensing effect of the bottom of the NMR tube, there are different scenarios depending on the position of the optic fiber output relative to the focal point of the NMR tube's bottom (approximately 2 mm). With a distance of 0-2 mm, the light beam is slightly divergent within the sample liquid, and with a distance superior to 2 mm, the light beam is focused within the NMR tube sample (Figure S3). In agreement with our postulate that the NMR tube bottom
80 is a lens, the light path is collimated when the distance between the optic fiber output and the NMR tube's bottom matches the focal point, i.e., approximately 2 mm (Figure S3). This agrees with the back focal length of 2.1 mm computed using the thick lens formula (Greivenkamp, 2004) and the values shown in Figure S2. The amount of light dispensed into the sample is stable for a distance range of 0.0 to 4.0 mm and starts decreasing for distances greater than this (Figure 1D). While the diameter of the optic fiber seems not to affect the coupling, we observe that the NA plays a critical role, and only $NA = 0.22$ allows a
85 robust setup (Figure 1D).

The light path inside the sample volume depends on the exact positioning of the optic fiber output to the sample. By mechanical construction of the Cryogenic probe flow accessory port, the fiber is coaxial to the sample tube. Nevertheless, since the fiber is flexible and does not fill the port fully, the alignment is expected to be imperfect, translating into a transverse offset. The sensitivity of the optical light path to transverse offset was evaluated for an optic fiber of $500 \mu\text{M}$ and $NA = 0.22$ fiber and a
90 3 mm NMR sample tube, using ray-tracing simulations (Figure S4). The optimal position is found where the fiber is coaxially centered with a distance to the tube on the order of 2 mm. In this case, the light beam is more or less collimated within the sample liquid. When a transverse offset of 0.5 mm is applied, the absorbed fraction reduces by 3% and 30% for a transverse offset of 1 mm (Figure 1E). Moreover, the light is confined to the NMR tube by the higher refractive index of the tube ($n=1.47$) compared to the air ($n=1.00$). Considering that the typical multimode optic fiber diameter is 0.4-0.9 mm and the accessory port
95 of the cryogenic probe is approximately 1.2 mm, an offset above 0.8 mm is unlikely. Therefore, even if the fiber position is offset, the light path is altered, but most of the light remains within the sample volume due to total internal reflection (Figure

S4). In summary, these results support our experimental findings that inserting the optical fiber near the NMR tube yields illumination of the sample in the NMR tube.




Figure 1: A) Sketch of the optical excitation scheme. The excitation laser is coupled to a multimode fiber. The fiber is inserted at the bottom of the cryogenic probe and is brought into the vicinity (distance of a few mm) of the NMR sample tube. B) Ray-tracing simulation of the light path. The method relies on the lensing effect of the standard NMR tube bottom. The light is confined within the NMR sample volume and NMR tube glass via total internal reflection. C) Optical bench top experiment with a multimode fiber in contact with a 3 mm diameter NMR tube. The light path is visible via the fluorescence of the fluorescein sample. D) Simulation of the absorbed power (AU) in the NMR sample volume as a function of longitudinal axial distance (Longitudinal z shift) for different NA values and optic fiber diameters (500 and 900 μm). E) Simulation of the absorbed power (AU) in the NMR sample volume as a function of lateral axial distance (Lateral y shift) for different NA values and optic fiber diameters (500 and 900 μm). The absorption fraction is the light flux absorbed within the NMR active sample volume divided by the light flux incident on the tube bottom. The absorption is computed by the Raytracing software using the Beer-Lambert law, assuming a fluorescein concentration of 10 $\mu\text{M/l}$ (0.10752 cm^{-1}).

105

110

115

To verify the absence of field homogeneity perturbation caused by inserting an optical fiber close to the B_1 coils, we recorded the spectra of 0.3% CHCl_3 in acetone- d_6 , the standard line shape reference sample (Figure 2A). The line shape did not exhibit any particular difference between the two setups; as shown in Table 1, the line widths at different levels from the maximum peak intensity are similar. The linewidth was also measured using the *peakw* command, resulting in two identical values of 1.147 Hz at 66% of the peak maximum.

Table 1: Line widths of the CHCl_3 peak for different setups, including an optical fiber in the cryogenic probe or not.

<i>setup</i>	<i>Linewidth at 0.11% maximum (Hz)</i>	<i>Linewidth at 0.55% maximum (Hz)</i>	<i>Linewidth at 50% maximum (Hz)</i>

<i>No fiber</i>	4.0	3.6	1.23
<i>Fiber</i>	4.3	3.7	1.23

Finally, the field map was recorded after shimming (*topshim map* command in Topspin®), as shown in Figure 2B. The extrema regions, i.e., between 1.0 and 1.5 cm, resp -1.0 and -1.5 cm, show differences in the measured field, which is unrelated to the optical fiber insertion as they appear at both ends. These differences are minor and result from how the algorithm fits the best polynomial in the area where the sample is measured (-0.5 to 0.5 cm). The similarity in shimming performances is reflected in the topshim reports with a B_0 standard deviation of 0.24 Hz and 0.22 Hz (*topshim report* command in Topspin®) for the setups with and without fiber, respectively. This positive outcome was expected as prior work from other groups did not report field homogeneity or shimmability issues;(Tolstoy et al., 2009; Koeppe et al., 2011) nevertheless, prior designs were for room temperature probes and not cryogenic probes. Therefore, it was essential to validate the compatibility of this setup for a different probe geometry.

To exemplify the performances of the fiber-coupled cryogenic probe, we recorded the most emblematic experiment of bioNMR, namely a [^{15}N , ^1H]-hetero single quantum coherence (HSQC)(Bodenhausen and Ruben, 1980) which is often used in fragment screening(Kerber et al., 2023) or for affinity determination.(Williamson, 2013) The [^{15}N , ^1H]-HSQC of the KRAS G13D mutant was recorded in the absence and presence of optical fiber (Figure 2C), and we examined the peak shape by extracting the row and column slices (*slice* command in Topspin®). The peaks showed no significant difference in the presence or absence of the optical fiber inserted inside the probe (Figure 2C). The two spectra were so similar that, in the spectra overlay (Figure 2C), it is impossible to observe the spectrum of KRAS that was measured with the fiber (blue). Therefore, we conclude there is no need to remove the optical fiber once installed, as it does not interfere with the non-illuminated NMR experiments.





Figure 2: A) Overlaid 1D 1H spectrum of 0.3% CHCl₃ in acetone-d₆. The spectrum was recorded with 32k points, i.e., 1 s acquisition. B) Field map recorded after shimming with (blue) and without optical fiber from the bottom (red). C) Overlaid [¹⁵N, ¹H]-HSQC spectra of KRAS G13D mutant with (blue) and without optical fiber from the bottom (red).

Finally, we assessed the photo-CIDNP NMR performances of our setup and compared them with the classic setup, where the optic fiber is coupled from the top of the NMR tube with the optic fiber dipped into the aqueous sample. The SNE is 74-fold and 60-fold when the sample is illuminated from the top and the bottom, respectively (Figure 3). The difference is largely explained by the length of the optical path, i.e., the distance between the light entering the sample and the sample volume detected by the probe. Indeed, with the optical fiber inserted into the sample, the light path before reaching the measurement zone is approximately 5 mm, while when the optical fiber is inserted at the bottom, the light path is approximately 10 mm. Using Beer-Lambert-Law,(Taniguchi et al., 2018; Dixon et al., 2005; Beer, 1852) and a molar absorption coefficient of approximately 11000 1/(M*cm) at a wavelength of 450nm and a fluorescein concentration of 10 μ M we estimated a transmission of 77.6%, after a 10 mm path length (bottom excitation) and 88.1% after a 5mm path length (top excitation). A second effect stems from Fresnel reflections,(Beer, 1852) at the interfaces between fiber core to air, air to glass and glass to water. The combined loss due to reflection was simulated to be 8.6%.

On the contrary, inserting the fiber directly in the aqueous solution results in only 0.3% Fresnel reflection due to the small difference in the index of refraction between the fiber core and the liquid. The combined effect of Fresnel reflection and absorption losses after a 10 mm path difference leads to power transmission of 71.1% of the initial fiber power in the case of the excitation from the bottom and 87.7% in the case of excitation of the immersed fiber (from the top). The ratio of the two transmitted powers for different setups, respectively for the bottom and from the top, is equal to 0.81 (0.711/0.877), which is in good agreement with the observed ratio of the differences in the photo-CIDNP signal of 60/74= 0.81. Therefore, a simple

way to improve the photo-CIDNP hyperpolarization could be to use stronger laser powers to compensate for the optical density and the Fresnel reflection. However, this would yield more photosensitizer bleaching and should only be considered for single or low number of scan applications. Optimizing the photosensitized concentration can also improve the homogeneity of the illumination and maximize the photo-CIDNP hyperpolarization (Figure S7). In both setups, detecting a low concentration of tryptophan ($30 \mu\text{M}$) with a single scan NMR experiment was possible. For comparison, in fragment screening with saturation transfer difference (STD) NMR screening, the ligand concentrations vary between 50 to $500 \mu\text{M}$, and a measurement time of 15-60 min. Boltzmann polarization NMR measurement time ranges from 15 minutes to 1 hour, depending on the instrument's sensitivity and concentration. With photo-CIDNP fragment screening, we previously demonstrated the possibility of screening at concentrations as low as $5 \mu\text{M}$ of ligand using single scan experiments,(Torres et al., 2023; Stadler et al., 2023) corresponding to 1-2 seconds. Recent advances in the ultralow concentration photo-CIDNP NMR measured tryptophan concentrations down to 20 nM ;(Yang et al., 2022) similar performances are expected to be possible with this setup as the light coupling is similar.



170 **Figure 3: A)** SNE evolution with illumination time, with illumination from the top (red) and from the bottom (blue). **B)** signal intensity after 2 seconds of illumination from the top (red) and from the bottom (blue). The sample contained $30 \mu\text{M}$ tryptophan and $10 \mu\text{M}$ fluorescein, with oxygen-quenching enzymes, GOCAT, and glucose (see Material and Methods).

Current autosamplers can exchange a sample every 2-3 minutes as the sample undergoes a temperature equilibration (1-2 min), a lift up and down of the sample (10-20 seconds), and the automated shimming procedure requires 30 seconds to 1 minute. 175 With the combination of photo-CIDNP hyperpolarized NMR experiments in the order of a few seconds and current commercial autosamplers, achieving a fully automated throughput of approximately 700 samples daily is possible. The implementation is straightforward and can likely be pushed by severalfold with optimization, fully automated. Irradiating from the bottom is also an easy solution for benchtops, which are often open on both sides.

Conclusion

180 The present work describes a simple way to introduce an optical fiber through a cryogenic probe's flow-cell accessory port to reach the near vicinity of NMR sample tubes. Such port is accessible in all 5-mm standard CryoProbes (but not for 3mm and 1.7mm CryoProbes) and Prodigy CryoProbes from 300 MHz-900 MHz. This allows the irradiation of the samples contained in NMR tubes from the bottom, taking advantage of the spherical bottom and resulting in a lensing effect. The light is, therefore, efficiently coupled with the sample tube, and a slight deviation in the positioning of the optical fiber is not critical. As the
185 optical fiber can be maintained sufficiently far away from the measuring and shimming regions of the probe, the influence on field homogeneity is minimal. It does not affect the performance of the spectrometer. This is not surprising as others have successfully designed similar systems before us for room temperature probes, using optical fiber and quartz rods, which are more prominent.(Tolstoy et al., 2009; Koeppe et al., 2011; Kuprov and Hore, 2004) Therefore, removing the optical fiber to perform other NMR experiments is unnecessary once installed. We verified that the photo-CIDNP performances were similar,
190 and we could measure relatively low tryptophan concentrations with single scan experiments. This design is the first to allow high throughput automated measurement of photo-CIDNP experiments as it is compatible with commercial autosamplers. It can be used equally for other light-coupled NMR experiments requiring high throughput. The presented setup will enable photo-CIDNP small molecule screening to achieve its maximal throughput, corresponding to the throughput of the commercially available autosamplers, as photo-CIDNP fragment screening is measured with single scan experiments (1-5
195 seconds). Such throughput is expected to increase with faster autosampler, adapted workflow (temperature equilibration), and accelerated shimming procedures.(Becker, 2022) Altogether, these advances will enable photo-CIDNP NMR small molecule screening to achieve a throughput of >1,500 samples daily, eventually up to 3,000.

Material and Methods

200 For the optical simulations, Ansys Zemax OpticStudio was used in non-sequential mode. The software performs a Monte Carlo simulation by computing the path of many randomized light rays through the geometry modeled according to the experimental setup. The simulation accounted for the absorption of fluorescein at 10 μM in water (0.10752 cm^{-1}), which was set to agree with the media's optical density (OD), and the scattering was neglected as it is expected to have a minor influence.
We built an optical benchtop setup that illuminated 3mm and 5mm NMR tubes with lensed tube bottoms with multimode
205 fibers of different NA and core diameters. The NMR tubes were coaxially held in standard Thorlabs 30mm cage plates, while positioners were used to change the lateral and axial distance of the fiber with respect to the NMR tube. We monitored optically reflected powers with a fiber-based beam splitter on the entrance port and transmitted powers with a photodiode sensor on top of the (uncapped) NMR sample tube. It was also possible to detect power levels at different axial positions within the liquid sample using an immersed photodiode combined with an external current-voltage converter (Figure S6).

210 The fiber is inserted from the entry of the air conduct, which is typically connected to the BCU. The connection to the BCU is maintained using a T-connector to accommodate the optic fiber and the BCU tubing. The fiber is inserted until the user feels the resistance of the NMR tube, then the optic fiber is pulled down by a few millimeters.

All NMR measurements were performed at 298 K on a Bruker Avance III HD 600 MHz spectrometer equipped with a Bruker TCI 600 $^1\text{H}/^{13}\text{C}/^{15}\text{N}$ MHz CryoProbe. The laser used was a Thorlabs L450P1600MM, a diode laser emitting at 450 nm. The 215 laser light was coupled (using appropriate coupling optics) into an optical fiber (Thorlabs, FG950UEC) with a length of 10 m and a diameter of 0.4 mm. The light power output by the laser diode is 1.6 W, and the light power measured at the optic fiber output is 1.0 W due to loss during laser diode-optic fiber coupling.

Photo-CIDNP experiments were performed at 30 μM tryptophan concentration in 100 mM potassium phosphate buffer at pH = 7.2, GOCAT enzyme with glucose as described herein, and 10 μM fluorescein. To prevent photosensitizer quenching, the 220 enzyme cocktail glucose oxidase (GO, 120 kDa), catalase (CAT, 240 kDa), and d-glucose (G, 180 Da) were used at a concentration of 200 nM, 140 nM, and 2.5 mM, respectively. (10,37) The stock solutions were 4.0 μM for Go and 4.0 μM for Cat, respectively, in 10 mM NaPO₄ buffer and pH = 7.2. The glucose stock solution was 500 mM in D₂O with 0.02% NaN₃.

The reference sample from Bruker was used for the line shape evaluation, i.e., 0.3% CHCl₃ in acetone-d₆.

[¹⁵N, ¹H]-HSQC spectrum was measured at 600 MHz ¹H NMR frequency with 600 μM ¹⁵N-labeled KRAS in 20 mM HEPES, 225 100 mM NaCl, 5 mM MgCl₂, 2mM TCEP, pH = 7.4, at a temperature of 298 K. Typically, 256 ($t_{1,\max}({}^{15}\text{N}) = 47.6$ ms) \times 2048 ($t_{2,\max}({}^1\text{H}) = 60.8$ ms) complex points, an interscan delay of 0.8 s, and 16 scans per increment were measured. The data was zero-filled to 4096 points in the direct proton dimension and 512 points in the ¹⁵N-dimension. Processing was done with a shifted cosine window function for both dimensions.

Data availability

230 All the data are available upon request to one or the corresponding authors.

Author contribution

WW designed, performed the experiments, and participated in writing; PG performed the simulations and participated in writing; AR brought the idea and consulted on the NMR tube as a lens; MB produced the KRAS sample; RR supervised the project and participated in writing; FT supervised the project and wrote the manuscript.

235 Competing interests

MB, RR and FT are co-founders of NexMR AG, which commercializes the technology as CryolightTM.

Acknowledgment

The authors thank Innosuisse for financial support (66876.1 INNO-LS).

References

- 240 Bargon, J. and Fischer, H.: Kernresonanz-Emissionslinien Wahrend Rascher Radikalreaktionen .2. Chemisch Induzierte Dynamische Kernpolarisation, Zeitschrift für Naturforschung A, 22, 1556-1562, 1967.
- Becker, M., Lehmkuhl, S., Kesselheim, S., Korvink, Y., Jouda, M.: Acquisitions with random shim values enhance AI-driven NMR shimming, Journal of Magnetic Resonance, 345, 107323, 2022.
- Beer, A.: Bestimmung der Absorption des rothen Lichts in farbigen Flüssigkeiten, Annalen der Physik, 162, 78-88, 1852.
- 245 Bodenhausen, G. and Ruben, D. J.: Natural Abundance N-15 NMR by Enhanced Heteronuclear Spectroscopy, Chemical Physics Letters, 69, 185-189, 1980.
- Bramham, J. E. and Golovanov, A. P.: Sample illumination device facilitates in situ light-coupled NMR spectroscopy without fibre optics, Communication Chemistry, 5, 90, 2022.
- Closs, G. L.: A Mechanism Explaining Nuclear Spin Polarizations in Radical Combination Reactions, Journal of the American 250 Chemical Society, 91, 4552-4554, 1969.
- Dixon, J. M., Taniguchi, M., and Lindsey, J. S.: PhotochemCAD 2: a refined program with accompanying spectral databases for photochemical calculations, Photochem Photobiol, 81, 212-213, 2005.
- Greivenkamp, J. E.: Field Guide to Geometrical Optics, SPIE., 2004.
- Hore, P. J. and Broadhurst, R. W.: Photo-CIDNP of Biopolymers, Progress in Nuclear Magnetic Resonance Spectroscopy, 25, 255 345-402, 1993.
- Kaptein, R. and Oosterhoff, L. J.: Chemically induced dynamic nuclear polarization III (anomalous multiplets of radical coupling and disproportionation products), Chemical Physics Letters, 4, 214-216, 1969.
- Kerber, P. J., Nunez, R., Jensen, D. R., Zhou, A. L., Peterson, F. C., Hill, R. B., Volkman, B. F., and Smith, B. C.: Fragment-based screening by protein-detected NMR spectroscopy, Methods in Enzymology, 690, 285-310, 2023.
- 260 Koeppe, B., Tolstoy, P. M., and Limbach, H. H.: Reaction pathways of proton transfer in hydrogen-bonded phenol-carboxylate complexes explored by combined UV-vis and NMR spectroscopy, Journal of the American Chemical Society, 133, 7897-7908, 2011.
- Kuhn, L. T.: Photo-CIDNP NMR spectroscopy of amino acids and proteins, Topics in Current Chemistry, 338, 229-300, 2013.
- Kuhn, L. T., Weber, S., Bargon, J., Parella, T., and Perez-Trujillo, M.: Hyperpolarization-Enhanced NMR Spectroscopy of 265 Unaltered Biofluids Using Photo-CIDNP, Analytical Chemistry, 96, 102-109, 2024.
- Kuprov, I. and Hore, P. J.: Uniform illumination of optically dense NMR samples, Journal of Magnetic Resonance, 171, 171-175, 2004.
- Morozova, O. B. and Ivanov, K. L.: Time-Resolved Chemically Induced Dynamic Nuclear Polarization of Biologically Important Molecules, Chemphyschem, 20, 197-215, 2019.
- 270 Okuno, Y. and Cavagnero, S.: Fluorescein: A Photo-CIDNP Sensitizer Enabling Hypersensitive NMR Data Collection in Liquids at Low Micromolar Concentration, The Journal of Physical Chemistry B, 120, 12416-12416, 2016.
- Sobol, A., Torres, F., Aicher, A., Renn, A., and Riek, R.: Atto Thio 12 as a promising dye for photo-CIDNP, The Journal of Chemical Physics, 151, 234201, 234201, 2019.
- Stadler, G. R., Segawa, T. F., Butikofer, M., Decker, V., Loss, S., Czarniecki, B., Torres, F., and Riek, R.: Fragment Screening 275 and Fast Micromolar Detection on a Benchtop NMR Spectrometer Boosted by Photoinduced Hyperpolarization, Angewandte Chemie International Edition, 62, e202308692, 2023.
- Taniguchi, M., Du, H., and Lindsey, J. S.: PhotochemCAD 3: Diverse Modules for Photophysical Calculations with Multiple Spectral Databases, Photochemistry Photobiology, 94, 277-289, 2018.
- Tolstoy, P. M., Koeppe, B., Denisov, G. S., and Limbach, H. H.: Combined NMR and UV/vis spectroscopy in the solution 280 state: study of the geometries of strong OH-O hydrogen bonds of phenols with carboxylic acids, Angewandte Chemie International Edition, 48, 5745-5747, 2009.
- Torres, F., Sobol, A., Greenwald, J., Renn, A., Morozova, O., Yurkovskaya, A., and Riek, R.: Molecular features toward high photo-CIDNP hyperpolarization explored through the oxidocyclization of tryptophan, Physical Chemistry Chemical Physics, 23, 6641-6650, 2021.
- 285 Torres, F., Büttikofer, M., Stadler, G. R., Renn, A., Kadavath, H., Bobrovs, R., Jaudzems, K., and Riek, R.: Ultrafast Fragment Screening Using Photo-Hyperpolarized (CIDNP) NMR, Journal of the American Chemical Society, 145, 12066-12080, 2023.

Torres, F., Renn, A.; Riek, R.: Exploration of the close chemical space of tryptophan and tyrosine reveals importance of hydrophobicity in CW-photo-CIDNP performances, *Magnetic Resonance*, 2, 321-329, 2021.

290 Ward, H. R. and Lawler, R. G.: Nuclear Magnetic Resonance Emission and Enhanced Absorption in Rapid Organometallic Reactions, *Journal of the American Chemical Society*, 89, 5518-5519, 1967.

Williamson, M. P.: Using chemical shift perturbation to characterise ligand binding, *Progress in Nuclear Magnetic Resonance Spectroscopy*, 73, 1-16, 2013.

295 Yang, H. M., Li, S. Y., Mickles, C. A., Guzman-Luna, V., Sugisaki, K., Thompson, C. M., Dang, H. H., and Cavagnero, S.: Selective Isotope Labeling and LC-Photo-CIDNP Enable NMR Spectroscopy at Low-Nanomolar Concentration, *Journal of the American Chemical Society*, 144, 11608-11619, 2022.