

Bimodal Q-band Probehead with Improved Signal-to-Noise Ratio in Pulse EPR

Vasyl Denysenkov¹, Alexey Fedotov², Burkhard Endeward¹, Thomas F. Prisner¹

¹Institute for Physical and Theoretical Chemistry, Goethe University Frankfurt, 60438, Germany

5 ²A.V. Gaponov-Grekhov Institute of Applied Physics of the Russian Academy of Sciences, Nizhny Novgorod, 603950, Russia

Correspondence to: Vasyl Denysenkov (Denysenkov@em.uni-frankfurt.de)

Abstract. In addition to the development of various resonators, the concept of a probehead equipped with an additional low noise amplifier (LNA) is becoming increasingly popular to enhance the sensitivity of EPR spectrometers. The low noise detection amplifier makes it possible to measure pulsed EPR signals with high sensitivity. However, a strong reflected pulse
10 signal can cause saturation and deterioration of the LNA characteristics, which requires protection of the LNA (for example, by using a protection switch in front of the LNA), which in turn reduces the signal-to-noise ratio. To overcome these limitations, we propose using an EPR probehead based on a bimodal cavity with strong isolation between the input and output ports, in combination with a low noise amplifier connected to the cavity output. The experiments demonstrate a 4-fold increase in the signal-to-noise ratio (SNR) of a bimodal probe head operating in transmission mode compared to its operation in
15 reflection mode, which was achieved thanks to the additional use of LNA. Performance of the probe was also compared with the Bruker EN 5107D2 probe available in our laboratory, which showed an improvement that can be achieved by increasing the SNR by 2 times due to additional LNA and isolation of the detection channel from the input signal, and by 3.3 times due to a larger sample volume in the bimodal probe (~20 μ l) at Q-band frequencies compared to the Bruker one (~6 μ l). The developed probehead can be used together with commercial Bruker ELEXYS EPR spectrometers without modification of the
20 microwave bridge.

1 Introduction

Electron paramagnetic resonance (EPR) spectroscopy is a well-established method for studying systems with unpaired electrons. It is widely used in research areas such as chemistry, physics, medicine, biology, and materials science. Increasing the sensitivity of EPR spectrometers is important for the development of new methods that open up new application
25 possibilities.

The central component of any conventional EPR spectrometer is a resonator, which amplifies the excitation as well as the induced microwave (mw) signal in the sample, thereby determining the sensitivity of measurements. Single-mode cavities (Reijerse et al., 2012) and dielectric resonators (Hyde and Mett, 2017; Raitsimring et al., 2012) are the most commonly used resonators operating in the reflection mode, which are used in almost all EPR spectrometers. For some specific applications,
30 more sophisticated options such as loop-gap resonators (LGR) (Hyde and Froncisz, 1989; Rinard and Eaton, 2005; Simovic et


al., 2006; Forrer et al., 2008; Tschaggelar et al., 2017), photonic band gap resonators (Milikisiyants et al., 2018), plasmonic metasurface resonators (Tesi et al., 2021), microresonators (Usevicius et al., 2025; Twig et al., 2013), Fabry-Perot resonators (Tipikin et al., 2010; Neugebauer and Barra, 2010; Budil and Earle, 2004) were developed for measurements in reflection mode depending on the used mw frequency range and the condition of the sample. A further example is a dual-mode cavity that matches both mw excitation frequencies in experiments with pulsed electron-electron double resonance (PELDOR/DEER), which increases the sensitivity of such experiments with a large frequency separation of the mw excitation pulses (Tkach et al., 2011). However, when operating in reflection mode, a significant part of the mw excitation power returns to the mw bridge due to resonator ring-down under pulse EPR conditions, and may reduce the sensitivity of the spectrometer receiver due to insertion losses in the protection gate switches in pulse mode, and due to the mw source noise in continuous wave (CW) mode.

A well-known approach to avoid the ringing as well as to reduce the problem with the source noise is the use of a bimodal cavity in which two modes with orthogonal H-field polarization resonate at the same frequency permitting detection of the orthogonal component of the circularly polarized induction signal, i.e. to excite the x-component of the magnetization and to detect the y-component (Hujsien and Hyde, 1974; Mailer et al., 1980; Barendswaard et al., 1984; Prisner and Dinse, 1989). This approach has also been used with loop-gap resonators (Piasecki et al., 1996) and cross-loop resonators (Rinard et al., 1996), as well as for a non-resonant probehead (Smith, et al., 2008).

Another key element for determining the SNR of the detected EPR signal is a low noise amplifier (LNA) used directly after the resonator. It is becoming increasingly popular to improve spectrometer sensitivity (Bienfait, et al., 2016; Pfenninger et al., 1995; Simenas et al., 2021; Kalendra et al., 2023; Jbara et al., 2025; Rinard et al., 1999). An additional LNA inserted into the probehead can help to minimize noise contribution of the circulator or any similar transmit/receive decoupling circuit, as well as a protection switch for the LNA inside the mw bridge of the EPR spectrometers. In principle, if the LNA is used as the first device after the resonator, then all other components no longer play any significant role for the SNR. However, in most cases, the compatibility of such probes with commercial EPR spectrometers becomes a non-trivial task requiring some modification of the mw bridge. In particular, strong ring-down signal can lead to the LNA saturation or damage, while protection switches, in turn, result in additional noise. Here we demonstrate the use of a bimodal cavity embedded in a Q-band probe, combined with an additional LNA to improve the sensitivity of a Q-band EPR spectrometer. For our test measurements, the LNA is placed outside the cryostat at room temperature to avoid complications caused by the external static magnetic field and low temperatures. The probehead can be used with commercial EPR spectrometers without any modification of the setup. The probe was tested by pulsed EPR with various samples at room temperature and at 80 K. The isolation between the excitation and detection mode makes it possible to take full advantages of the LNA and noticeably improve the sensitivity of the spectrometer.

2 Development of the probehead

The block-diagram of the probehead with a cavity is shown in the Fig. 1.



65 **Figure 1: Block-diagram of the Q-band EPR probehead. The waveguide switch is shown in the position 1-2 for transmission mode operation. Reflection mode is available when the switch is in the position 1-3.**

The probehead is designed to operate with the Bruker ELEXSYS E580 EPR spectrometer. The probehead dimensions are compatible with that of Bruker flexline Q-band resonators and it fits perfectly into an Oxford CF935 Helium flow cryostat to
70 be able to operate at temperatures in the range of 5 – 300 K. The probe can be connected to the Q-band bridge of the Bruker ELEXSYS E580 EPR spectrometer by means of a standard WR-28 waveguide.

In the probe we use a bimodal cavity based on the design described by James Hyde and coworkers, scaling it up for Q-band applications. This is a bimodal cavity (Fig. 2) in which two rectangular TE₁₀₃ modes are polarization-crossed and have two half-wavelengths common (Hyde et al., 1968). The great merit of this kind of cavity is the significant isolation between the
75 input and output modes that can be achieved with sample volume of 20-50 microliters over a wide temperature range.





Figure 2: Bimodal TE103 cavity. On the left: 1 – input port with coupling iris 2.2 mm diameter, 2 – quartz rod 1 mm diameter for input mode frequency tuning; 3 – teflon tube 4mm OD / 3 mm ID for resonator protection against impurities; 4 – quartz sample tube 2.8 mm OD; 5 – quartz rod 1.5 mm diameter for output mode frequency tuning; 6 – output port with coupling iris 3 mm diameter. Input and output ports have similar 7.4×3.7 mm cross-sections. All other dimensions of the structure are included in a 3D model in SI (Supplementary Information). On the right: microwave B_1 field distribution along the cavity is demonstrating the decoupling between output and input ports (simulated by CST Suite).

The probehead can operate in reflection or transmission mode, depending on the position of the manual waveguide switch (530B/383 MI-Wave Inc., USA). The reflection mode of operation is typical for conventional probes and will not be described further here. In transmission mode the probehead operates in combination with a LNA. Our used LNA (Model JS-426004000-27-10P Narda-MITEQ, USA) has a 1.9 dB noise figure at 20 °C, which corresponds to an equivalent noise temperature of 160 K. For our test experiments it is placed outside the sample cryostat. The uniline (4IWN32-2 Dorado Int., USA) protects the LNA from reflected mw power. A circulator (Model 179B-34/383 Anritzu Inc., Japan) with 0.13 dB insertion loss and 33 dB isolation at 34 GHz was chosen to direct the mw power from the Q-band bridge to the bimodal cavity and from the LNA back to the receiver part of the spectrometer.

3 Experimental results

The probehead loaded with a frozen aqueous solution (with a dielectric constant of $\epsilon = 3.4$ and a loss tangent of $\text{tg}\delta = 0.01$) in a quartz sample tube has been simulated in the 33 - 34.5 GHz range by finite element calculations with CST Suite version 2021. The simulation results are frequency-dependent s-parameters: S11 is the reflection coefficient of the input cavity mode, S22 is the reflection coefficient of the output cavity mode, and S21 is the transmission coefficient between input and output of the structure. Both modes of the cavity are tuned to the same frequency of 33.192 GHz. In this case isolation between the input and output modes characterized by S21 curve is approximately 51 dB at the frequency of interest (Fig. 3) which was reached without additional tuning paddles.



105 **Figure 3: Microwave properties of TE103 bimodal c simulated by CST Suite: S11 (red) - input port reflection coefficient; S22 (black)**
 110 **- output port reflection coefficient; S21 (blue) - transmission coefficient that indicates output-to-input isolation between input and**
output ports.

The resonance frequencies of the empty cavity are higher than the spectrometer frequency range and are shifted by the sample in the quartz tube OD = 2.8 mm, $\epsilon = 3.8$ (Rototec-Spintec, USA) down into the 33 – 34.5 GHz range. Microwave performance of the probehead was tested using a network analyzer (ZVA-40 Rohde&Schwarz) at 294 K on a sample of BDPA:PS powder and on a sample of 0.1 mM OXO TEMPO (Sigma-Aldrich GmbH) in toluene at 80 K. The results of the experimental tests on the OXO TEMPO sample are shown in Fig. 4 with traces representing the reflection coefficient S11 of the input mode (a), the reflection coefficient S22 of the output mode (b), and the transmission coefficient S21 which characterizes the input-to-output isolation (c).






Figure 4: a) Return loss of the probehead input for switch position 1-3; b) Return loss of the cavity output when the network analyzer is connected to the output position as indicated in Fig. 1; c) Input-to-output isolation of the probehead without the LNA. On the left: input and output modes are tuned to the same frequency of 34.208 GHz. On the right: output mode was detuned to 34.342 GHz (134 MHz above the input mode).

Traces (a) and (b) on the network analyzer screen are return loss traces. Their shift down to ~ 4 dB and ~ 8 dB relative to 0 dB level is due to propagation losses in UT 141 coaxial cables with SMA connectors connecting the probe to the network analyzer and insertion losses in the mw components inside the probe. All together mismatches in long transmission paths cause some standing waves in traces a) and b). Trace (c) indicates the minimum isolation at the resonant position. Logically, if the two modes resonate at the same frequency (Fig.4 on the left), then the isolation will be lower compared to the case of any frequency offset between the two modes (Fig. 4, on the right). In the presented graph a 134 MHz frequency offset between both modes provides higher isolation by 5 dB. This dependence of isolation on frequency offset can be used as a tool for indirectly monitoring the output mode and adjusting it to the frequency of the input mode by means of Xepr on the Bruker ELEXYS E580.

The measured decoupling between the input and output of the resonator is approximately: $(52 \text{ dB} - (4 \text{ dB} + 8 \text{ dB})/2) = 46 \text{ dB}$.

130 This experimental value is few dBs worse than the simulated one due to the imperfections of the inner surfaces of the fabricated resonator structure and due to the presence of a sample tube which may slightly shift from the axis in the experiments. Q-factors of the input and output modes of the bimodal resonator are $Q_{\text{input}} = 250$ and $Q_{\text{output}} = 180$.

The assembled probehead was tested using electron spin echo (ESE) experiments on a Bruker ELEXYS E580 EPR
135 spectrometer equipped with a 150 W TWT amplifier (Applied Systems Engineering Inc., USA). For testing at room temperature we use BDPA:PS powder with a total of 10^{15} spins in a 0.5 mm ID fused quartz capillary (VitroCom, USA). The sample capillary was placed in the probehead with a 2.8 mm OD sample tube and measured in both reflection and transmission modes when both modes were set to the same frequency. The same sample in the same capillary was also measured with a Bruker EN 5107D2 probehead. The echo signals of both probeheads (bimodal and Bruker EN 5107D2), normalized to the
140 same noise level are presented in Fig. 5.



145 **Figure 5: Hahn echo with BDPA:PS in a 0.5 mm ID quartz capillary measured at 294 K in: 1 – bimodal probehead in transmission mode with LNA (Patt=18 dB; $\pi/2$ -pulses = 80 ns; π -pulses = 160 ns); 2 - bimodal probehead without LNA in reflection mode; 3 - EN 5107D2 probehead (Patt= 24 dB; $\pi/2$ -pulses = 80 ns; π -pulses = 160 ns); 4 - EN 5107D2 probehead (Patt= 0 dB; $\pi/2$ -pulses = 6 ns; π -pulses = 12 ns). All traces were recorded with single shot per point, and with 400 ns delay between the pulses.**

In the case of the bimodal probehead the mw attenuation was set to 18 dB in order to avoid overload and damage of the LNA that resulted in optimal $\pi/2$ -pulses = 80 ns; π -pulses = 160 ns pulses. A similar pulse sequence was used with the EN 5107D2 probehead, setting mw attenuation to 24 dB. This test showed a 4-fold improvement in the signal-to-noise ratio (SNR) for the bimodal probehead equipped with the LNA. For comparison, full mw power (0 dB attenuation) was also applied to the EN
150 5107D2 probehead with optimal $\pi/2$ -pulses = 6 ns; π -pulses = 12 ns pulses. In this case the SNR improvement using the bimodal probe is still 2 times achieved by measuring the peak amplitudes of the echo signals. For inhomogeneously broadened spectra the echo width follows to the excitation pulse width in time domain traces. That is the reason that the echoes excited by short (12 ns) pulses look much shorter in comparison with the echoes excited by 160 ns pulses (Fig. 5). In case of longer

pulses the burned hole in the EPR spectrum is narrower causing smaller number of excited spins with respect to the case of shorter excitation pulses resulting in a difference of the signal intensities between trace 3 and trace 4 that is usually measured by integrated echoes. In the case of excitation by pulses of similar pulse length the echoes can be compared also by peak intensities because it is the simplest and accurate approach to show the better performance of the bimodal probehead with LNA (trace 1) with respect to its operation without LNA (trace 2) as well as the Bruker probe (trace 3). However, in comparison with shorter pulses (trace 4) integrated echo values can be used. In this case the better performance of the bimodal probe with LNA is also evident if integrated intensities instead of peak intensities for traces 1 and 4 will be used. Q factors of the input and output modes of the bimodal resonator have been chosen low enough ($Q_{\text{input}} = 250$ and $Q_{\text{output}} = 180$) to reach a broad EPR excitation.

We also accomplished another Hahn echo experiment using a 0.1 mM TEMPO in toluene sample in a 2.8 mm OD and a 1.6 mm OD sample tubes for the bimodal and Bruker EN 5107D2 probeheads respectively. This experiment was performed at 80 K. The experimental results of this test are shown in Fig. 6.




Figure 6: Spin echo with 0.1 mM OXO TEMPO in toluene measured at 80 K by a typical Hahn echo experiment in: black – a 2.8 mm OD sample tube inside the bimodal probehead (Patt=15 dB; $\pi/2$ -pulses = 50 ns; π -pulses = 100 ns); red – a 1.6 mm OD sample tube inside the EN 5107D2 probehead (Patt= 22 dB; $\pi/2$ -pulses = 50 ns; π -pulses = 100 ns). All traces were recorded with single shot per point, and with 400 ns delay between the pulses. Video gain is 6 dB. In the region between 200 ns and 1100 ns where the spectrometer protection gate is ON the noise for the bimodal cavity trace (black one) is lower because it is reduced proportionally due to normalization step of the large signal produced by the LNA which has 30 dB gain.

175 In this case signal-to-noise enhancement is reaching factor of 7 but only with a lower mw power and longer pulse lengths. The difference of the results of this test and the previous one is mainly due to larger number of spins in the 2.8 mm OD sample tube used in the bimodal probehead with respect to the 1.6 mm OD sample tube used in the Bruker probehead resulted in sample volumes of 20 μ l and 6 μ l correspondingly. Thus the 3.5 times enhancement is achieved due to a larger sample volume in the bimodal probe, and additional 2 times SNR improvement is due to the LNA application. It should be noted that the reliability of this comparison depends on the specific performance of the commercial EN 5107D2 probe. Our EN 5107D2 probehead is not a new one and may have a slightly lowered mw power conversion factor compared to other (new) commercial probes. All presented experiments have been accomplished with the available spectrometer hardware and software. 180
Obtained SNR enhancements (peak amplitude values) are summarized in the Table 1:

185 Table 1. Signal-to-noise enhancements for the bimodal probehead with respect to the EN 5107D2

Sample	Pulsewidth, ns	SNR enhancement, times	Remarks
BDPA:PS @ 294 K	80/160 6/12	4 2	due to mw power limitations in bimodal probe
OXO TEMPO @80 K	50/100	7	
			3.5 times – due to difference in sample volumes and 2 times - due to the additional LNA

4 Discussion

190 The aim of this study is a proof-of-principle demonstration of the improvements obtained by using a bimodal resonator combined with a LNA in the detection channel directly after the resonator. There is still the prerequisite to use reduced mw pulse power for the spin excitation due to the limited isolation of 46 dB between the input and output of the mw resonator. This means that at the moment it is not possible to use the full 150 W power of the travelling wave tube amplifier. However, despite the restrictions caused by the modest input-to-output isolation of our newly developed probehead, we were able to demonstrate a significant SNR enhancement compared to the standard reflection mode operation of the probe, as well as with a commercial Q-band probehead. The developed probehead may be interesting for use in time-resolved EPR methods (Biskup, 2019) that do not require such high mw power. Such methods include transient EPR (Niklas and Poluektov, 2017; Tait et al., 2015) and non-adiabatic rapid scan EPR (NARS) (Kittell et al., 2011; Stoner et al., 2004; Rokeakh and Artyomov, 2023). In 195

addition, for pulse EPR applications with broadband mw pulses the bimodal resonator might offer a significant advantage in
200 avoiding standing waves compared to resonators in the reflection mode (Trenkler et al., 2025).
The isolation level can be further improved by introducing tune paddles into the resonator (Mailer et al., 1980), which we plan
to do in the future to extend the probe to high mw power. Another improvement will be a cryogenic LNA which is placed into
the cryostat to reduce the noise temperature at the input of the LNA from 160 K to supposed 10-50 K (Kalendra et al., 2023)
as well as to further reduce the noise figure of the LNA itself. However, the presence of strong magnetic fields can impair the
205 operation of the LNA due to the Hall effect (Harrysson Rodrigues et al., 2019) which should be eliminated by proper shielding
from the magnetic field or careful orientation of the LNA. In addition, the problem associated with repeated cooling and
warming cycles of the probe can lead to a shorter LNAs lifetime.

5 Conclusions

This work showed that a bimodal cavity in combination with a LNA connected to the mw resonator in transmission mode led
210 to an improvement in SNR in pulse EPR experiments performed at Q-band frequencies. For simplicity the LNA is placed
outside the cryostat at room temperature which provides a noise figure of 1.9 dB at 34 GHz. As a result, we achieved an
experimental SNR enhancement factor of 2 to 4 regardless of the sample temperature and composition of the sample. Another
feature of the proposed probehead is its compatibility with commercial Bruker ELEXYS EPR spectrometers without any
modification of the mw bridge.

215 Data availability

The Supplement contains the following files: bimodal cavity simulation TE103_induction.cst that was used to optimize
geometry of the structure

Supplement

The supplement information related to this article is available online at:

220 Author contribution

VD and TP designed the experiments and BE carried them out. VD and AF developed the model and performed the
simulations. VD prepared the manuscript with contributions from all co-authors.

Competing interests

At least one of the (co-)authors is a member of the editorial board of Magnetic Resonance.

225 Acknowledgements

We thank the BMRZ for financial support of the EPR infrastructure and the DFG for funding of this project No 471089264. AF thanks financial support from the Institute of Applied Physics RAS under project No FFUF-2024-0027. VD, BE, TP would like to thank Thorsten Maly for the BDPA:PS sample prepared during his stay in the lab.

References

- 230 Barendswaard, W., Disselhorst, J. A. M., and Schmidt, J.: A bimodal cavity for reducing the dead-time in electron spin-echo spectroscopy, *J. Magn. Reson.*, 58, 477-483, [https://doi.org/10.1016/0022-2364\(84\)90152-5](https://doi.org/10.1016/0022-2364(84)90152-5), 1984.
- Bienfait, A., Pla, J., Kubo, Y., Stern, M., Zhou, X., Lo, C. C., Weis, C. D., Schenkel, T., Thewalt, M. L. W., Vion, D., Esteve, D., Julsgaard, B., Moelmer, K., Morton, J. J. L., and Bertet, P., Reaching the quantum limit of sensitivity in electron
235 spin resonance, *Nature Nanotech.*, 11, 253–257, <https://doi.org/10.1038/nnano.2015.282>, 2016.
- Biskup, T.: Structure–Function Relationship of Organic Semiconductors: Detailed Insights From Time-Resolved EPR Spectroscopy, *Front. Chem.* 7, 10, <https://doi.org/10.3389/fchem.2019.00010>, 2019.
- 240 Budil, D. E. and Earle, K. A.: Sample Resonators for Quasioptical EPR, in: Very High Frequency (VHF) ESR/EPR. Biological Magnetic Resonance, edited by: Grinberg, O. Y. and Berliner, L. J., Springer, Boston, MA, https://doi.org/10.1007/978-1-4757-4379-1_11, 2004.
- Forrer, J., Garcí'a-Rubio, I., Schuhmam, R., Tschaggelar, R., and Harmer, J.: Cryogenic Q-band (35 GHz) probehead featuring
245 large excitation microwave fields for pulse and continuous wave electron paramagnetic resonance spectroscopy: Performance and applications, *J. Magn. Reson.*, 190, 280–291, <https://doi.org/10.1016/j.jmr.2007.11.009>, 2008.
- Harrysson Rodrigues, I., Niepce, D., Pourkabirian, A., Moschetti, G., Schlee, J., Bauch, T., and Grahn, J.: On the angular dependence of InP high electron mobility transistors for cryogenic low noise amplifiers in a magnetic field, *AIP Advances*, 9,
250 085004, <http://doi.org/10.1063/1.5107493>, 2019.

- Hyde, J.S. and Froncisz, W.: Loop gap resonators, in: *Advanced EPR: Applications in biology and biochemistry*. A. J. Hoff (Ed.). Elsevier, Amsterdam, 277-305, ISBN 0-444-88050-X, 1989.
- 255 Hyde, J.S. and Mett, R.R.: EPR uniform field signal enhancement by dielectric tubes in cavities, *Appl. Magn. Reson.*, 48, 1185-1204, <https://doi.org/10.1007/s00723-017-0935-4>, 2017.
- Hyde, J.S., Chien, J.C.W., and Freed, J.H.: Electron–electron double resonance of free radicals in solution, *J. Chem. Phys.*, 48, 4211-4226, <https://doi.org/10.1063/1.1669760>, 1968.
- 260 Huisjen, M. and Hyde, J.S.: A pulsed EPR spectrometer, *Rev. Sci. Instrum.*, 45, 669–675, <https://doi.org/10.1063/1.1686710>, 1974.
- Jbara, M., Zgadzai, O., Harneit, W., and Blank, A.: Cryogenic W-band Electron Spin Resonance Probehead
 265 with an Integral Cryogenic Low Noise Amplifier, *Appl. Magn. Reson.*, 56, 265–284, <https://doi.org/10.1007/s00723-024-01732-1>, 2025.
- Kalendra, V., Turcak, J., Banyš, J., Morton, J.L., and Šimenas, M.: X- and Q-band EPR with cryogenic amplifiers independent of sample temperature , *Journal of Magnetic Resonance* 346, 107356, <https://doi.org/10.1016/j.jmr.2022.107356>, 2023.
- 270 Kittell, A. W., Camenisch, T. G., Ratke, J. J., Sidabras, J. W., and Hyde, J. S.: Detection of undistorted continuous wave (CW) electron paramagnetic resonance (EPR) spectra with non-adiabatic sweep (NARS) of the magnetic field, *J. Magn. Reson.*, 211, 228–233, <https://doi.org/10.1016/j.jmr.2011.06.004>, 2011.
- 275 Mailer, C., Thomann, H., Robinson, B.H., and Dalton, L.R.: Crossed TM_{110} bimodal cavity for measurement of dispersion electron paramagnetic resonance and saturation transfer electron paramagnetic resonance signals for biological materials, *Rev. Sci. Instrum.*, 51, 1714-1721, <https://doi.org/10.1063/1.1136162>, 1980.
- Milikisiyants, S., Nevzorov, A.A., and Smirnov, A.I.: Photonic band-gap resonators for high-field/high frequency EPR of
 280 microliter-volume liquid aqueous samples. *J. Magn. Reson.* 296, 152–164, <https://doi.org/10.1016/j.jmr.2018.09.006>, 2018.
- Neugebauer, P. and Barra, A-M.: New Cavity Design for Broad-Band Quasi-Optical HF-EPR Spectroscopy, *Appl. Magn. Reson.* 37, 833–843, <https://doi.org/10.1007/s00723-009-0092-5>, 2010.

- 285 Niklas, J., and Poluektov, O. G.: Charge transfer processes in OPV materials as revealed by EPR spectroscopy. *Adv. Energy Mater.* 7:1602226, <https://doi.org/10.1002/aenm.201602226>, 2017.
- Piasecki, W., Froncisz, W., and Hyde, J. S.: Bimodal loop-gap resonator, *Rev. Sci. Instrum.* 67, 1896–1904, <https://doi.org/10.1063/1.1147001>, 1996.
- 290 Pfenninger, S., Froncisz, W., and Hyde, J. S.: Noise analysis of epr spectrometers with cryogenic microwave preamplifiers, *J. Magn. Reson., Ser. A*, 113, 32–39, <https://doi.org/10.1006/jmra.1995.1052>, 1995.
- Prisner, T. and Dinse, K. P.: ESR with stochastic excitation, *J. Magn. Reson.*, 84, 296-308, [https://doi.org/10.1016/0022-](https://doi.org/10.1016/0022-2364(89)90373-9)
 295 2364(89)90373-9, 1989.
- Raitsimring, A., Astashkin, A., Enemark, J. H., Blank, A., Twig, Y., Song, Y., and Meade, T. J.: Dielectric resonator for K-band pulsed EPR measurements at cryogenic temperatures: probehead construction and applications, *Appl. Magn. Reson.*, 42, 441-452, <https://doi.org/10.1007/s00723-012-0313-1>, 2012.
- 300 Reijerse, E., Lendzian, F., Isaacson, R., and Lubitz, W.: A tunable general purpose Q-band resonator for CW and pulse EPR/ENDOW experiments with large sample access and optical excitation, *J. Magn. Reson.* 214, 237-243, <https://doi.org/10.1016/j.jmr.2011.11.011>, 2012.
- 305 Rinard, G. A. and Eaton, G.R.: Loop-Gap Resonators, in: *Biomedical EPR, Part B: Methodology, Instrumentation, and Dynamics. Biological Magnetic Resonance*, edited by: Eaton, S. R., Eaton, G. R., Berliner, L. J., 24/B, 19-52, Springer, Boston, MA, https://doi.org/10.1007/0-306-48533-8_2, 2005.
- Rinard, G. A., Quine, R. W., Ghim, B. T., Eaton, S. S., and Eaton, G. R.: Easily tunable crossed-loop (bimodal) EPR resonator, *J. Magn. Reson. A*, 122, 50-57, <https://doi.org/10.1006/JMRA.1996.0173>, 1996.
- 310 Rinard, G. A., Quine, R. W., Song, R., Eaton, G. R., and Eaton, S. S.: Absolute EPR spin echo and noise intensities, *J. Magn. Reson.*, 140, 69–83, <https://doi.org/10.1006/jmre.1999.1823>, 1999.
- 315 Rokeakh, A. I. and Artyomov, M. Y.: Low-Frequency NARS (LF NARS) by the use of a superheterodyne EPR spectrometer, *J. Magn. Reson.*, 349, 107402, <https://doi.org/10.1016/j.jmr.2023.107402>, 2023.

- Šimėnas, M., O'Sullivan, J., Zollitsch, C.W., Kennedy, O., Seif-Eddine, M., Ritsch, I., Hülsmann, M., Qi, M., Godt, A., Roessler, M. M., Jeschke, G., and Morton, J. L.: A sensitivity leap for X-band EPR using a probehead with a cryogenic preamplifier, *J. Magn. Reson.* 322, 106876, <https://doi.org/10.1016/j.jmr.2020.106876>, 2021.
- Simovic, B., Studerus, P., Gustavsson, S., Leturcq, R., Ensslin, K., Schumann, R., Forrer, J., and Schweiger, A.: Design of Q-Band loop-gap resonators at frequencies 34-36 GHz for single electron spin spectroscopy in semiconductor nanostructures, *Rev. Sci. Instrum.* 77 (6), <https://doi.org/10.1063/1.2206776>, 2006.
- Smith, G., Cruickshank, P., Bolton, D. R., and Robertson, D. A., High-field pulse EPR instrumentation, *Electron Paramagnetic Resonance*, 21, 216–233, <https://doi.org/10.1039/b807958g>, 2008.
- Stoner, J. W., Szymanski, D., Eaton, S. S., Quine, R. W., Rinard, G. A., and Eaton, G. R.: Direct-detected rapid-scan EPR at 250 MHz, *J. Magn. Reson.*, 170, 127–135, <https://doi.org/10.1016/j.jmr.2004.06.008>, 2004.
- Tait, C. E., Neuhaus, P., Peeks, M. D., Anderson, H. L., and Timmel, C.: Transient EPR Reveals Triplet State Delocalization in a Series of Cyclic and Linear π -Conjugated Porphyrin Oligomers, *J. Am. Chem. Soc.*, 137, 8284–8293, <https://doi.org/10.1021/jacs.5b04511>, 2015.
- Tesi, L., Bloos, D., Hrtoň, M., Beneš, A., Hentschel, M., Kern, M., Leavesley, A., Hillenbrand, R., Krápek, V., Šikola, T., and van Slageren, J.: Plasmonic Metasurface Resonators to Enhance Terahertz Magnetic Fields for High-Frequency Electron Paramagnetic Resonance, *Small Methods*, 5(9):e2100376, <https://doi.org/10.1002/smtd.202100376>, 2021.
- Tipikin, D. S., Earle, K. A., and Freed, J. H.: Variable Coupling Scheme for High-Frequency Electron Spin Resonance Resonators Using Asymmetric Meshes, *Appl. Magn. Reson.*, 37, 819-832, <https://doi.org/10.1007/s00723-009-0088-1>, 2010.
- Tkach, I., Sicoli, G., Höbartner, C., and Bennati, M.: A dual-mode microwave resonator for double electron–electron spin resonance spectroscopy at W-band microwave frequencies, *J. Magn. Reson.*, 341-346, <https://doi.org/10.1016/j.jmr.2011.01.012>, 2011.
- Trenkler, P., Endeward, B., Sigurdsson, S., and Prisner, T.: Optimized shaped pulses for 2D-SIFTER, *Magn. Reson. Discuss.* [preprint], <https://doi.org/10.5194/mr-2025-11>, in review, 2025.
- Tschaggelar, R., Breitgoff, F. D., Oberhänsli, O., Mian, Qi., Godt A., and Jeschke, G.: High-Bandwidth Q-Band EPR Resonators, *Appl. Magn. Reson.*, 48, 1273–1300, <https://doi.org/10.1007/s00723-017-0956-z>, 2017.

Twig, Y., Dikarov, E., and Blank, A.: Ultra miniature resonators for electron spin resonance: Sensitivity analysis, design and construction methods, and potential applications, Mol. Phys., 111:18-19, 2674-2682, 355 <https://doi.org/10.1080/00268976.2012.762463>, 2013.

Usevicius, G., Šimenas, M., Geoghegan, B. L., Kennedy, O. W., Pocius, I., Hogan, P., Villanueva Ruiz de Temino, A., Verstraete, J.-B., Verbaitytė, P., Chatziathanasiou, A., Antilen Jacob, G., Kamarauskas, M., Treideris, M., Gecys, P., Alexander, J., Kalendra, V., Banys, J., Roessler, M. M., and Morton J. L., Versatile High-Sensitivity EPR Using 360 Superconducting Spiral Microresonators, Small Methods, e01451, <https://doi.org/10.1002/smt.202501451>, 2025.

365