

1 **3D-printed microcell for protein NMR at high ionic strengths**
2 **and small sample volumes**

3

4

5 Tayeb Kakeshpour, Martin D. Gelenter, Jinfa Ying and Ad Bax

6

7 Laboratory of Chemical Physics, National Institute of Diabetes and Digestive and Kidney
8 Diseases, National Institutes of Health, Bethesda, MD 20892-0520, U.S.A.

9

10

11

12 ***Not for publication:***

13 Contact: Ad Bax, email: bax@nih.gov

14 Building 5, room 126


15 NIH, Bethesda, MD 20892-0520

16 USA

17 Ph +1 301 496 2848

18

Abstract. Standard solution NMR measurements use 5-mm outer diameter (OD) sample tubes that require *ca* 0.5 mL of solvent to minimize “end effects” on magnetic field homogeneity in the active volume of the sample. Shigemi cells reduce the solvent requirement to *ca* 0.29 mL. At high ionic strength, or at ultrahigh magnetic fields, smaller OD samples are needed to study samples in conductive, radiofrequency absorbing solvents such as water. We demonstrate an effective and inexpensive alternative for reducing the active sample volume to 0.13 mL by 3D printing of ellipsoidal shaped cells that are inserted into 5-mm OD NMR tubes. Static magnetic susceptibility, χ , of printer resin was measured using a simple slice-selection pulse sequence. We found that the χ of water increases linearly with NaCl concentration, from -9.05 ppm to -8.65 ppm for 0 to 2 M NaCl. The χ of D₂O was measured to be -9.01 ppm. The susceptibility difference between the resin (χ = -9.40 ppm) and water can be minimized by paramagnetic doping of the resin. Such doping was found unnecessary for obtaining high quality protein NMR spectra when using ellipsoidal shaped cells that are insensitive to susceptibility mismatching. The microcells offer outstanding RF and good B₀ homogeneities. Integrated 600-MHz HSQC signal intensities for the microcell sample in PBS buffer were 6.5±4% lower than for 0.5 mL of the same protein solution in a regular 5-mm sample tube. The cell is demonstrated for N-acetylated α -synuclein in PBS buffer, and for observing tetramerization of melittin at 2 M NaCl.



1 Introduction

Nearly all solution NMR measurements are carried out using standard 5-mm outer diameter (OD) sample tubes that require *ca* 0.5 mL of solvent to minimize “end effects” on magnetic field homogeneity in the active volume of the sample. Use of a Shigemi cell, which consist of glass with a magnetic susceptibility that is close to that of the selected solvent, can reduce the amount of sample required and thereby limit the cost of expensive protein preparations. However, the tubes are expensive and need to be matched to the magnetic susceptibility of the NMR solvent. The latter can be challenging considering that the susceptibility of water is somewhat temperature-dependent (Schenck, 1996) and increases substantially (becomes less negative) upon the addition of NaCl, whereas it increases by *ca* 0.04 ppm in D₂O versus H₂O (see below). The sample volume also can be reduced by inserting plugs, of a magnetic susceptibility close to that of the solvent, above and below the active volume of the sample (Barbara, 2009).

Magnetic susceptibility measurements of ionic solutions are often carried out using a magnetic field that oscillates at frequencies ranging from ~50 Hz to low MHz (Tsukada et al., 2006; Gutiérrez-Mejía and Ruiz-Suárez, 2012). However, for magnetic resonance purposes it is the magnetic susceptibility measured in a static magnetic field that is relevant to distortions in field homogeneity (Sangal et al., 2023) and to our efforts to develop a small volume sample cell for protein NMR.

We describe a simple method for measuring magnetic susceptibility of solid material in a high-resolution NMR spectrometer, and the development of a 3D printed microcell that can be inserted into a regular 5-mm NMR tube. The resin that we used for 3D printing at high resolution (25 μ m) has a reported magnetic susceptibility of $\chi = -9.34$ ppm (Sangal et al., 2023), which is well below that of water. However, by printing the cell with a spherical (Hizawa et al., 2017) or ellipsoidal geometry, magnetic field homogeneity within the cell becomes insensitive to the susceptibility mismatch between solvent and the printer resin (Schenck, 1996; VanderHart, 1996). The shape of our cell is perturbed by a narrow diameter, cylindrical access port that is needed to fill it with the NMR sample. Magnetic susceptibility mismatching effects, resulting from the deviation of a perfect ellipsoidal shape caused by this access port and by the finite resolution of the printer, can be minimized by paramagnetic doping of the printer resin with organic paramagnetic salt (Evans, 1959)

but is found unnecessary for routine applications. The cells are reusable, but the cost of printing such cells is minimal and recycling them therefore may not be necessary.

We note that susceptibility mismatching for non-ellipsoidal shapes can also be obtained by introducing nearby suitably shaped small “compensation structures” of different magnetic susceptibility, that effectively shim the sample to homogeneity for microfluidic applications (Ryan et al., 2014). Alternatively, the magnetic susceptibility of an aqueous solution can be increased by addition of Eu^{3+} -complexed diethyl-triamine pentaacetate (DTPA), which due to the short Eu^{3+} electron T_1 has minimal broadening effects on other solutes (Hale et al., 2018) but is restricted to cases where the solvent susceptibility is more negative than the surrounding material, which does not apply for 3D printer resin.

The microcell introduced by us for biological NMR spectroscopy purposes is particularly useful when the available sample quantity is limited, or when high ionic strength is required. High ionic strength lowers the quality factor (Q) of radiofrequency (RF) coils, and thereby negatively impacts NMR sensitivity, an effect that scales steeply with frequency (Ugurbil, 2018). RF penetration of water has been studied extensively for applications to magnetic resonance imaging, where it impedes the observation of tissue far from the body surface (Roschmann, 1987). RF absorption at high and ultrahigh magnetic fields also challenges solution ^1H NMR spectroscopy, where the use of pulses that are compensated for both offset and RF inhomogeneity (Freeman et al., 1980; Xia et al., 2017) becomes essential, in particular for the vast majority of advanced experiments that include multiple 180° pulses (Manu et al., 2023). Use of the 3D printed microcell greatly reduces problems with probe detuning, lowering of Q, and RF absorption. Consequently, use of the microcell results in short ^1H pulse widths with superior RF homogeneity, even at elevated ionic strength.

Magnetic susceptibilities of materials with application to magnetic resonance imaging (MRI) as well as microfluidic NMR technology commonly have employed MRI scanner equipment, and elegant methods applied to a wide range of materials were presented by Wapler et al. (Wapler et al., 2014). The same approach was recently used to measure magnetic susceptibility of 3D printed materials (Sangal et al., 2023), we demonstrate simple methods for deriving magnetic susceptibility differences between ionic solutions and printer material in a high-field solution NMR magnet. We also demonstrate that while

using only 130 μL of solvent, the sensitivity in common multi-dimensional NMR experiments, such as the gradient-enhanced HSQC (Kay et al., 1992), is comparable to what is obtained on a regular 500- μL sample. In another application, we show that high-quality spectra can be obtained for 90 μg of recombinantly expressed and chemically amidated melittin at 2 M NaCl concentration. Milligram-scale expression and purification of this uniformly ^{15}N -enriched peptide in its post-translationally modified state, which tetramerizes in a salt-dependent manner, is expensive and very labor-intensive (Gelenter and Bax, 2023).

2 Results and Discussion

2.1 Measurement of magnetic susceptibility

For all materials pertinent to high-resolution solution NMR, the induced magnetization \mathbf{M} depends linearly on the applied magnetic field \mathbf{H}_0 :

$$\mathbf{M} = \chi \mathbf{H}_0 \quad (1)$$

where χ is the volume magnetic susceptibility, often expressed in units of parts per million (ppm), i.e. χ is negative for diamagnetic media and positive for paramagnetic substances.

The magnetic flux density, \mathbf{B}_0 , is directly proportional to \mathbf{H}_0 :

$$\mathbf{B}_0 = \mu \mathbf{H}_0 \quad (2)$$

where the magnetic permeability μ corresponds to $\mu = (1+\chi)\mu_0$, and $\mu_0 = 4\pi \cdot 10^{-7} \text{ H/m}$ is the permeability of vacuum. In an NMR sample, the magnetic flux density corresponds to the sum of the applied magnetic field \mathbf{H}_0 and the integral of the magnetic field contributions from the induced magnetization over all sample volume elements at locations \mathbf{r} relative to the point of interest (Barbara, 1994). For a cylindrical sample of infinite length, with its axis parallel to a homogeneous applied magnetic field, the integral over all space within the cylinder is uniform across all locations, resulting in

$$\mathbf{B}_0 = \mu_0(1+\chi) \mathbf{H}_0 \quad (3)$$

Magnetic flux is conserved at any interface orthogonal to \mathbf{H}_0 between two media with magnetic susceptibilities χ_1 and χ_2 . For the example of a Shigemitsu sample cell, where the

bottom segment of the cylindrical tube consists of glass with susceptibility χ_g , and χ_s is the solvent susceptibility above it (Figure 1a), the flux density at the interface is given by

$$\mathbf{B}_0 = \mu_0[1 + (\chi_g + \chi_s)/2] \mathbf{H}_0 \quad (4)$$

Away from the interface, the integrated contributions to \mathbf{B}_0 from volume elements below and above the interface depend on the height above the interface as well as the transverse location, with values converging to $\mathbf{B}_0 = \mu_0(1 + \chi_s)\mathbf{H}_0$ at distances above the interface that are large relative to the tube inner diameter, D . Therefore, when selectively observing solvent signal with a chemical shift of δ_0 from a very thin slice perpendicular to the sample axis at height h above the interface (Figure 1a), a narrow signal at a frequency $\delta(h) = [1 + (\chi_g + \chi_s)/2]\delta_0$ is observed for $h \ll D$, with $\delta(h)$ approaching $(1 + \chi_s)\delta_0$ for $h \gg D$, and a complex line shape for slices taken at intermediate values (Barbara, 1994).

Hence, for a solution above the solid glass of a Shigemi cell with a total solvent height that is large relative to both D and the height of the receiver coil, the line shape corresponds to $\delta(h)$ contributions ranging from $[1 + (\chi_g + \chi_s)/2]\delta_0$ to $(1 + \chi_s)\delta_0$. In other words, the total width of the line shape at its base corresponds to half the difference between χ_g and χ_s in units of ppm. The above analysis does not take into account the B_0 gradient inside the NMR tube that is caused by the large magnetic susceptibility mismatch at the bottom of the tube. However, when assuming that the magnetic susceptibility of the Shigemi glass is much closer to that of water than air, shimming on a water-filled sample tube inserted to the same depth into the probe can be used to remove this B_0 gradient, to first order.




Figure 1. Measurement of magnetic susceptibility in a Shigemi sample tube. (a) Prior to measurements on this sample, the magnetic field homogeneity was optimized for a regular 5-mm NMR sample tube, filled to the same height of 70 mm above the bottom of the tube, and inserted to the same depth into the probehead. Dimensions in mm. A thin slice (pink) through the sample at height h above its bottom was selected by excitation sculpting. (b) HDO resonances for the Shigemi sample tube containing 0.3% H₂O/99.7% D₂O (no plunger). The glass/liquid interface of the Shigemi cell was positioned 5 mm below the center of the receiver coil, using the shim settings of a regular 5-mm sample tube filled to the same total height with the same solution, inserted to the same depth into the NMR probe (see left half of panel a). Overlaid are spectra recorded for the entire sample (green) and for 0.6-mm thickness horizontal slices through the Shigemi cell, centered at $h = 0$ mm (blue), and at $h = 13$ mm (red) by using excitation sculpting while applying z gradients. Spectra were scaled to show the same intensity.

2.2 Measurement of magnetic susceptibility in a high-resolution NMR magnet

Careful measurements of solvent susceptibility in a high-resolution magnet that used the gas phase of TMS as an internal reference have been reported by Hoffman (Hoffman, 2020;Hoffman, 2022). Here, we describe a different approach that also permits susceptibility measurements of solid objects. First, we demonstrate the method for measurement of the susceptibility of the glass bottom section of a Shigemi cell.

Bruker's topshim program was used to minimize magnetic field inhomogeneity of a regular 5-mm sample that contained a 1-mL solution of a variable NaCl concentration in 97% D₂O/3% H₂O, that was inserted into the probe to have its bottom ~13 mm below the center of the receiver coil. Then, without changing the shim settings, the HDO line shape on the same solvent composition was observed for a Shigemi sample tube, inserted to the same depth into the probehead. Because the Shigemi tube had an 8-mm bottom segment of solid glass, the flat solvent-glass interface then is located 5 mm below the center of the receiver coil (Figure 1a). If χ_g were identical to χ_s , the same perfect line shape would be expected, but with approximately 15% lower intensity because the bottom ~15% of the receiver coil now was filled with solid glass from the Shigemi tube.

For the 97% D₂O/3% H₂O Shigemi tube sample, a pronounced upfield shoulder was observed (Figure 1b, green), indicative of susceptibility mismatching. Excitation sculpting (Stott et al., 1995) (Appendix A1) while applying a pulsed z-gradient was then used to select a slice of 0.6-mm thickness, centered at the glass-solvent interface, i.e. selecting a ~0.3-mm solvent layer just above the interface. A resonance for this layer was observed that was 0.032 ppm upfield from the most intense segment of the solvent obtained with a 30° pulse, without slice selection (Figure 1b). When selecting a slice at a height of 13 mm above the interface, the signal (Figure 1b, red) coincided with the maximum of the resonance obtained without slice selection. Therefore, the difference in frequency between the red and blue resonances provides a good measure for the difference between $(\chi_g + \chi_s)/2$ and χ_s . The precision at which the frequency at the glass-water interface can be measured is impacted by the slice thickness of the selected solvent layer, which is subject to lateral gradients that increase with slice layer thickness and towards the edges of the slice. By varying the position of the center of the slice from ~0.2 mm below the interface to ~0.2 mm above the interface, increased total intensity with a strong downfield shoulder is

197 observed. The upfield edge of this line shape (Figure 1b, blue) remains invariant to the
198 precise position of the selected slice and represents the true $(\chi_g + \chi_s)/2$ value, which can be
199 determined at a precision of ~ 0.02 ppm.

200 Repeating the same measurement but using 97% H₂O/3% D₂O and strongly mistuning the
201 probehead to reduce radiation damping, showed a shoulder that was 0.0183 ppm closer to
202 the frequency observed 8 mm above the center of the coil (i.e. $h = 13$ mm; Appendix A2).
203 Accounting for the solutions not being fully deuterated or protonated, their difference in
204 static magnetic susceptibility then equals $\chi_{\text{H}_2\text{O}} - \chi_{\text{D}_2\text{O}} = 2 \times 0.0183 \times (100/94) = 0.04$ ppm. The
205 widely used literature value for $\chi_{\text{H}_2\text{O}}$ is -9.05 ppm (Sangal et al., 2023), yielding $\chi_{\text{D}_2\text{O}} = -$
206 9.01 ppm, in fair agreement with Hoffman's measurements (Hoffman, 2022). Using $\chi_{\text{D}_2\text{O}}$
207 = -9.01 ppm as a reference, the susceptibility of the Shigemi glass used in our
208 measurements is $\chi_g = -9.08$ ppm.




Figure 2. Effect of salt on magnetic susceptibility of 97% D₂O. Resonances shown correspond to a *ca* 0.6-mm thick slice through a Shigemi tube, centered at $h = 0$ (blue) and $h = 13$ mm (red) above the glass bottom of the Shigemi sample cell, using the same protocol as for Figure 1b at (a) 0 mM NaCl; (b) 1.0 M NaCl; (c) 2.0 M NaCl. (d) Magnetic susceptibility of D₂O as a function of NaCl concentration, using $\chi_{D_2O} = -9.01$ ppm as a reference. The intensities of the slices at $h = 0$ are upscaled *ca* 16-fold to account for the 50% smaller aqueous volume and the strong magnetic field inhomogeneity in these lower slices.

2.3 Salt dependence of water magnetic susceptibility

The effect of dissolved NaCl on magnetic susceptibility of water is important to protein NMR. We therefore repeated the above measurements of the 97% D₂O/3% H₂O Shigemi tube sample after addition of 0.5, 1, 1.5, and 2 M analytical grade NaCl (Sigma-Aldrich) to the solvent (Figure 2a-c), showing a linear increase in solvent susceptibility with salt


concentration over this range: $\chi_{\text{D}_2\text{O}+\text{NaCl}} = ([\text{NaCl}] \times 0.2 - 9.01)$ ppm, where $[\text{NaCl}]$ denotes the concentration molarity in units of moles per liter (Figure 2d).

2.4 Susceptibility and paramagnetic doping of 3D printer resin

Clear V4 resin was used because it enabled the highest precision (25- μm resolution) of printing on a Formlabs Form3+ 3D printer available in our laboratory, and its optical transparency facilitated sample handling for 3D printed microcells. The susceptibility of the printed Clear V4 resin was measured in the same manner as described above for the Shigemi tube. A cylindrical plug of 2 cm length was printed and pushed to the bottom of a standard 5-mm NMR sample tube that was prefilled with 0.7 mL 1% $\text{H}_2\text{O}/99\%$ D_2O , such that the top of the plug was again 5 mm below the center of the receiver coil once inserted into the magnet. Shimming of the magnetic field was carried out on a sample without the plug, filled to the same level by using ~ 1 mL of the same solvent, inserted into the probehead at the same depth, i.e. with the bottom of the sample tube at 25 mm below the center of the RF coil.

The frequency difference observed at the $\text{D}_2\text{O}/\text{resin}$ interface versus the top of the coil was -0.195 ppm (Figure 3a), corresponding $\chi_{\text{ClearV4}} = -9.40$ ppm, which is close to the value of -9.33 ppm measured for this resin by magnetic resonance imaging (Sangal et al., 2023). Considering that a mismatch in magnetic susceptibility between the sample cell and the solution impacts the achievable B_0 homogeneity, and that $\chi_{\text{ClearV4}} < \chi_{\text{H}_2\text{O}}, \chi_{\text{D}_2\text{O}}$, increasing the value of χ_{ClearV4} by paramagnetic doping of the resin in principle allows elimination of this difference. Finding a paramagnetic doping substance that is miscible with the printer resin and does not impact the performance of the 3D printer proved challenging. For example, the use of concentrated CuCl_2 in methanol strongly impacted the polymerization kinetics. The same problem was encountered for a range of strong paramagnetic chelated substances, such as gadodiamide (Omniscan), which is commonly used in protein paramagnetic relaxation enhancement measurements and as a contrast agent in magnetic resonance imaging. However, the hydrophobic cobalt(II) complex, cobalt(II) bis(2-ethylhexanoate), available from Sigma-Aldrich as a 65 wt. % solution in mineral spirits

(product number: 444545), proved miscible with the Clear V4 printer resin without major adverse impact on print quality.



252

Figure 3. Effect of cobalt(II)2-ethylhexanoate doping on the magnetic susceptibility of printed Formlabs Clear V4 resin. A solid plug of printed resin that is 2 cm in length and has a 4.0-mm outer diameter (OD) was inserted into a regular 5mm OD NMR sample tube (Wilmad-507; ID 4.2 mm), prefilled with 0.7 mL 99%D₂O/1%H₂O. Overlaid spectra are shown from 0.6-mm thickness slices centered at the top of a Clear V4 plug (blue), and 13 mm above the plug (red). (a) No doping; (b) 2 mM cobalt(II) doping; (c) 5 mM cobalt(II) doping.

Comparison of the difference in HDO resonance frequencies obtained for slices at 13 mm above the interface between solvent and printed plug and at the interface for three different levels of the cobalt(II) doping, 0 mM (Figure 3a); 2 mM (Figure 3b) and 5 mM (Figure 3c), shows a doping-dependent decrease from +117 Hz to -80 Hz for the NaCl-free 99% D₂O sample. This result indicates that it is possible to match the susceptibility of the solvent to that of the printed resin. However, that would require a large number of printed cells to

cover the ionic strength range from 0 to 2 M salt, while also accounting for the difference between D₂O and H₂O samples.

In practice, printing sample cells with different levels of cobalt(II) doping is labor-intensive because it requires thorough cleaning of the printing vat used by the Formlabs 3D laser printer. Printing with doped resin requires mixing of the viscous resin at the molecular level with the doping agent and keeping it homogenous during printing. The latter also required some modification of the printer to prevent refreshing of the printer resin with undoped resin from a sealed cassette during printing. We therefore resorted to printing the sample cells with an ellipsoidal shape, that to a good approximation were insensitive to the susceptibility mismatch between solvent and printed resin (Schenck, 1996; VanderHart, 1996).

2.4 Performance of a Clear V4 ellipsoidal microcell

For applications to proteins, we settled on an ellipsoidal microcell design with a volume of 130 μ L (Figure 4a). A printed access channel of 1.6 mm diameter (measured 1.3 mm) and 8.5-mm length and a volume of \sim 11 μ L was used for cleaning of the sample after the initial print and prior to further hardening of the resin in a Formlabs light chamber (see Methods). A subsequent overnight rinse with Milli-Q H₂O at 60 $^{\circ}$ C was used to remove small water-soluble contaminants. Comparison of the NMR spectrum of these impurities with those in a sample obtained by briefly vortexing a mixture of unpolymerized resin and water followed by entering the aqueous phase into an NMR sample tube (Figure A3) suggests that these impurities consist of residual, methacrylate-based small oligomeric species that apparently can diffuse from the cell walls into the aqueous contents of the cell. Leaving the sample cells filled with H₂O for a week prior to a final rinse reduces the impurity levels further, but in our experience is not necessary considering we have not observed any interaction between these very low impurity concentrations and isotopically enriched proteins. For applications to samples in D₂O, leaving the microcell filled with D₂O for 24 h, prior to using it, reduces the intensity of a weak, very broad (\sim 1 kHz) signal at \sim 3.8 ppm, that results from H₂O diffusing into the resin.

Because the digital printer increases the size of printed parts by a small amount due to partial polymerization adjacent to the laser-selected spots, the printed walls of the chamber are actually slightly thicker than designed, and the total volume of the cell including its access channel was measured gravimetrically to be 130 μL (Figure 4b).




Figure 4. Images of the printed microcell. (a) Technical drawing; note that the actual dimensions of printed material are slightly larger due to polymerization of a thin (~ 0.15 mm) sticky surface layer that remains on the printed cell prior to subsequent UV curing. (b) Photographs of the printed cells (left) prior to and (right) after filling with a blue dye solution; (c) Contour plot of a sagittal xz cross-section through a 3D image of a 97% D_2O , 3% H_2O sample containing 1.8 mM CuCl_2 to shorten the $^1\text{H}T_1$ value to ~ 0.7 s. The image was recorded in absorption mode on an 800-MHz NMR spectrometer equipped with a 3-axis pulsed field gradient probehead. The total measurement time was 9 minutes.

A sagittal (xz) cross-section through the absorption mode 3D image of the cell, recorded on a Bruker Neo 800-MHz instrument equipped with a 3-axis pulsed field gradient probehead, yielded a shape that matched the ellipsoidal design (Figure 4c) but that did not

include the access channel because the solvent in that channel falls outside the RF coil. Small distortions near the bottom of the image correspond to the drop off in RF coil receptivity. Distortions at the top of the sample originate from the access channel which also distorts the ellipse. However, cross-sections taken through the 3D image along the x and z axis through the center of the sample, plotted along the sides of the 3D image, show the expected nearly rectangular shape, indicative of linear imaging gradients.

^1H non-spinning linewidths of ~ 1 Hz at half height (600 MHz) obtained with the cell (expanded anomeric doublet in Fig.5) were slightly larger than those obtained for a regular 5-mm NMR sample in the same probehead. With a width of only ~ 13 Hz at 0.55% of the HDO peak height the lineshape is also very good, which is most important for protein NMR studies, where protein ^1H line widths at half height commonly exceed 10-20 Hz due to fast transverse relaxation and the absence of a “hump” in the water line shape is critical for good solvent suppression.




Figure 5. ^1H background of printed microcell. The regular ^1H NMR spectrum of a 1 mg/mL solution of sucrose in 99% D_2O , recorded without presaturation (red) and with ^1H presaturation at -2 ppm (blue), using a 100 Hz RF field strength to suppress the ^1H background of the solid resin. Resonances from small impurities released from the printed cell when it has only briefly (\sim ca 2 h) been rinsed with H_2O are marked with an asterisk. However, a subset of these resonances very slowly reappear over a period of multiple days, at concentrations much lower than seen in the

figure (see Figure A3). For display purposes, the not-presaturated spectrum has been shifted upfield by 2 ppm and offset vertically. The inset shows the splitting for the sucrose anomeric doublet at 5.39 ppm.

We note that sometimes microscopic air bubbles can form inside the cell after filling it with the NMR sample solution. When this happens, it has a very strong adverse impact on both lineshape and linewidth. To eliminate the potential presence of such air bubbles, we briefly (~20 sec) spin the filled sample cell in a speedvac, operating at a pressure of *ca* 150 Pa, and replenish any lost volume by adding a few μL of the protein solution to the access channel.

For observation of simple one-dimensional ^1H spectra, without echo delays, the protons of the resin yield a strong background signal that is broad due to the rapid transverse relaxation of this solid material. This background signal is readily suppressed by spin echo delays prior to the start of signal acquisition, as are already present in nearly all protein NMR experiments. It can also be effectively reduced by saturating this broad background with a weak RF field outside the spectral region of interest. Saturation with a 100-Hz RF field, applied at -2 ppm in the ^1H spectrum, attenuates the background signal by about six-fold (Figure 5).

The ^1H RF field homogeneity was compared for the microcell and for a Shigemi sample (straight wall) containing 280 μL of phosphate buffered saline (PBS) solution in 99% D_2O . Due to the substantial detuning of the 600 MHz cryoprobe used for this work by the ionic solution in the Shigemi sample, the RF power for this sample was increased by 2.7 dB over the power used for the microcell to yield the same 90° pulse width of $\sim 17 \mu\text{s}$. Comparison of the decay of the signal when the excitation pulse is increased from 0.2 to 497.7 μs using the Bruker “paropt” module (Figure 6) shows slightly better RF field homogeneity for the smaller 3D printed microcell than for the Shigemi sample. Notably, the intensity of the signal after a 90° pulse was only $\sim 10\%$ lower for the printed microcell than for the Shigemi tube that contained more than double the volume of the same solution.





Figure 6. Comparison of ^1H RF field homogeneity in a 280- μL Shigemi sample cell (red) and in the 3D printed 130- μL 23 \times 3.3-mm ellipsoidal microcell (blue). Signal intensity is shown as a function of ^1H pulse duration, ranging from 0.2 to 497.7 μs , generated by the Bruker macro 'paropt'. Both samples contained the same solution of PBS buffer in 99% D_2O . The RF power for the microcell sample was adjusted 2.7 dB lower than for the Shigemi sample to equalize the 90° pulse lengths.

2.6 Observation of high-resolution protein NMR spectra

The high resolution and sensitivity obtained with the microcell are illustrated for two proteins, N-acetylated α -synuclein and native C-terminal amidated melittin in its monomeric and tetrameric forms. Complete or nearly complete N-acetylation of α -synuclein is invariably present in mammalian cells (Bartels et al., 2011) and strongly impacts its interaction with phospholipids (Kang et al., 2012; Maltsev et al., 2012). By simultaneously including a plasmid for expressing the NatB complex, needed for acetylation of α -synuclein, together with a plasmid for α -synuclein, fully N-acetylated protein can also be obtained from bacterial expression systems (Johnson et al., 2010). Although this combined expression reduced protein yields in our hands, it enabled the pivotal incorporation of stable isotopes, such as ^{15}N , in the biologically relevant state of the protein.

As can be seen, in *ca* 20 min, a high-quality gradient-enhanced ^1H - ^{15}N HSQC spectrum was obtained for 130 μg of N-acetylated α -synuclein (14.5 kDa) in PBS buffer, pH 6.5, when using the microcell, which approaches the sensitivity and resolution that was

380 obtained for 500 μg protein in a regular 5-mm sample cell, using a 0.5 mL sample volume
 381 (Figure 7).



382

383 **Figure 7.** Comparison of small regions of the 600-MHz ^1H - ^{15}N HSQC spectra of 70 μM N-acetylated
 384 α -synuclein at 20 $^\circ\text{C}$ in PBS-buffer, pH 6.5. Each spectrum results from $175^* \times 4000^*$ data points
 385 with two transients per FID, for total measuring times of 21 min each. (a) 130 μL in the 3D printed
 386 microcell; (b) 500 μL in a standard 5-mm NMR tube. Lowest contours are drawn at the same level
 387 above the respective RMS noise. Insets show the expansion of the L113 cross peak, with cross-
 388 sections through L113 above the panels at locations marked by the red dotted lines. The
 389 comparison shows slightly lower resolution of the $^1\text{H}^{\text{N}}$ - $^1\text{H}^{\alpha}$ doublets for the microcell, reflected in
 390 12% lower S/N versus 6.5% lower peak integrals relative to RMS noise.

391 The utility of the microcell becomes even more compelling for the study of the
 392 tetramerization of native melittin, which requires C-terminal amidation. The latter involves
 393 a large number of chemical derivatization and purification steps (Gelenter and Bax, 2023),
 394 making it very challenging to generate adequate quantities of peptide for driving it to its
 395 tetrameric state within a standard 5-mm NMR tube. Tetramerization is promoted by
 396 increasing the NaCl concentration, but the tuning of the RF circuitry in cryoprobes often
 397 limits salt concentrations to be below *ca* 0.5 M. Here, we demonstrate that the microcell
 398 enables observation of the monomer-tetramer equilibrium at salt concentrations of 2 M,
 399 and that indeed even at a moderate peptide concentration of 250 μM the peptide is fully
 400 tetrameric in the presence of 2 M NaCl (Figure 8). Interestingly, residues L9-R24, close
 401 to the center of the peptide, are attenuated by an exchange process that is much less

prevalent at low ionic strength and high peptide concentration (Gelenter et al., 2024). We speculate that the line-broadening associated with these weaker resonances arises from the exchange process between the asymmetric arrangement of the two dimers seen in its X-ray tetrameric structure (Terwilliger and Eisenberg, 1982) switching from fast exchange at lower ionic strength to intermediate exchange at 2 M NaCl.

Notably, the same microcell was used for the two melittin spectra. After recording of the low-ionic strength spectrum, contents of the microcell was removed and used to dissolve 15.2 mg NaCl, prior to insertion of the high ionic strength sample into the original cell, thereby demonstrating the recyclability of the tube.




Figure 8. 800-MHz ^1H - ^{15}N HSQC spectra from 90 μg (250 μM) of ^{15}N -labeled native melittin in a 130- μL microcell. Both spectra were collected at 288 K in 10 mM sodium phosphate buffer, pH 7.0, containing 3% D_2O . (a) Sample containing no NaCl, where melittin remains an intrinsically disordered monomer. L13_{cis} and A15_{cis} correspond to residues in monomers with P14 in the *cis* conformation. The data was collected with 2 transients per FID, using 200* points in the indirect dimension corresponding to an evolution of 103 ms and a total measurement time of 27 min. (b) 15.2 mg of NaCl was added to the sample from (a) to reach a final concentration of 2 M NaCl. Under these conditions melittin adopts an α -helical tetrameric conformation. The data was collected with 8 transients, using 100* points in the indirect dimension ($t_{1,\text{max}} = 51$ ms) and a total experimental time of 52 min.

3 Concluding remarks

3D printing enables the efficient and relatively inexpensive creation of complex, customized products with minimal waste. It is extensively used for prototyping new designs of items with intricate or complex geometries and has broad impact in science and engineering. 3D printing also enabled the design and construction of high performance solid-state NMR probes, offering similar or improved filling factors due to the coil being in close proximity to the sample, resulting in high RF transmit and receive efficiencies (Long et al., 2021;Pereira et al., 2023).

In solution NMR spectroscopy, 3D-printed bioreactor platforms were introduced that are compatible with low-field NMR spectrometers that accommodate bioengineered 3D cell models (Mangas-Florencio et al., 2025). That work consisted of a bioreactor made of biocompatible materials and included a microfluidic system for optimization of cell culture conditions during the actual NMR data collection process.

The application of 3D printing to high-field solution NMR spectroscopy has remained rather limited, largely due to the requirements of high magnetic field homogeneity and minimal background signals. Our study demonstrates that the homogeneity requirement can be met by printing small sample cells with an ellipsoidal shape. The half-height line widths achievable for our microcell is *ca* 1 Hz and remains limited by the precision at which the cell's surface can be printed. For shimming purposes, we first used Bruker's topshim program to adjust field homogeneity to starting values on a regular solution NMR sample filled to the same height (~40 mm) as the length of the microcell. Subsequently, after entering the microcell into the magnet, we used topshim followed by iterative tuning of z , z^2 , x , y , xz and yz gradients. The microcells can easily be recycled as they slide in and out of standard high-quality Wilmad-507, New Era NE-HP5, or Norell Standard Series 5-mm NMR tubes. The access channel of the microcell is sufficiently small that surface tension prevents the aqueous solution from leaving the microcell when the 5-mm NMR tube is fully inverted while the microcell slides out of it.

Although inexpensive, sample cells are easily recycled as highlighted for the melittin sample, where 15.2 mg of salt was added to the initial sample by first removing the sample

451 solution, dissolving the NaCl, and re-injecting the solution into the original cell, all with
452 minimal losses. It is advisable to briefly spin and expose the sample to vacuum to remove
453 dissolved gasses, in particular for lengthy experiments. Formation of even microscopic air
454 bubbles deteriorates homogeneity for the microcell sample more than for the Shigemi or
455 larger conventional NMR sample tubes.

456 The sample cell appears unsuitable for the use of organic solvents which dissolve and
457 release resin components, resulting in strong narrow background signals. Even when using
458 water as the solvent, slowly increasing signals from micromolar quantities of resin-derived
459 small molecules appear in the ^1H spectrum over a period of days (marked by an asterisk in
460 Figure 5). However, the standard use of isotope-enriched multi-dimensional multinuclear
461 NMR experiments keeps these resonances well below the signal-to-noise threshold level,
462 and unless such contaminants have a strong interaction with the protein studied they have
463 no effect on the acquired spectra. We note that commercial 3D printing resins are
464 formulated in a way that tries to meet various performance specifications, such as material
465 strength, printing resolution, and printing speed. As such, they may not be optimal for high
466 precision work such as required for protein NMR. Future development of resins is needed
467 to allow high precision printing while minimizing release of small, incompletely
468 polymerized precursors into the aqueous phase.

469 While a small amount of solvent ($<0.1\%$) can diffuse into cavities of the polymerized resin,
470 no detectable loss of solute signal was observed. For example, the intensity of the sucrose
471 NMR resonances remained unchanged, to within $\pm 0.2\%$ over a duration of more than one
472 week.

473 We cleaned the microcells by soaking them overnight at $60\text{ }^\circ\text{C}$ in Milli-Q water and
474 subsequent removal of most of the solvent with a standard gel micropipette tip that didn't
475 quite reach the bottom of the cell, followed by upside down centrifugation after insertion
476 into an Eppendorf tube to remove the remainder of the solvent. An additional rinse with
477 $130\text{ }\mu\text{L}$ D_2O , followed by centrifugation and vacuum exposure can be used to remove any
478 residual solvent protons if the sample is intended for measurements in highly deuterated
479 D_2O . These are the most labor-intensive steps in preparing the sample cells, but limited
480 quantities of unrinsed sample cells are available upon request.

4. Methods

4.1 Magnetic susceptibility measurements

Selective excitation while applying a 11.5 G/cm (20% on the Bruker Neo-600 instrument) was used for collecting the HDO solvent resonance measurements (Appendix A1) of 0.6-mm thickness slices at various heights, h , above the flat interface between a solid printed plug and the aqueous solvent. The printed plug with outer diameter of 4-mm and a length of 20-mm, including its hemispheric bottom, was inserted and pushed to the bottom of a regular NMR tube, prefilled with 0.7 mL of 97% D₂O or 97% H₂O solution. A similar tube without the plug was filled to the same height with the same solution and used for shimming the magnetic field using topshim prior to inserting the sample with the plug at its bottom, where the same shim settings of the sample without the plug were used. The thickness of the aqueous fraction of the slice collected for $h=0$ was two-fold smaller than for slices at $h>0.3$ mm, with a correspondingly lower integrated volume. The frequency of slices collected at heights $>\sim 10$ mm above the plug became essentially independent of h . The slice collected at $h=0$ shows extensive line broadening due to the large field gradient at the solvent-plug interface (Figure 2a-c). The difference in ppm frequency was used as a measure for $(\chi_{\text{solvent}} - \chi_{\text{resin}})/2$.

4.2 Printing of the microcell

Cells were printed in Clear V4 resin on a Form 3+ printer at 25-micron resolution to achieve a smooth surface finish. The designs were created in OpenSCAD (2021.01) using standard STL export settings, then prepared for printing in PreForm (3.43.2). Models were oriented so that the bottom of each cell faced downward on the build platform, with 0.2 mm touchpoint supports attached only at the bottom of the cells. The cells were washed with 15 mL isopropyl alcohol (IPA) at a rate of 5 mL/min through a syringe needle (0.58 mm ID) with a blunt tip, connected to a Fast Protein Liquid Chromatography (FPLC) pump. Cells were kept vertical during printing and subsequent UV-curing using Form Cure for 16 hours at 60 °C. Subsequently, cells were filled with Milli-Q water and immersed in a water-

filled falcon tube that was heated at 60 °C for 12 hours to remove water-soluble components. After this cleaning, to further minimize the amounts of resin-derived impurities from leaching from the cell wall into the solvent, which occurs at a rate that decreases steadily with time, cells can be stored filled with and immersed in water for a week at room temperature. After removal of the water from the cells by pipetting followed by centrifugation upside down in Eppendorf tubes, they were briefly dried under vacuum. Subsequent use of such a cell showed strongly reduced intensities of these impurity signals (Figure A3).

The paramagnetically doped resin was prepared by mixing Formlabs Clear V4 Resin with a cobalt(II) bis(2-ethylhexanoate) solution (65 wt.% in mineral spirits, Sigma-Aldrich, 444545) on a shaker at 37 °C and 200 rpm for 30 minutes. The plugs printed from doped resin were washed with IPA and cured at 60 °C for 16 hours using Form Cure.

4.3 NMR sample preparation

The microcells were filled to the top with the sample solution (~130 µL) using a gel-tip pipette and degassed at *ca* 150 Pa pressure for 20-30 seconds using a SpeedVac (Savant, SVC-100-H). After degassing, the cells were topped off with *ca* 2 µL of additional sample solution and inserted into standard Wilmad-507, New Era NE-HP5, or Norell Standard Series 5-mm NMR tubes.

4.4 Imaging of the microcell

Although MRI normally uses absolute value mode displays, higher resolution absorption mode spectra can also be obtained (Bretthorst, 2008). For generating images of the microcell, we used a very simple one-pulse sequence with variable durations of the x, y, and z gradients for encoding the three spatial dimensions (Appendix A4). The experiment used Rance-Kay quadrature selection (Palmer et al., 1991; Kay et al., 1992) in both the x and y dimensions, by collecting four scans per hypercomplex time domain data point (ns=1). NMRPipe (Delaglio et al., 1995) processing of the 3D time domain matrix was used to generate regular, amplitude-modulated quadrature (States et al., 1982) in both indirect dimensions (see Appendix A5 for NMRPipe processing script). A total of 20*(x)×

538 $20^*(y) \times 512^*(z)$ data points were collected with a total acquisition time of *ca* 9 minutes
539 for a sample that contained 1.8 mM CuCl₂ in 99% D₂O.

540

Appendix A: Supplementary data and code

A1: Pulse diagram used for slice selection




Figure A1. Diagram of the slice selection pulse sequence used to measure magnetic susceptibility. The narrow bar represents the 90° ^1H excitation pulse. The shaped pulses ϕ_1 and ϕ_2 have a profile of Reburp (Geen and Freeman, 1991) with a duration of 2 ms and a variable offset frequency, ranging from *ca* +23 to -30 kHz from the water signal resonating at 4.7 ppm. The gradient pulses are either sine-bell shaped or rectangular, with durations of 0.5, 2.1 and 0.5 ms and a strength of 2, -11.5 and 3.5 Gauss/cm for G_1 , G_2 and G_3 , respectively. Delays: $\delta = 50 \mu\text{s}$ and $\tau = 60 \mu\text{s}$. The following phase cycling scheme was used: $\phi_1 = y, y, y, y, -x, -x, -x, -x, -y, -y, -y, -y, x, x, x, x$; $\phi_2 = x, y, -x, -y$; $\phi_{\text{rec}} = x, -x, x, -x, -x, x, -x, x$. The interscan delay was set to 4 s, number of scans to 4, spectral width to 8620 Hz and the acquisition time to 2 s.

A2: Spectra used for measuring $\chi_{\text{glass}} - \chi_{\text{H}_2\text{O}}$




Figure A2. Same as Figure 1b (main text), but for a solution containing 97% H_2O /3% D_2O . The resonance of the full Shigemi tube sample (green) is broadened by radiation damping and shown together with slices taken at $h = 0$ mm (blue), and at $h = 13$ mm (red). The center of the green resonance is shifted relative to its regular position due to radiation damping effect in the deliberately mistuned probe (Torchia, 2009).

A3: Spectra of polymer precursors and oligomeric impurities




Figure A3. ^1H NMR spectra (500 MHz) of liquid Formlabs Clear V4 resin, saturated in D_2O (5 mm tube, 128 scans; blue trace), and 1 mg/mL sucrose in D_2O recorded in a 3D-printed sample cell, 48 h after preparation (512 scans; green trace). The blue spectrum is scaled down 10-fold relative to the green spectrum and corresponds to *ca* 10 mg/mL precursor concentration. The dissolved oligomer impurity concentration in the 3D-printed microcell sample corresponds to *ca* 24 $\mu\text{g mL}^{-1}$. Insets correspond to the olefinic and aliphatic regions, upscaled 25-fold. According to the Formlabs Clear V4 datasheet, the monomeric precursors in the blue spectrum correspond to methacrylic acid monoester with propane-1,2-diol and 7,7,9(or 7,9,9)-trimethyl-4,13-dioxo-3,14-dioxo-5,12-diazahexadecane-1,16-diylbismethacrylate (see <https://formlabs-media.formlabs.com/datasheets/1801037-SDS-ENEU-0.pdf>).

576 A4: Bruker pulse program used for imaging the microcell:

```
577 #include<Avance.incl>
578 #include<Grad.incl>
579 #include<De.incl>
580
581
582 "d11=30m"
583
584 "in0=inf1".      ;125u
585 "in10=inf2"      ;125u
586
587 "d0=0"
588 "d10=0"
589
590 define list<gradient> EA2 = { 1.000 -1.000}
591
592 aqseq 321
593
594
595 1 ze
596   1m
597 2 d11 groff
598   10u BLKGRAD
599   10u pl9:f1      ;set to 1000dB unless presat is needed
600   10u fq=cnst2(bf ppm):f1 ;chemical shift for the resin to be
601 presat'ed
602   d1 cw:f1
603   10u do:f1
604   10u UNBLKGRAD
605   p20:gp20.      ;3m at 33% of z-grad
606   1m fq=0:f1      ;shift carrier back to be on-resonance with water
607   1m pl1:f1
608   (p1 ph1):f1     ;20 degree flip angle
609   1u gron1*EA      ;x-grad 8%
610   d0
611   1u groff
612
613   1u gron2*EA2     ;y-grad 8%
614   d10
615   1u groff
616   600u            ;delay to dephase a very broad hump
617   1u gron3         ;z-grad -1%
618   go=2 ph31
619   d11 groff mc #0 to 2
620   F1EA(calgrad(EA), caldel(d0))      ;TD1=40
621   F2EA(calgrad(EA2), caldel(d10))    ;TD2=40
622   d11 BLKGRAD
623 exit
624
625
626 ph1=0
627 ph31=0
```

628 A5: NMRPipe conversion and processing scripts

```
629 #!/bin/csh
630
631 bruk2pipe -verb -in ./ser \
632   -bad 0.0 -ext -aswap -DMX -decim 1240 -dspfv 20 -grpdly 68 -ws 8 -
633 noi2f \
634   -xN          1024  -yN          40  -zN          40
635 \
636   -xT          512  -yT          20  -zT          20
637 \
638   -xMODE       DQD  -yMODE  Echo-AntiEcho  -zMODE  Echo-AntiEcho
639 \
640   -xSW         16129.032  -ySW         8000.000  -zSW         8000.000
641 \
642   -xOBS        800.134  -yOBS        800.134  -zOBS        800.134
643 \
644   -xCAR        4.821  -yCAR        4.821  -zCAR        4.821
645 \
646   -xLAB        Hz  -yLAB        Hy  -zLAB        Hx
647 \
648   -ndim        3  -aq2D        Complex
649 \
650 | pipe2xyz -x -out ./fid/test%03d.fid -ov
651
652
653 xyz2pipe -in fid/test%03d.fid -x \
654 | nmrPipe -fn SP -off 0.5 -end 0.98 -pow 2 -c 0.5 \
655 | nmrPipe -fn ZF -zf 2 -auto \
656 | nmrPipe -fn FT \
657 | nmrPipe -fn PS -p0 149 -p1 38 -di \
658 | nmrPipe -fn TP \
659 | nmrPipe -fn SP -off 0.5 -end 0.98 -pow 1 -c 0.5 \
660 | nmrPipe -fn ZF -zf 2 -auto \
661 | nmrPipe -fn FT \
662 | nmrPipe -fn PS -p0 135 -p1 0 -di \
663 | nmrPipe -fn ZTP \
664 | nmrPipe -fn SP -off 0.5 -end 0.98 -pow 1 -c 0.5 \
665 | nmrPipe -fn ZF -zf 2 -auto \
666 | nmrPipe -fn FT \
667 | nmrPipe -fn PS -p0 90 -p1 0 -di \
668 | nmrPipe -fn TP \
669 | nmrPipe -fn POLY -ord 4 -nw 300 -nl 301 1740 \
670 | pipe2xyz -out ft/test%04d.ft3 -x -ov -verb
671
```

672 Author contributions

673 TK: conceptualization (equal), 3D printing and cleaning, figures 1-6, review and editing
674 (equal). MDG: review and editing (equal), data collection (equal), figure 8. JY: review and
675 editing (equal), data collection (equal), figure 7. AB: conceptualization (equal), writing
676 original draft, review and editing (equal), recording data (equal).

677 Acknowledgments

678 This work was supported by the Intramural Research Program of the NIDDK at the
679 National Institutes of Health. We thank James L. Baber for assistance with
680 instrumentation, and Philip Anfinrud, Marshall J. Smith, Nico Tjandra and David A.
681 Nyenhuis for useful suggestions.

682 Competing interests

683 AB is a member of the editorial board of *Magnetic Resonance*. The peer-review process
684 was guided by an independent editor, and the authors also have no other competing
685 interests to declare.

686 Financial support

687 This research was supported by grant DK029046 from the Intramural Research Program
688 of the National Institute of Diabetes and Digestive and Kidney Diseases.

689

690

691 References

692
693

694 Barbara, T. M.: Cylindrical Demagnetization Fields and Microprobe Design in High-
695 Resolution NMR, *Journal of Magnetic Resonance Series A*, 109, 265-269,
696 <https://doi.org/10.1006/jmra.1994.1168>, 1994.

697 Barbara, T. M.: NMR Probes for Small Sample Volumes, *Encyclopedia of Magnetic*
698 *Resonance*, eds-in-chief R. K. Harris and R. Wasylishen, John Wiley: Chichester.,
699 <https://doi.org/10.1002/9780470034590.emrstm1084>, 2009.

- 700 Bartels, T., Choi, J. G., and Selkoe, D. J.: Alpha-Synuclein Occurs Physiologically as a
701 Helically Folded Tetramer that Resists Aggregation, *Nature*, 477, 107-110,
702 <https://doi.org/10.1038/nature10324>, 2011.
- 703 Bretthorst, G. L.: Automatic Phasing of MR Images. Part II: Voxel-wise Phase Estimation,
704 *J. Magn. Reson.*, 191, 193-201, <https://doi.org/10.1016/j.jmr.2007.12.011>, 2008.
- 705 Delaglio, F., Grzesiek, S., Vuister, G. W., Zhu, G., Pfeifer, J., and Bax, A.: NMRPipe: a
706 Multidimensional Spectral Processing System Based on UNIX Pipes., *J. Biomol. NMR*, 6,
707 277-293, <https://doi.org/10.1007/BF00197809>, 1995.
- 708 Evans, D. F.: The Determination of the Paramagnetic Susceptibility of Substances in
709 Solution by Nuclear Magnetic Resonance, *J. Chem. Soc.*, 2003-2005,
710 <https://doi.org/10.1039/jr9590002003>, 1959.
- 711 Freeman, R., Kempell, S. P., and Levitt, M. H.: Radiofrequency Pulse Sequences which
712 Compensate their Own Imperfections, *J. Magn. Reson.*, 38, 453-479,
713 <https://doi.org/10.1016/j.jmr.2011.08.03>, 1980.
- 714 Geen, H., and Freeman, R.: Band-selective Radiofrequency Pulses, *J. Magn. Reson.*, 93,
715 93-141, [https://doi.org/10.1016/0022-2364\(91\)90034-Q](https://doi.org/10.1016/0022-2364(91)90034-Q), 1991.
- 716 Gelenter, M., and Bax, A.: Recombinant Expression and Chemical Amidation of
717 Isotopically Labeled Native Melittin, *J. Am. Chem. Soc.*, 145, 3850-3854,
718 <https://doi.org/10.1021/jacs.2c12631>, 2023.
- 719 Gelenter, M. D., Yau, W. M., Anfinrud, P. A., and Bax, A.: From Milliseconds to Minutes:
720 Melittin Self-Assembly from Concerted Non-Equilibrium Pressure-Jump and Equilibrium
721 Relaxation Nuclear Magnetic Resonance, *J. Phys. Chem. Lett.*, 15, 1930-1935,
722 <https://doi.org/10.1021/acs.jpcllett.3c03563>, 2024.
- 723 Gutiérrez-Mejía, F., and Ruiz-Suárez, J. C.: AC Magnetic Susceptibility at Medium
724 Frequencies Suggests a Paramagnetic Behavior of Pure Water, *Journal of Magnetism and*
725 *Magnetic Materials*, 324, 1129-1132, <https://doi.org/10.1016/j.jmmm.2011.10.035>, 2012.
- 726 Hale, W., Rossetto, G., Greenhalgh, R., Finch, G., and Utz, M.: High-resolution Nuclear
727 Magnetic Resonance Spectroscopy in Microfluidic Droplets, *Lab on a Chip*, 18, 3018-
728 3024, <https://doi.org/10.1039/c8lc00712h>, 2018.
- 729 Hizawa, T., Takahashi, M., and Iwase, E.: Sample Shape Design for a Micro-volume NMR
730 Spectroscopy, *Micro & Nano Letters*, 12, 550-553, <https://doi.org/10.1049/mnl.2017.0138>,
731 2017.
- 732 Hoffman, R.: Magnetic Susceptibility Measurement by NMR: 2. The Magnetic
733 Susceptibility of NMR Solvents and their Chemical Shifts, *J. Magn. Reson.*, 335,
734 <https://doi.org/10.1016/j.jmr.2021.107105>, 2022.
- 735 Hoffman, R. E.: Magnetic Susceptibility Measurement by NMR: 1. The Temperature
736 Dependence of TMS, *J. Magn. Reson.*, 312, <https://doi.org/10.1016/j.jmr.2020.106689>,
737 2020.
- 738 Johnson, M., Coulton, A. T., Geeves, M. A., and Mulvihill, D. P.: Targeted Amino-
739 Terminal Acetylation of Recombinant Proteins in *E. coli*, *PLoS One*, 5, e15801,
740 <https://doi.org/10.1371/journal.pone.0015801>, 2010.

- 741 Kang, L., Moriarty, G. M., Woods, L. A., Ashcroft, A. E., Radford, S. E., and Baum, J.: N-
742 Terminal Acetylation of Alpha-Synuclein Induces Increased Transient Helical Propensity
743 and Decreased Aggregation Rates in the Intrinsically Disordered Monomer, *Protein Sci.*,
744 21, 911-917, <https://doi.org/10.1002/pro.2088>, 2012.
- 745 Kay, L. E., Keifer, P., and Saarinen, T.: Pure Absorption Gradient-Enhanced Heteronuclear
746 Single Quantum Correlation Spectroscopy with Improved Sensitivity, *J. Am. Chem. Soc.*,
747 114, 10663-10665, <https://doi.org/10.1021/ja00052a088>, 1992.
- 748 Long, Z., Ruthford, J., and Opella, S. J.: 3D Printed Sample Tubes for Solid-State NMR
749 Experiments, *J. Magn. Reson.*, 327, <https://doi.org/10.1016/j.jmr.2021.106957>, 2021.
- 750 Maltsev, A. S., Ying, J. F., and Bax, A.: Impact of N-Terminal Acetylation of α -Synuclein
751 on Its Random Coil and Lipid Binding Properties, *Biochemistry*, 51, 5004-5013,
752 <https://pubs.acs.org/doi/10.1021/bi300642h>, 2012.
- 753 Mangas-Florencio, L., Herrero-Gómez, A., Eills, J., Azagra, M., Batlló-Rius, M., and
754 Marco-Rius, I.: A DIY Bioreactor for in Situ Metabolic Tracking in 3D Cell Models via
755 Hyperpolarized ^{13}C NMR Spectroscopy, *Anal. Chem.*, 97, 1594-1602,
756 <https://doi.org/10.1021/acs.analchem.4c04183>, 2025.
- 757 Manu, V. S., Olivieri, C., and Veglia, G.: AI-designed NMR Spectroscopy RF pulses for
758 Fast Acquisition at High and Ultra-High Magnetic Fields, *Nat. Comm.*, 14,
759 <https://doi.org/10.1038/s41467-023-39581-4>, 2023.
- 760 Palmer, A. G., Cavanagh, J., Wright, P. E., and Rance, M.: Sensitivity Improvement in
761 Proton-Detected 2-Dimensional Heteronuclear Correlation Nmr-Spectroscopy, *J. Magn.*
762 *Reson.*, 93, 151-170, [https://doi.org/10.1016/0022-2364\(91\)90036-S](https://doi.org/10.1016/0022-2364(91)90036-S), 1991.
- 763 Pereira, D., Sardo, M., Marín-Montesinos, I., and Mafra, L.: One-Shot Resin 3D-Printed
764 Stators for Low-Cost Fabrication of Magic-Angle Spinning NMR Probeheads, *Anal.*
765 *Chem.*, 95, 10384-10389, <https://doi.org/10.1021/acs.analchem.3c01323>, 2023.
- 766 Roschmann, P.: Radiofrequency Penetration and Absorption in the Human Body -
767 Limitations to High-Field Whole-Body Nuclear Magnetic Resonance Imaging, *Med.*
768 *Phys.*, 14, 922-931, <https://doi.org/10.1118/1.595995>, 1987.
- 769 Ryan, H., Smith, A., and Utz, M.: Structural Shimming for High-Resolution Nuclear
770 Magnetic Resonance Spectroscopy in Lab-on-a-Chip Devices, *Lab on a Chip*, 14, 1678-
771 1685, <https://doi.org/10.1039/c3lc51431e>, 2014.
- 772 Sangal, M., Anikeeva, M., Priese, S. C., Mattern, H., Hövener, J. B., and Speck, O.: MR
773 based Magnetic Susceptibility Measurements of 3D Printing Materials at 3 Tesla, *J. Magn.*
774 *Reson. Open*, 16-17, <https://doi.org/10.1016/j.jmro.2023.100138>, 2023.
- 775 Schenck, J. F.: The Role of Magnetic Susceptibility in Magnetic Resonance Imaging: MRI
776 Magnetic Compatibility of the First and Second Kinds, *Med. Phys.*, 23, 815-850,
777 <https://doi.org/10.1118/1.597854>
778 , 1996.
- 779 States, D. J., Haberkorn, R. A., and Ruben, D. J.: A Two-dimensional Nuclear Overhauser
780 Experiment with Pure Absorption Phase in Four Quadrants, *J. Magn. Reson.*, 48, 286-292,
781 [https://doi.org/10.1016/0022-2364\(82\)90279-7](https://doi.org/10.1016/0022-2364(82)90279-7), 1982.

- 782 Stott, K., Stonehouse, J., Keeler, J., Hwang, T. L., and Shaka, A. J.: Excitation Sculpting
783 in High-Resolution Nuclear Magnetic Resonance Spectroscopy - Application to Selective
784 NOE Experiments, J. Am. Chem. Soc., 117, 4199-4200,
785 <https://doi.org/10.1021/ja00119a048>, 1995.
- 786 Terwilliger, T. C., and Eisenberg, D.: The Structure of Melittin. 2. Interpretation of the
787 Structure, J. Biol. Chem., 257, 6016-6022, [https://doi.org/10.1016/S0021-9258\(20\)65098-](https://doi.org/10.1016/S0021-9258(20)65098-0)
788 [0](https://doi.org/10.1016/S0021-9258(20)65098-0), 1982.
- 789 Torchia, D. A.: Slight Mistuning of a Cryogenic Probe Significantly Perturbs the Water H-
790 1 Precession Frequency, J. Biomol. NMR, 45, 241-244, [https://doi.org/10.1007/s10858-](https://doi.org/10.1007/s10858-009-9363-6)
791 [009-9363-6](https://doi.org/10.1007/s10858-009-9363-6), 2009.
- 792 Tsukada, K., Kiwa, T., and Masuda, Y.: AC Magnetic Properties of Large Volume of
793 Water - Susceptibility Measurement in Unshielded Environment, Jap. J. Appl. Phys. Part
794 2-Letters & Express Letters, 45, L1097-L1099, <https://doi.org/10.1143/jjap.45.L1097>,
795 2006.
- 796 Ugurbil, K.: Imaging at Ultrahigh Magnetic Fields: History, Challenges, and Solutions,
797 Neuroimage, 168, 7-32, <https://doi.org/10.1016/j.neuroimage.2017.07.007>, 2018.
- 798 VanderHart: Magnetic Susceptibility & High Resolution NMR of Liquids and Solids, in:
799 Encyclopedia of Nuclear Magnetic Resonance, edited by: Grant, D. M., and Harris, R. K.,
800 Wiley, New York, 1996.
- 801 Wapler, M. C., Leupold, J., Dragonu, J., von Elverfeld, D., Zaitsev, M., and Wallrabe, U.:
802 Magnetic Properties of Materials for MR Engineering, Micro-MR and Beyond, J. Magn.
803 Reson., 242, 233-242, <https://doi.org/10.1016/j.jmr.2014.02.005>, 2014.
- 804 Xia, Y. L., Rossi, P., Subrahmanian, M. V., Huang, C. D., Saleh, T., Olivieri, C.,
805 Kalodimos, C. G., and Veglia, G.: Enhancing the Sensitivity of Multidimensional NMR
806 Experiments by Using Triply-Compensated π Pulses, J. Biomol. NMR, 69, 237-243,
807 <https://doi.org/10.1007/s10858-017-0153-2>, 2017.
- 808